

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0030.0570 Neg Flow Verify: dP2 at -118,0 (Setpoint 120) [15m]	SLPM	-126,2 to -109,8	-118,8	✓	
37	0030.0580 Neg Flow Verify: dP2 at -108,3 (Setpoint 110) [15m]	SLPM	-116,1 to -100,5	-109,0	✓	
38	0030.0590 Neg Flow Verify: dP2 at -98,7 (Setpoint 100) [15m]	SLPM	-106,1 to -91,2	-99,4	✓	
39	0030.0600 Neg Flow Verify: dP2 at -88,8 (Setpoint 90) [15m]	SLPM	-95,9 to -81,8	-89,2	✓	
40	0030.0610 Neg Flow Verify: dP2 at -78,3 (Setpoint 80) [15m]	SLPM	-85,0 to -71,7	-78,7	✓	
41	0030.0620 Neg Flow Verify: dP2 at -68,2 (Setpoint 70) [15m]	SLPM	-74,4 to -61,9	-68,9	✓	
42	0030.0630 Neg Flow Verify: dP2 at -58,0 (Setpoint 60) [15m]	SLPM	-63,8 to -52,2	-58,7	✓	
43	0030.0640 Neg Flow Verify: dP2 at -48,3 (Setpoint 50) [15m]	SLPM	-53,8 to -42,9	-48,5	✓	
44	0030.0650 Neg Flow Verify: dP2 at -38,8 (Setpoint 40) [15m]	SLPM	-43,9 to -33,8	-39,2	✓	
45	0030.0660 Neg Flow Verify: dP2 at -28,5 (Setpoint 30) [15m]	SLPM	-33,1 to -23,9	-28,8	✓	
46	0030.0670 Neg Flow Verify: dP2 at -19,8 (Setpoint 20) [15m]	SLPM	-24,8 to -14,8	-19,2	✓	
47	0030.0680 Neg Flow Verify: dP2 at -10,0 (Setpoint 10) [15m]	SLPM	-14,5 to -5,5	-9,7	✓	
48	0030.0690 Neg Flow Verify: dP2 at -5,0 (Setpoint 5) [15m]	SLPM	-8,8 to -1,2	-5,0	✓	
49	0030.0700 Neg Flow Verify: dP2 at -0,0 (Setpoint 0) [15m]	SLPM	-2,0 to 2,0	-0,0	✓	
50	0030.0950 Contr.Press.Ver.(8361): P1 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,07	✓	
51	0030.0960 Monit.Press.Ver.(3670): P2 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,00	✓	
52	0030.0160 Pprox Press. Ver.(8015): P2 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,05	✓	
53	0030.0970 Contr.Press.Ver.(10727): P1 at 3,89 (SP 4, 1, 1)	cmH2O	2,89 to 4,89	3,99	✓	
54	0030.0980 Monit.Press.Ver.(6397): P2 at 3,89 (SP 4, 1, 1)	cmH2O	2,89 to 4,89	3,88	✓	
55	0030.0161 Pprox Press. Ver.(10390): P2 at 3,89 (SP 4, 1, 1)	cmH2O	2,89 to 4,89	3,98	✓	
56	0030.0990 Contr.Press.Ver.(19163): P1 at 17,89 (SP 18, 1, 1)	cmH2O	16,89 to 18,89	17,99	✓	
57	0030.1000 Monit.Press.Ver.(16139): P2 at 17,89 (SP 18, 1, 1)	cmH2O	16,89 to 18,89	17,85	✓	
58	0030.0162 Pprox Press. Ver.(18862): P2 at 17,89 (SP 18, 1, 1)	cmH2O	16,89 to 18,89	17,99	✓	
59	0030.1010 Contr.Press.Ver.(27594): P1 at 31,91 (SP 32, 1, 1)	cmH2O	30,91 to 32,91	32,00	✓	
60	0030.1020 Monit.Press.Ver.(25905): P2 at 31,91 (SP 32, 1, 1)	cmH2O	30,91 to 32,91	31,89	✓	
61	0030.0163 Pprox Press. Ver.(27355): P2 at 31,91 (SP 32, 1, 1)	cmH2O	30,91 to 32,91	32,03	✓	
62	0030.1030 Contr.Press.Ver.(36038): P1 at 45,89 (SP 46, 1, 1)	cmH2O	44,89 to 46,89	45,99	✓	
63	0030.1040 Monit.Press.Ver.(35658): P2 at 45,89 (SP 46, 1, 1)	cmH2O	44,89 to 46,89	45,90	✓	
64	0030.0164 Pprox Press. Ver.(35756): P2 at 45,89 (SP 46, 1, 1)	cmH2O	44,89 to 46,89	45,98	✓	
65	0030.1050 Contr.Press.Ver.(38446): P1 at 49,89 (SP 50, 1, 1)	cmH2O	48,89 to 50,89	50,01	✓	
66	0030.1060 Monit.Press.Ver.(38459): P2 at 49,89 (SP 50, 1, 1)	cmH2O	48,89 to 50,89	49,90	✓	
67	0030.0165 Pprox Press. Ver.(38224): P2 at 49,89 (SP 50, 1, 1)	cmH2O	48,89 to 50,89	49,97	✓	
68	0030.1070 Contr.Press.Ver.(45413): P1 at 61,48 (SP 65, 1, 1)	cmH2O	60,48 to 62,48	61,49	✓	
69	0030.1080 Monit.Press.Ver.(46575): P2 at 61,48 (SP 65, 1, 1)	cmH2O	60,48 to 62,48	61,51	✓	
70	0030.0166 Pprox Press. Ver.(45192): P2 at 61,48 (SP 65, 1, 1)	cmH2O	60,48 to 62,48	61,46	✓	

Test Started On 04/08/24 09:59:44Serial Number TV01602190BElapsed Test Time 19m 25sStatus **PASS**

Page 2 of 3

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Troubleshooting: Group Level mode.
 File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
 Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 77D6, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20