Test Software P/N 4102483

Tested Part Number LA1054096

 $\textbf{Test Software Name} \ \underline{\texttt{Trilogy FSA (REPAIR TEST)}}$



Software Revision 24.0.0.0

Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0020.0570 Pos Flow Cal: dP2 at 77,0 (Setpoint 75) [23m 44	ADC Cnts	LE 47525	46568	√	
37	0020.0580 Pos Flow Cal: dP2 at 87,5 (Setpoint 85) [23m 44	ADC Cnts	LE 48391	47526	√	
38	0020.0590 Pos Flow Cal: dP2 at 96,3 (Setpoint 95) [23m 44	ADC Cnts	LE 49300	48392	√	
39	0020.0600 Pos Flow Cal: dP2 at 107,3 (Setpoint 105) [23m	ADC Cnts	LE 49797	49301	√	
40	0020.0610 Pos Flow Cal: dP2 at 114,2 (Setpoint 115) [23m	ADC Cnts	LE 50914	49798	√	
41	0020.0620 Pos Flow Cal: dP2 at 128,2 (Setpoint 125) [23m	ADC Cnts	LE 51594	50915	√	
42	0020.0630 Pos Flow Cal: dP2 at 136,8 (Setpoint 135) [23m	ADC Cnts	LE 52229	51595	√	
43	0020.0640 Pos Flow Cal: dP2 at 147,8 (Setpoint 145) [23m	ADC Cnts	LE 53583	52230	√	
44	0020.0650 Pos Flow Cal: dP2 at 169,3 (Setpoint 170) [23m	ADC Cnts	LE 54753	53584	√	
45	0020.0660 Pos Flow Cal: dP2 at 187,8 (Setpoint 190) [23m	ADC Cnts	LE 65435	54754	√	
46	0020.0900 Atmospheric Press Sensors Cal Patm: Slope (m) [lnHg/ADC	EQ 0,000517	0,000517	√	
47	0020.0910 Atmospheric Press Sensors Cal Patm: Intercept (inHg	1,91 to 4,09	2,69	√	
48	0020.0920 Write Sensors Cal Table [25m 5s]	N/A	Pass	TRUE	√	
49	0020.0970 Write Pprox Cal Table [25m 26s]	N/A	Pass	TRUE	√	
5 0	0020.0935 O2 Sensor Cal (not attempted for SW ver.14.2.05	N/A	None	None	√	
51	0020.1030 Write Image Table [25m 46s]	N/A	Pass	TRUE	1	
						_
						
						
						
						<u> </u>
						_
						<u> </u>
						i

Test Started On 06/11/24 03:31:01

Serial Number TV112010980 Elapsed Test Time 25m 46s

Status PASS

Page 2 **of** 2