Test Software P/N 4102483

Tested Part Number LA1054096

 $\textbf{Test Software Name} \ \underline{\texttt{Trilogy FSA}} \ (\texttt{PM TEST})$



Software Revision 24.0.0.0

Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass F	ail
71	0030.1030 Contr.Press.Ver.(35621): P1 at 45,66 (SP 46, 1,	cmH20	44,66 to 46,66	45,99		_
72	0030.1040 Monit.Press.Ver.(36460): P2 at 45,66 (SP 46, 1,	cmH20	44,66 to 46,66	45,59	→	
73	0030.0164 Pprox Press. Ver.(35357): P2 at 45,66 (SP 46, 1	cmH20	44,66 to 46,66	45,95	√	_
74	0030.1050 Contr.Press.Ver.(38024): P1 at 49,64 (SP 50, 1,	cmH20	48,64 to 50,64	49,99	√	_
75	0030.1060 Monit.Press.Ver.(39234): P2 at 49,64 (SP 50, 1,	cmH20	48,64 to 50,64	49,59	√	_
76	0030.0165 Pprox Press. Ver.(37788): P2 at 49,64 (SP 50, 1	cmH20	48,64 to 50,64	49,95	√	_
77	0030.1070 Contr.Press.Ver.(46847): P1 at 64,00 (SP 65, 1,	cmH20	63,00 to 65,00	64,70	√	_
78	0030.1080 Monit.Press.Ver.(49421): P2 at 64,00 (SP 65, 1,	cmH20	63,00 to 65,00	64,21	√	_
79	0030.0166 Pprox Press. Ver.(46585): P2 at 64,00 (SP 65, 1	cmH20	63,00 to 65,00	64,67	√	_
8 0	0030.1150 Atmospheric Pressure Verify: Patm at 28,05,28,0	inHg	27,45 to 28,65	27,91	√	—
81	0030.1250 Airstream Temperature (Tair) at 26,00 [27m 50s]	C deg	21,00 to 41,00	30,10	√	_
					\perp	
					\perp	
					\perp	
					\perp	
						_
						_
					+	_
					+	_
					++	—
					++	—

Test Started On 01/15/24 09:51:07

Elapsed Test Time 27m 50s

Serial Number TV11412164B

Page 3 **of** 3