Test Software P/N 4102483

**Tested Part Number** 1040000

Test Software Name Trilogy FSA (REPAIR TEST)



**Software Revision** 24.0.0.0

Prepared By P. Pascal

## Test Report

Tost Report					
Step	Description	Units	Limits	Results	Pass Fail
36	0021.0570 02 Positive Flow Cal: dP1 at 65,0 (SP 65 @74,4P	\DC Cnts	LE 1922	1782	<u> </u>
37	0021.0580 02 Positive Flow Cal: dP1 at 74,5 (SP 75 @72,6P	ADC Cnts	LE 2048	1923	<b>√</b>
38	0021.0590 02 Positive Flow Cal: dP1 at 84,0 (SP 85 @72,3P	ADC Cnts	LE 2180	2049	<b>√</b>
39	0021.0600 O2 Positive Flow Cal: dP1 at 94,3 (SP 95 @70,6P	ADC Cnts	LE 2300	2181	<b>√</b>
40	0021.0610 02 Positive Flow Cal: dP1 at 103,7 (SP 105 @70,	ADC Cnts	LE 2423	2301	<b>√</b>
41	0021.0620 O2 Positive Flow Cal: dP1 at 114,9 (SP 115 @68,	ADC Cnts	LE 2522	2424	<b>√</b>
42	0021.0630 O2 Positive Flow Cal: dP1 at 124,0 (SP 125 @67,	ADC Cnts	LE 2620	2524	<b>√</b>
43	0021.0640 O2 Positive Flow Cal: dP1 at 134,0 (SP 135 @65,	ADC Cnts	LE 2716	2621	✓
44	0021.0650 O2 Positive Flow Cal: dP1 at 144,3 (SP 145 @63,	ADC Cnts	LE 2980	2717	<b>√</b>
45	0021.0660 02 Positive Flow Cal: dP1 at 174,2 (SP 175 @59,	ADC Cnts	LE 3114	2981	✓
46	0021.0670 O2 Positive Flow Cal: dP1 at 190,2 (SP 190 @55,	ADC Cnts	LE 3999	3115	✓
47	0021.0880 02 Sensor Cal (NO_O2_ERROR): 560 mV at ambient	N/A	Pass	TRUE	<b>√</b>
48	0021.0930 Write OBM Sensors Cal Table (V: TRUE, A: TRUE)	N/A	Pass	TRUE	✓

**Test Started On** 07/10/24 12:08:55

Serial Number TV017031585 Elapsed Test Time 16m 32s

Status PASS

Page 2 **of** 2