

Test Software P/N 4102483Tested Part Number 1040000Test Software Name Trilogy FSA (REPAIR TEST)Software Revision 24.0.0.0Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
1	0021.0007 OBM O2 Sensor Heater On (NO_O2_ERROR) [0s]	N/A	Pass	TRUE	✓	
2	0021.0010 O2 Press Sensor Cal P1@ 25,35 & 75,20 PSI: Slope	ADC Cnts	0,03010 to 0,0510	0,03341	✓	
3	0021.0020 O2 Press Sensor Cal P1@ 25,35 & 75,20 PSI: Inter	PSI	-25,00 to -10,00	-13,37	✓	
4	0021.0030 O2 Press Sensor Cal P1@ 25,35 & 75,20 PSI: Zero	ADC Cnts	320 to 490	414	✓	
5	0021.0120 Neg Flow Cal: dP2 at -48,0 (Setpoint 45) [4m 49s]	ADC Cnts	GE 101	192	✓	
6	0021.0130 Neg Flow Cal: dP2 at -33,0 (Setpoint 35) [4m 49s]	ADC Cnts	GE 192	421	✓	
7	0021.0140 Neg Flow Cal: dP2 at -23,0 (Setpoint 25) [4m 49s]	ADC Cnts	GE 422	588	✓	
8	0021.0150 Neg Flow Cal: dP2 at -15,0 (Setpoint 15) [4m 49s]	ADC Cnts	GE 589	784	✓	
9	0021.0160 Neg Flow Cal: dP2 at -5,0 (Setpoint 5) [4m 49s]	ADC Cnts	GE 785	1065	✓	
10	0021.0170 Neg Flow Cal: dP2 at -0,0 (Setpoint 0) [4m 49s]	ADC Cnts	GE 1066	1180	✓	
11	0021.0250 Raw Zero Flow: dP2 at 0,0 (Setpoint 0) [4m 49s]	ADC Cnts	GE 0	1180	✓	
12	0021.0300 Pos Flow Cal: dP2 at 5,0 (Setpoint 5) [8m 57s]	ADC Cnts	LE 1561	1297	✓	
13	0021.0310 Pos Flow Cal: dP2 at 15,0 (Setpoint 15) [8m 57s]	ADC Cnts	LE 1770	1562	✓	
14	0021.0320 Pos Flow Cal: dP2 at 24,0 (Setpoint 25) [8m 57s]	ADC Cnts	LE 1996	1771	✓	
15	0021.0330 Pos Flow Cal: dP2 at 38,0 (Setpoint 35) [8m 57s]	ADC Cnts	LE 2114	1997	✓	
16	0021.0340 Pos Flow Cal: dP2 at 47,0 (Setpoint 45) [8m 57s]	ADC Cnts	LE 2255	2115	✓	
17	0021.0350 Pos Flow Cal: dP2 at 57,0 (Setpoint 55) [8m 57s]	ADC Cnts	LE 2379	2256	✓	
18	0021.0360 Pos Flow Cal: dP2 at 67,0 (Setpoint 65) [8m 57s]	ADC Cnts	LE 2493	2380	✓	
19	0021.0370 Pos Flow Cal: dP2 at 77,7 (Setpoint 75) [8m 57s]	ADC Cnts	LE 2601	2494	✓	
20	0021.0380 Pos Flow Cal: dP2 at 87,7 (Setpoint 85) [8m 57s]	ADC Cnts	LE 2694	2602	✓	
21	0021.0390 Pos Flow Cal: dP2 at 97,0 (Setpoint 95) [8m 57s]	ADC Cnts	LE 2784	2695	✓	
22	0021.0400 Pos Flow Cal: dP2 at 107,9 (Setpoint 105) [8m 57s]	ADC Cnts	LE 2861	2785	✓	
23	0021.0410 Pos Flow Cal: dP2 at 117,0 (Setpoint 115) [8m 57s]	ADC Cnts	LE 2948	2862	✓	
24	0021.0420 Pos Flow Cal: dP2 at 128,0 (Setpoint 125) [8m 57s]	ADC Cnts	LE 3030	2949	✓	
25	0021.0430 Pos Flow Cal: dP2 at 137,5 (Setpoint 135) [8m 57s]	ADC Cnts	LE 3101	3031	✓	
26	0021.0440 Pos Flow Cal: dP2 at 148,0 (Setpoint 145) [8m 57s]	ADC Cnts	LE 3272	3102	✓	
27	0021.0450 Pos Flow Cal: dP2 at 173,3 (Setpoint 175) [8m 57s]	ADC Cnts	LE 3366	3273	✓	
28	0021.0460 Pos Flow Cal: dP2 at 188,3 (Setpoint 190) [8m 57s]	ADC Cnts	LE 3999	3367	✓	
29	0021.0500 O2 Positive Flow Cal: dP1 at 0,0 (SP 0 @80,9PSI)	ADC Cnts	LE 373	243	✓	
30	0021.0510 O2 Positive Flow Cal: dP1 at 4,3 (SP 5 @83,6PSI)	ADC Cnts	LE 730	374	✓	
31	0021.0520 O2 Positive Flow Cal: dP1 at 15,0 (SP 15 @83,2PSI)	ADC Cnts	LE 968	731	✓	
32	0021.0530 O2 Positive Flow Cal: dP1 at 24,0 (SP 25 @82,8PSI)	ADC Cnts	LE 1185	970	✓	
33	0021.0540 O2 Positive Flow Cal: dP1 at 34,9 (SP 35 @82,5PSI)	ADC Cnts	LE 1354	1186	✓	
34	0021.0550 O2 Positive Flow Cal: dP1 at 44,0 (SP 45 @82,2PSI)	ADC Cnts	LE 1524	1355	✓	
35	0021.0560 O2 Positive Flow Cal: dP1 at 54,5 (SP 55 @81,7PSI)	ADC Cnts	LE 1680	1525	✓	

Test Started On 02/06/24 05:22:32Serial Number TV02001060EElapsed Test Time 14m 15sStatus **FAIL**

Page 1 of 2

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Troubleshooting: Group Level mode.
 File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
 Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 77D6, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20