


▪ **System Components (Hardware Components):**

Internal (Implanted):

- Rechargeable battery-powered computer that is responsible for generating neuro stimulation pulses.
- Bluetooth communication system linked to the remote control.
- A cable that carries those pulses to the hypoglossal nerve.
- A sensor that detects the breathing of the patient.



<https://www.inspiresleep.com/learn/>

External:

A remote control that allows patients to start, stop or pause the therapy (neurostimulation)



<https://www.inspiresleep.com/learn/>