

I. Dave Boldridge

Digestive:

To evaluate the currently available data on the vapor pressure of nicotine at moderate temperatures (20 - 100°C) and to evaluate the potential for using the Tapered Element Oscillating Microbalance (TEOM) to obtain experimental values of the vapor pressure of nicotine in the temperature region of interest.

Summary:

Several studies of the vapor pressure of nicotine have been published. ~~Only two of these~~ studies address overlapping temperature regions, and there are gross differences in the vapor pressures reported. Some obvious sources of potential error are ~~reported~~ identified and discussed for each of these experiments.

Vapor pressure measurements with the TEOM are reported and compared to the published values. The major weaknesses in the experimental technique are discussed. The values obtained can be considered lower limits for the vapor pressure of nicotine.

last
to range
of values
not the
best (in
our judgment)
value on
our own
quarantined
material at
ambient pressure