

Fw: Final revisions for manuscript COMMSBIO-24-7622A

Florian MULLER

lun. 16/02/2026 19:01

Boîte de réception

À Christian WEBER <christian.weber@pasteur.fr>; Chloé STURMACH <chloe.sturmach@pasteur.fr>; Jacques BOURG
: <jacques.bourg@pasteur.fr>;

1 pièce(s) jointe(s) (237 Ko)

CommsBio_AIP_table-7622A_1771261884_31.docx;

Finally!

Some minor modifs and then we are good!

Congratulations

Sent from [Outlook for Android](#)

From: kaliya.georgieva@nature.com <kaliya.georgieva@nature.com>

Sent: Monday, February 16, 2026 6:12:01 PM

To: florian.muller@pasteur.fr <florian.muller@pasteur.fr>

Subject: Final revisions for manuscript COMMSBIO-24-7622A

** Please ensure you delete the link to your author homepage in this email if you wish to forward it to your coauthors **

Dear Dr Mueller,

Your manuscript entitled "autoFISH - a modular toolbox for sequential smFISH experiments" has now been seen again by our referees, whose comments appear below. In light of their advice I am delighted to say that we are happy, in principle, to publish a suitably revised version in Communications Biology.

We therefore invite you to revise your paper one last time to address the remaining minor points of our reviewers. At the same time we ask that you edit your manuscript to comply with our format requirements and to maximise the accessibility and therefore the impact of your work.

* Please see the attached document for editorial requests for the final version (.docx file). Please ensure a completed version of this file is uploaded as a Related Manuscript with your final submission.

* Please review our [final submission file checklist](#) to ensure all necessary files are present with your final submission and to avoid delays in accepting your manuscript. For your reference, a style and formatting guide is available [here](#) and includes all of our style requirements.

It is important that you pay careful attention to the requests in these documents to avoid a delay in formal acceptance of the article.

If you have any questions or concerns about any of our requests, please do not hesitate to contact me.

Open access

Communications Biology is a fully open access journal. Articles are made freely accessible on publication. For further information about article processing charges, open access funding, and advice and support from Nature Research, please visit <https://www.nature.com/commsbio/open-access>

At acceptance, you will be provided with instructions for completing the open access licence agreement on behalf of all authors. This grants us the necessary permissions to publish your paper. Additionally, you will be asked to declare that all required third party permissions have been obtained, and to provide billing information in order to pay the article-processing charge (APC).

Please use the following link to upload your revised files:

<https://mts-commsbio.nature.com/cgi-bin/main.plex?el=A5Cx3LDP5B1BlwH7l4A9ftdb3bnplvmBmP64ksFNQbdNQZ>

** This url links to your confidential home page and associated information about manuscripts you may have submitted or be reviewing for us. If you wish to forward this email to co-authors, please delete the link to your homepage first **

We hope to hear from you within two weeks. If you expect the process to take longer than one month, please let us know.

Congratulations on an excellent paper!

Best regards,

Kaliya Georgieva, PhD
Associate Editor
Communications Biology
<https://orcid.org/0009-0006-2251-2578>

PS: At acceptance, the corresponding author will be provided with instructions for completing the license on behalf of all authors. This grants us the necessary permissions to publish your paper. Additionally, you will be asked to declare that all required third party permissions have been obtained, and to provide billing information in order to pay the article-processing charge (APC).

REVIEWERS' COMMENTS:

Reviewer #1 (Remarks to the Author):

In the manuscript by Weber et al. titled, "autoFISH: a modular toolbox for sequential smFISH experiments," the authors made a great effort to lay out a comprehensive workflow for implementing a custom-built fluidic device-assisted sequential fluorescence in situ hybridization (smFISH) experiments with the goal of making the very approach more accessible and available for the research community. In this work, they provided a manual with the critical points and tips for the designing, building and operating a sequential smFISH experiment. Furthermore, they made detailed manuals publicly available through the Github repository. As a proof of principle, the authors performed an exemplary sequential smFISH on cultured HeLa cells and cleared mouse lung tissues, and the results of which demonstrated impressive robustness and utility of the autoFISH workflow.

As I pointed out previously, this work is inspiring in that it provides service to broader research community and carries great values as a manual that may further inspire the research community to join the collective effort in optimizing fluidic device-assisted sequential smFISH protocols and the expansion of instrumental modules. The revised manuscript has appropriately addressed all points I made and further improved the overall quality of the manuscript substantially. Therefore, I recommend its publication.

Reviewer #2 (Remarks to the Author):

Review Paper # COMMSBIO-24-7622A

The manuscript presents autoFISH, a modular, open-source platform for automated sequential smFISH. The system integrates customizable hardware, flexible Python-based control software, and experimental protocols supporting both signal amplification (SABER) and tissue clearing. The authors validate the platform in HeLa cells and cleared mouse lung tissue, demonstrating robust signal detection, efficient probe stripping across 20 hybridization rounds, and improved nuclear signal retention using a GMA-based clearing modification. Overall, this work provides a cost-effective and adaptable alternative to commercial spatial transcriptomics systems.

In the revised version, the authors have addressed all reviewer comments, substantially improved the clarity of the manuscript, and included the requested experimental controls. We therefore support publication, pending the following minor revisions:

- Introduction: Key references to multiplexed FISH approaches are still missing (besides your own publication) after the sentence at line 3 ("Various adaptations of this method...").
- Methods (Microscopy): The filter sets used for image acquisition should be specified.
- Methods (smFISH readout and stripping oligos): The authors describe a strategy to test the quality of readout sequences ("We observed that some of the original readout sequences yielded lower signals during RNA smFISH..."). As this is a valuable control, we suggest adding a schematic to the Supplementary Information to illustrate the strategy, along with representative image examples of both poor and robust readout sequences.
- Figure 2: The figure legends for panels C, F, G, and H do not describe what the lighter background shading represents in the line plots.
- Figure 3: The figure legend does not specify whether the central line in the boxplots represents the median or the mean, nor does it define the boxes and whiskers.
- Supplementary Figures 1 and 2: Scale bars are missing from the images and should be added.

** See the Nature Portfolio author and referees' website at www.nature.com/authors for information about policies, services and author benefits

This email has been sent through the Springer Nature Tracking System NY-610A-NPG&MTS

Confidentiality Statement:

This e-mail is confidential and subject to copyright. Any unauthorised use or disclosure of its contents is prohibited. If you have received this email in error please notify our Manuscript Tracking System Helpdesk team at <http://platformsupport.nature.com>.

Details of the confidentiality and pre-publicity policy may be found here <http://www.nature.com/authors/policies/confidentiality.html>

[Privacy Policy](#) | [Update Profile](#)

autoFISH: a modular toolbox for sequential single-molecule RNA FISH experiments

Christian Weber^{1,2,*}, Thomas Defard^{1,2,3,4,5*}, Chloé Sturmach¹, Mickael Lelek⁶, Hugo Laporte^{6,7}, Ayan Mallick⁶, Maria Isabella Gariboldi^{1,2}, José-Arturo Londoño-Vallejo⁶, Thomas Walter^{3,4,5}, Charles Fouillade⁶, Jacques Bourg¹, Florian Mueller^{1,2,*,#}

¹Institut Pasteur, Université Paris Cité, Photonic Bio-Imaging, Centre de Ressources et Recherches Technologiques (UTechS-PBI, C2RT), Paris, France

² Institut Pasteur, Université Paris Cité, Imaging and Modeling Unit, Paris, France

³ Centre for Computational Biology (CBIO), Mines Paris, PSL University, Paris, France

⁴ Institut Curie, PSL University, Paris, France

⁵ INSERM, U900, 75005, Paris, France

⁶ Institut Curie, Inserm U1021-CNRS UMR 3347, University Paris-Saclay, PSL Research University, Centre Universitaire, Orsay, France.

⁷ Institute of Cell Biology (Cancer Research), University Hospital Essen, Essen, Germany.

* These authors contributed equally to this work

email: fmueller@pasteur.fr

Abstract

Understanding the role of gene expression in cellular function and tissue organization requires spatial and quantitative detection of individual RNA molecules; however, the widespread adoption of automated single-molecule fluorescence in situ hybridization (smFISH) has been limited by the cost of equipment and the complexity of experimental procedures. We present autoFISH, an affordable, user-friendly platform that removes these barriers through open-source hardware components, accessible control software, and integrated analysis tools. The system demonstrates broad applicability by enabling both conventional and signal-amplified smFISH protocols and incorporates an optimized tissue-clearing method that preserves nuclear structures. Testing across multiple cell types and tissue preparations validates the system's reliability and reproducibility, offering a practical solution for scaling spatial transcriptomics research and advancing discoveries in cellular and developmental biology, while significantly reducing costs and the technical expertise required.

Introduction

Transcription is a fundamental biological process pivotal to development and disease. Individual cells exhibit considerable heterogeneity in their transcriptional profiles, including RNA abundance and intracellular RNA localization. Accurately quantifying the number of RNAs within cells and their locations provides a precise readout of a cell's molecular state and enhances our understanding of RNA regulation. Such measurements require spatially resolved RNA detection

in the native cellular environment with sub-cellular accuracy. Here, single-molecule FISH (smFISH) is widely regarded as the gold standard for detecting individual RNA molecules with high sensitivity and precise sub-micrometer localization (Femino et al., 1998). Various adaptations of this method have been developed, all based on the principle of targeting RNAs with multiple oligonucleotide fluorescent probes, either directly labeled (Raj et al., 2008) or indirectly labeled via secondary probes (Tsanov et al 2016) or amplifying probe systems (e.g, branched-DNA (Xia et al., 2019), HCR-FISH (Choi et al., 2018), SABER (Kishi et al., 2019)).

RNAs then appear as diffraction-limited spots in fluorescence microscopy images, and their positions can be measured with high spatial accuracy using dedicated analysis workflows (Bahry et al., 2021; Imbert et al., 2022; Xu et al., 2024). When combined with cell segmentation (Stringer et al., 2020), the RNA abundance in each cell can be estimated, and statistical analysis can be used to quantitatively describe subcellular RNA localization (Chouaib et al., 2020; Mah et al., 2024). Crucially, smFISH preserves the spatial context of cells while revealing their transcriptional states. This dual output facilitates advanced quantitative analysis of tissue architecture by integrating molecular profiles, cellular identities, and spatial information. Cell types can be identified using standard scRNA-seq clustering approaches (Traag et al., 2019), and their annotation can be further refined by incorporating spatial data (Teng et al., 2022). This enables large-scale investigation of features such as cell-cell communication, tissue composition, and complex tissue gradients (Palla et al., 2022).

A significant limitation of the original smFISH methods is that the number of spectrally distinguishable fluorophores constrains the number of distinct RNA species that can be visualized simultaneously. To overcome this limitation, new methods have been developed that rely on iterative hybridizations, i.e., alternating cycles of probe hybridization, imaging, and signal removal (Chen et al., 2015; Codeluppi et al., 2018; Lubeck et al., 2014). A general feature of these approaches is that the primary probes carry one or more readout sequences that can be detected by a secondary, complementary readout oligo. Usually, primary probes against all RNA targets are hybridized on the bench, while the readout oligos are hybridized directly on the microscope with an automated fluidic system. Signal removal can be achieved by various methods; the most commonly used are stripping with high-stringency buffers (Lohoff et al., 2022), fluorophore removal via chemical cleavage (Moffitt et al., 2016b), or displacement with a tertiary stripping oligo (Mateo et al., 2021). In sequential smFISH experiments, the number of targeted species is linearly correlated with the number of iterations and the number of fluorophores used. Multiplexed barcoding approaches have driven the field of spatial transcriptomics by dramatically increasing RNA detection throughput (Cheng et al., 2023). Here, each RNA species carries several readout sequences forming unique RNA-specific barcodes. Although readout sequences are reused across RNAs, each hybridization round visualizes a different set of genes. RNA identities must be decoded post-analysis. When combined with error correction, this combinatorial method enables the detection of hundreds to thousands of RNA species.

While several commercial systems offer convenient turnkey solutions with increasingly large panel sizes, they are costly in terms of instrumentation and reagents and lack flexibility.

Exploratory analyses, time courses, dose-response studies, and replicates thus demand substantial financial resources. While large RNA panels are often necessary, not all studies require such depth; smaller, more flexible panels may suffice. For example, a small probe panel with 10 genes can be adequate to resolve the main cell types in tumor cryosections (Rademacher et al., 2024).

While these commercial methods are user-friendly, they often lack flexibility across microscopy (e.g., imaging modalities) and experimental protocols (e.g., RNA detection approaches). This limitation can be critical for challenging samples, such as tissue sections, where weak RNA signals benefit from amplification to improve detection reliability. Several widely used amplification strategies include branched DNA (Wan et al., 2024; Xia et al., 2019), hybridization chain reaction (HCR) (Choi et al., 2018; Gandin et al., 2025), and SABER (Kishi et al., 2019). SABER is appealing due to its simple base protocol and tunable amplification levels.

Taken together, home-built solutions remain a crucial option for addressing many research questions. To address this need, we propose autoFISH, an end-to-end open-source workflow comprising hardware, software, and experimental protocols for automated smFISH experiments designed for flexibility and affordability. The autoFISH toolbox relies on a versatile and affordable fluidic system, for which we provide detailed building plans. Flexibility comes in several flavors. First, the imaging modality can be chosen depending on the desired application, as the fluidics system required for sequential hybridizations can be mounted on various microscopes. Second, protocols can be readily modified to meet experimental needs, such as by incorporating signal amplification or tissue clearing. Lastly, smaller-scale experiments can be readily performed because probe design and synthesis are amenable to control.

We demonstrate the applicability of autoFISH on two experimental datasets. Firstly, HeLa cells, which we use to benchmark the performance of the protocol and illustrate its flexibility by alternating between standard and amplified FISH experiments with SABER (Kishi et al., 2019), and secondly, mouse lung tissue, which we use to visualize several marker genes from our earlier study (Curras-Alonso et al., 2023). Here, we introduce a modified tissue-clearing protocol that improves nuclear signal. In summary, we provide a comprehensive toolbox for establishing automated, sequential smFISH experiments at a lower cost.

Results

Here, we present autoFISH (Figure 1), a universal toolbox for performing sequential smFISH, with instrumentation that can be built in-house at an affordable cost and adapted to specific needs and experimental protocols using standard buffers. autoFISH is based on several previous studies (Bintu et al., 2018; Imbert et al., 2022; Kishi et al., 2019; Moffitt and Zhuang, 2016; Tsanov et al., 2016) and is entirely open-source. We also provide Python workflows to analyze these data and infer single-cell RNA quantification. autoFISH is hosted on GitHub (see Methods) and includes detailed build plans, control software, and experimental protocols. To validate and demonstrate the capability of autoFISH, we implemented and validated different protocols, including workflows for signal amplification and tissue clearing on both cultured cells

and tissue sections.

Computer-controlled fluidics system for automated experiments.

A sequential FISH experiment consists of several key steps: hybridization of primary oligos overnight at 37 °C, sample mounting on the microscope, and alternating imaging and incubation steps for fluorescently labeled secondary probe hybridization and displacement of secondary probes from the previous round (Figure 1A).

The core of autoFISH is a computer-controlled fluidics system that performs the sequential hybridization steps and automates imaging. This system enables flushing buffers from syringes, providing either large buffer reservoirs or single-use buffers stored in multiwell plates, which can be accessed via a pipetting robot. A computer-controlled valve allows buffer selection, and flow control is achieved via a peristaltic pump, with flow rates monitored by a flow sensor. Lastly, air bubbles are removed from the buffers with a bubble trap. Samples can be mounted in an imaging chamber, which is ideal for single-condition cell culture or tissue samples (Figure 1C). We emphasize that placing the pump immediately after the valve yielded better performance than placing it at the end of the fluidic system (after the imaging chamber). Based on our experience, this resulted in higher bubble-trap efficiency and reduced deformation of the imaging chamber during pump action, which can otherwise lead to loss of microscope autofocus.

We implemented a modular, open-source Python package that controls all components of the fluidic setup to perform automated buffer exchanges. The code's modularity facilitates adding new components from other providers, such as a different peristaltic pump. The user can flexibly define fluidics runs in a human-readable text file that contains all buffers, pump durations, and incubation times, as well as optional steps, such as performing DAPI staining only for one run or adding hybridization steps, e.g., for signal amplification with SABER.

For fully automated experiments, imaging must be automated as well. As a design choice, we do not directly control microscopes within our code; instead, we provide methods to interface with acquisition software packages. Currently, we support acquisition in Micromanager via Pycromanager, or triggered acquisition via a TTL pulse or a shared synchronization text file (see Methods for details). With these solutions, autoFISH can already be used with various imaging systems, and new trigger approaches can be added as needed (Figure 1D).

Lastly, we provide a graphical user interface (GUI) to facilitate usage. This interface can be used to control the entire workflow: load the configuration files, initiate the fluidics system, trigger the microscope, and then launch either an individual fluidics run or a fully automated run, in which imaging and fluidic runs are performed sequentially and iteratively.

autoFISH with 20 hybridization rounds

We first demonstrated the system's robustness in generating specific smFISH signals and its

efficiency in signal removal. Assessing RNA detectability in such sequential smFISH experiments relies on several key analyses. Probe stripping efficiency is a fundamental test, evaluated by comparing images for the same target RNA before stripping and in subsequent rounds after stripping and re-hybridization. The absence of signal in the stripped rounds confirms efficient probe removal, which is critical to preventing false-positive detection. Consistency in spot counts across identical hybridization rounds provides another crucial metric; maintaining a stable number of detected RNA spots indicates robust, reproducible detection efficiency across multiple rounds. Cross-round colocalization analysis directly assesses whether the same individual transcripts are reliably detected across different rounds. A high colocalization rate (e.g., ~90%) strongly indicates consistent identification of the same transcripts. Monitoring the signal-to-noise ratio (SNR, spot intensity divided by the mean local background fluctuations) is essential for assessing whether the signal-to-background ratio remains sufficient for reliable detection. Tracking signal and background intensities complements SNR analysis, helping to diagnose the cause of any changes. For instance, observing a slight decrease in both the actual RNA signal and background intensity, while SNR remains stable, suggests a minor loss of probe binding without compromising detectability. These analysis metrics collectively demonstrate the robustness of detection over the entire multiplexing cycle sequence, proving the technique's capability for large-scale experiments.

We therefore performed an experiment comprising 20 rounds in which we alternated probing two genes with distinct subcellular localization patterns: *XPO1*, which is randomly distributed, and *KIF1C*, which preferentially localizes to cell protrusions (Chouaib et al., 2020). The same RNAs are observed in the first and last hybridization rounds for each gene (Figure 2A). Importantly, probe removal is efficient as the stripped signal from previous rounds is not visible (Figure 2B). More detailed analysis revealed a consistent number of detected spots per round (Figure 2C). Importantly, the detected spots were the same, as indicated by a high colocalization percentage (approx. 90%; Figure 2D-E and Supplementary Figure 1), which is consistent with other studies that perform dual-color labeling of the same RNA (Tsanov et al., 2016). To evaluate signal conservation, we calculated the signal-to-noise ratio for each round. It remained constant (Figure 2F). Closer inspection revealed that subsequent washing rounds slightly reduced both the actual RNA signal and the background (Figure 2G-H). In summary, we demonstrated robust RNA detection over 20 hybridization rounds. Using multiple distinct fluorophores allows the visualization of several tens of RNAs.

Flexible signal amplification with SABER

In smFISH, each RNA molecule is targeted by multiple oligonucleotides. This leads to a strong local signal at each RNA molecule, detectable under a microscope. Signal amplification can help to increase RNA detection efficiency for more challenging samples. In autoFISH, we place readout sequences on both 3' and 5' of the primary oligos. Further, each imaging oligo carries two fluorophores. So each primary oligo is labeled with up to 4 fluorophores. If the obtained signal is not strong enough, we occasionally also label the secondary oligos with two fluorophores. However, this is rather costly. An elegant alternative is SABER, which offers a robust, programmable solution (Kishi et al., 2019). In short, it allows oligos with a repeated

sequence to be extended in vitro before use in hybridization (Figure 3A). These extended oligos can then be bound by multiple imaging oligos, thereby amplifying the signal. The length of extended oligos can be precisely controlled by adjusting the amplification reaction duration (Figure 3A). A significant advantage of amplification prior to hybridization is that the duration of the actual FISH experiment is minimally affected. To increase flexibility and simplify the approach, we explored additional strategies: (1) instead of amplifying oligos, we used synthesized oligos containing multiple binding sites. Due to synthesis length limitations (typically ~150 nt), these oligos generally include three binding sites. We also evaluated labeling each imaging oligo with two fluorophores—one at the 3' end and one at the 5' end—to further enhance signal strength. Our comparisons revealed that signal intensity generally scales with the number of binding sites. Adding fluorophores at both ends produced an additional increase in signal, though not a twofold improvement, potentially due to steric hindrance or quenching effects (Supplementary Figure 1). These results indicate that for moderate amplification, directly synthesized oligos with three repeats can already provide sufficient signal enhancement.

We integrated SABER into autoFISH by replacing the imager sequence on the readouts with the SABER PER sequence required for amplification (Figure 3A). We then optimized the protocol to be compatible with the fluidics system. In short, we replaced the formamide from the original buffers with the less toxic Ethylene Carbonate. We also hybridized the SABER oligos in two steps on the fluidics system: first, the amplicons, then the imaging oligos. We performed the hybridization of the amplicons in a more stringent buffer, resulting in a substantial decrease in background. We also found that adding even a moderate amount of dextran (5% was compatible with our fluidics system) reduced the amount of required imaging oligo, thereby reducing experimental costs. With these changes, SABER can be performed on the fluidics system. Interestingly, the flexibility in the fluidics control allows us to perform SABER hybridization, which requires an additional hybridization step, in selected runs only, as demonstrated in the validation experiments.

We then performed an experiment targeting *XPO1* and *NFKB1A*, alternating between normal and SABER hybridization, and included a control round in which we stripped SABER without a new hybridization (Figure 3B). Qualitative inspection of the images shows clear amplification of SABER for both genes, with the same RNAs visible and SABER resulting in increased signal intensity (Figure 3C). Importantly, the SABER amplicon can also be efficiently stripped and the signal removed (Figure 3C). A more detailed analysis confirmed this qualitative impression, showing a constant amount of detected RNA per round (Figure 3D). This also revealed that SABER substantially increased RNA intensity (Figure 3E) and slightly increased background (Figure 3E). Importantly, co-localization between detection rounds (both classic and SABER FISH) for the same gene is high (Figure 3F). In contrast, co-localization between the two genes is very low, indicating no cross-detection.

These experiments demonstrate how SABER can be used in autoFISH to amplify signals. They also demonstrate the control software's flexibility by varying hybridization conditions for specific runs.

autoFISH on cleared mouse lung tissue

In a recent study, we performed smFISH of selected RNA markers on mouse lung tissue (Curras-Alonso et al., 2023). While imaging quality was sufficient for analysis, we also encountered substantial background fluorescence. It has been demonstrated that this background arises from both tissue autofluorescence and nonspecific probe binding to various cellular structures (Moffitt et al., 2016a). Tissue-clearing methods help to efficiently reduce this background. Here, the biological sample is embedded into a nonswellable polyacrylamide (PA) matrix (Figure 4A). Several strategies exist to anchor RNAs to the PA, either through chemical compounds (Chen et al., 2016) or the more widely used approach of oligos targeting the poly (A) tails of mRNAs that carry a terminal acrydite moiety (Moffitt et al., 2016a). The tissue can then be treated with Proteinase K and SDS to remove proteins and lipids, which are significant sources of autofluorescence and nonspecific binding (Moffitt et al., 2016a).

When applying the protocol with the Poly-T anchoring oligos to lung tissue, the background was reduced, and RNAs were detected. Still, we found that the nuclear signal was often less distinct than in samples anchored with epoxide (Figure 4B). For samples such as lung tissue with a locally high nuclear density, this could adversely affect nuclear segmentation. We speculated that this could be due to reduced chromatin constraints after digestion. In a recent study, an epoxide (GMA) was introduced as a universal linking agent in Expansion Microscopy (ExM) for proteins and RNA (Cui et al., 2023). GMA possesses two reactive sites, allowing covalent linkage of various functional groups—particularly amine moieties—to the acrylamide gel matrix. Since the polymers used for clearing and ExM are similar, we speculated that GMA could also be used in clearing protocols and serve as a structural scaffold.

We compared DAPI staining of HeLa cells after clearing under three conditions: no anchoring, anchoring with GMA, and anchoring with GMA plus poly-T oligos. When GMA was used—either alone or with poly-T anchor oligos—nuclear staining appeared crisper (Figure 4B). To provide a more comprehensive comparison, we analyzed these images using a quantitative sharpness metric (see Methods, (Helmlí and Scherer, 2001)). This analysis confirmed our qualitative observations and demonstrated increased sharpness of the DAPI signal upon addition of GMA (Supplementary Figure 2).

We then performed autoFISH in cleared mouse lung tissue against 6 targets from our previous study (Curras-Alonso et al., 2023). Individual RNA molecules for each target were successfully detected (Figure 4C). Because dedicated membrane markers were not used, we applied our analysis package, ComSeg (Defard et al., 2024b), to segment cells using nuclear staining and to detect RNA point clouds (Figure 4D). The resulting gene expression profiles then enable clustering (Figure 4E) and mapping the identified cell types back onto the image, revealing their spatial locations.

In summary, we developed an improved tissue-clearing procedure that utilizes GMA to enhance nuclear signal retention. This method was successfully applied to mouse tissue, enabling the detection of various marker genes.

Discussion

smFISH is a powerful technique for visualizing individual RNAs within their native cellular context. By performing multiple cycles of probe hybridization, imaging, and probe removal, the number of RNA species that can be analyzed is greatly increased. However, this also requires more complex protocols and specialized equipment. Although commercial solutions are available, they are costly and lack flexibility. Here, we introduce autoFISH, a modular toolkit for automated, sequential FISH that leverages validated, optimized workflows.

In this work, we demonstrate the performance of autoFISH across cell lines and tissues, including up to 20 hybridization runs. While the throughput of this method is lower than that of commercial multiplexed solutions, it offers some advantages. First, because each RNA is imaged individually, spatial crowding is not a concern, thereby allowing even highly expressed genes to be targeted. Additionally, by ordering oligos for each gene individually, we can customize the selection of targeted genes. One of the main strengths of autoFISH is its flexibility. We demonstrate this by applying various experimental protocols, showcasing flexible control over fluidic runs via optional steps, and by implementing different methods to trigger acquisitions on the microscope.

For more challenging samples, we provide experimental protocols for signal amplification and tissue clearing that are compatible with the autoFISH fluidic system. We illustrate how signal amplification can be achieved with SABER and demonstrate that it can be performed on specific runs via the flexibility of the fluidic design. The proposed use of epoxide in tissue clearing enhances nuclear signal preservation, promising to improve segmentation in densely packed samples. Additionally, the epoxide used (GMA) is inexpensive, allowing for cost-effective experiments. Finally, epoxide anchoring is more versatile than targeting poly(A) tails, which limits its application to mature mRNAs. This versatility ensures that other RNA species, such as nascent or long non-coding RNAs, can also be anchored and are less likely to be washed out during the clearing step.

We offer several options for triggering acquisitions, with the potential to add more in the future, making it straightforward to integrate autoFISH into different imaging systems. The open design of the control software facilitates the addition of new components, such as valves or pumps. The control software is implemented in Python, enabling easy implementation of additional steps, such as quality control, and leveraging powerful libraries within the Python ecosystem.

In summary, we believe that autoFISH will lower the barrier to entry for research groups seeking to perform sequential FISH. We believe that in-house solutions, such as the one presented here, are also well-suited to use alongside commercial approaches. The latter enables large-scale, high-multiplexed experiments that are highly challenging. Appropriate analysis of these and other data (e.g., scRNAseq) can then identify a smaller set of target genes. With in-house solutions, these genes can be probed at substantially reduced cost. Lastly, autoFISH is developed as Open Source, and we hope that it will spark community effort and be expanded to include more protocols and modules.

Methods

This section provides an overview of the methods. For detailed, up-to-date information, readers

are referred to the following resources: <https://github.com/fish-quant/autofish> (for the fluidics system), <https://doi.org/10.5281/zenodo.17965896> (for experimental protocols), and <https://github.com/fish-quant/autofish-analysis> (analysis code).

Instrumentation and control

AutoFISH relies on an automated fluidics system for buffer exchanges. Python control software was developed to operate the fluidics system and communicate with the microscope to initiate acquisitions. AutoFISH is implemented as a lightweight Python package with minimal dependencies. It utilizes PySerial for communication with various hardware components, and its modular design enables the easy addition of new components, such as pumps or valves from different suppliers. AutoFISH can be controlled with a graphical user interface.

Microscopy

AutoFISH does not allow direct control of microscopes but provides several options for triggering image acquisition. This makes autoFISH highly flexible, as it can be used across various microscopes with minimal adaptation. At the time of publication, the following options are available, but new ones can be easily added:

Pycromanager (Pinkard et al., 2021) is suitable for microscopes controlled by Micromanager. The user can define a settings file with the relevant acquisition parameters and a position list. autoFISH will then create a multi-D acquisition event that enables imaging after each hybridization round.

Synchronization with a text file: either the file's content (e.g., a switch from 0 to 1) or its existence can be used to start an acquisition. Once the acquisition is complete, the file content can be set back to 0, or the file can be deleted, signaling to autoFISH that the next hybridization round can begin.

Communication with TTL trigger: This is suitable for microscopes that integrate incoming and outgoing TTL trigger events into their acquisition routines. In autoFISH, we use an Arduino to send and receive TTL triggers: a TTL trigger initiates an acquisition, and an incoming TTL trigger signals that the acquisition is terminated and that the next hybridization rounds can begin.

Fluidics system

We use a custom-built fluidics system combining elements from published approaches (merFISH and ORCA (Bintu et al., 2018; Moffitt and Zhuang, 2016)). Detailed building plans and part lists are on GitHub (<https://github.com/fish-quant/autofish>). The system consists of the following main components, which are widely available and affordable: a peristaltic pump to accurately control flow rate, a bubble trap connected to a vacuum pump to remove air bubbles, and a flow sensor to monitor flow and detect potential issues, such as obstructed fluidic lines or large air bubbles.

Buffers are stored either in deep-well 96-well plates (each well holding up to 2 mL) or in syringes. The former usually contains hybridization buffers used only once, and the syringes contain general-purpose buffers for washing and imaging. The plates are placed on a pipette robot built from a CNC milling machine. These robots can be readily adapted by incorporating a 3D-printed insert into the spindle holder to accommodate a needle connected to a fluidic line. The needle can aspirate liquid from the multi-well plates. A computer-controlled valve selects buffers from different inlets from the fluidic lines connected to the pipette robot or the syringes.

Sequential smFISH: instrumentation and microscopy

Detailed experimental protocols are available at <https://doi.org/10.5281/zenodo.17965896>.

In experiments on the fluidic system, we used the less-toxic Ethylene Carbonate (EC) instead of formamide, as previously described (Moffitt et al., 2016a). In most previous studies, hybridizations on the fluidics system were done without Dextran. For some experiments, such as signal amplification with SABER, we found that even low concentrations of dextran help reduce oligo concentrations, improve signal, and decrease background. We found that, with our peristaltic pump (Ismatec Reglo Digital), buffers containing up to 5% dextran can be used without detrimental effects on flow rates. It is essential to filter all buffers (using 0.22 μm filters) to prevent clogging of the fluidic lines. The imaging buffer contains enzymes. An isothermal chamber with a cold block maintains a low temperature throughout the experiment. In the syringe, a layer of mineral oil on top of the imaging buffer prevents oxygen from diffusing into the buffer.

Two microscopy systems were used in this study. Data for Figures 1, 2, and 4 were acquired on a Nikon Ti Eclipse equipped with a Lumencor Spectra X LED light source, a 60 $\times$ 1.4 NA objective, and a Prime BSI Express sCMOS camera, and dedicated monoband filters from Semrock (DAPI-5060C-OMF, Cy3-4040D-000, Cy5-4040D-000). The system was controlled via Micro-Manager, with automated acquisition through Pycro-Manager in autoFISH. Data for Figure 3, and Supplementary Figures 1 and 2 were acquired on a Leica Thunder (DMI8) with a Leica K8 camera, 613X 1.4 NA objective, Leica LED8 light source, Leica quad-band DFT51011 filter set with a motorized emission wheel containing 460/80, 535/70, 590/50, 642/80, and 100% open positions. The system was operated using Leica LAS X software with the trigger module, and automated acquisition was achieved via TTL signals exchanged between an Arduino and the microscope through the autoFISH program.

Sequential smFISH: oligos for RNA detection and fluorescent readout

Oligos against the target genes can be designed using several tools (Defard et al., 2024a). Here, we used Oligostan (Tsanov et al., 2016), a computational tool we built in-house for the bioinformatic design of RNA probes. The oligo synthesis method depends on the number of required oligos. We use oligo pools provided by Integrated DNA Technologies (IDT) at a concentration of 50 pmol/oligo for small-scale studies involving only a few hundred oligos. The pools can be used without amplification and are sufficient for several hundred hybridization

rounds when hybridizing 12mm coverslips. Individual pools can be ordered for each gene, enabling the custom assembly of hybridization panels tailored to specific biological questions. For larger oligo numbers, oPools that require amplification can be used (Moffitt and Zhuang, 2018).

In this study, we employed the strategy described by Bintu et al. (2018), which uses four oligos (Figure 1): primary oligos, readout oligos, imager oligos, and displacement oligos. For each gene, 20-50 primary oligos are designed, consisting of two parts. The first part is complementary to a portion of the target gene's RNA. The other part is an identical readout sequence placed on 3' and 5' of each oligo (provided in (Bintu et al., 2018)). All oligos targeting the same RNA share this barcode. Note that these primary oligos are not fluorescently labeled. A readout oligo with the complementary barcode sequence is then used to detect a specific RNA. This oligo is pre-hybridized with an imager oligo carrying two fluorophores. Each primary oligo is thus labeled with up to 4 fluorophores. This design is cost-effective because the same imager oligo can be used with different readout oligos. A third region on the readout oligo—the toehold—can be used to remove the fluorescent signal. Using a displacement oligo complementary to the toehold and readout sequence, the readout oligo can be efficiently stripped off, and the resulting oligo duplex can be washed out. In the experiment, all primary oligos are hybridized simultaneously during a primary incubation step, while readout oligos are hybridized sequentially to image each RNA species. Multiple RNA species can be imaged simultaneously using imager oligos conjugated to spectrally distinct fluorophores. Interestingly, displacement oligos to remove readout oligos from previous rounds can be hybridized simultaneously with new readout oligos. This allows for faster experiments.

We observed that some of the original readout sequences yielded lower signal intensities in RNA smFISH. We developed a simple strategy to systematically test their quality (Supplementary Figure 3). In short, we used a probe set against a housekeeping gene carrying a validated readout sequence (RO1). We then design an intermediate oligo with a 5' region with a reverse complementary sequence to RO1 and a 3' region with a reverse complementary sequence to the readout to be tested. The to-be-tested readout can then be hybridized with an imager oligo and visualized. This approach eliminates the need to synthesize a reference oPool for each new readout, but requires only a simple adaptor oligo. For an up-to-date list of validated readout and imager sequences, please refer to <https://doi.org/10.5281/zenodo.17965896>.

SABER provides a simple, programmable method for signal amplification (Kishi et al., 2019). For us, ten-fold amplification was successfully implemented without inducing large oligo aggregations. Specific hybridization rounds can be amplified and stripped by using the amplified SABER sequences instead of the imager sequences on the readout oligos. During SABER optimization, we observed abundant nonspecific dots arising from either nonspecific primary probes or amplicons, likely due to signal amplification. To reduce this background, we found that a 3-step hybridization protocol performs best when stringency is optimized at each step.

First, primary probes are hybridized on the bench, using a more stringent buffer (1X SSC, 30% formamide, 10% dextran) rather than our standard buffer (2X SSC, 20% formamide). Second, the amplicons are hybridized in the fluidics system for each round using 1xSSC, 20% EC, and 5% dextran. Lastly, the imaging oligos are hybridized on the fluidics system (2X SSC, 10% EC, 5% dextran). The melting temperature of these oligos is relatively low because they contain no

Guanine, and a more stringent buffer results in a loss of signal intensity. The low dextran concentration does not affect flow rates, allowing a reduction in the concentration of the imaging oligos.

Clearing

Sample clearing reduces background signal in smFISH by removing tissue autofluorescence and reducing off-target binding of probes to lipids and proteins (Moffitt et al., 2016a). Briefly, RNAs are first anchored to a polyacrylamide matrix using poly (A)-targeting oligos and/or an epoxide (GMA). Once the matrix has solidified, tissues are incubated overnight at 37°C in a Proteinase K solution (20mg/mL). The next day, tissues are washed, and hybridization of readout oligos is performed.

Sequential smFISH in HeLa cells

To validate the autoFISH workflow, HeLa cells (ATCC CCL-2) were cultured in DMEM (gibco 21063-029), supplemented with 10% FBS and 1% penicillin-streptomycin at 37°C with 5% CO₂. Cells were then fixed in 4% PFA and permeabilized with 70% ethanol at -20°C overnight. After rehydration, the samples were hybridized overnight with primary probes against *XPO1* and *KIF1C*. Following mounting on the appropriate support, the samples were installed on the fluidics system. A series of 20 alternating hybridizations and stripping cycles for *XPO1* and *KIF1C* was then performed on the fluidic system. This enabled monitoring of signal quality over time, observing probe dehybridization, and tracking its progression throughout the experiment.

Mouse lung tissue

Studies were performed in accordance with the European Community's (2010/63/UE) recommendations for the care and use of laboratory animals. Experimental procedures were specifically approved by the ethics committee of the Institut Curie CEEA-IC #118 (Authorization number APAFIS#5479-201605271 0291841 given by the National Authority) in compliance with international guidelines. Ten to sixteen weeks old C57Bl6/J female mice were housed in the Institut Curie animal facility in compliance with European ethical procedures. Before tissue processing, mice were euthanized by cervical dislocation. To validate the hybridization of several genes on a section of fixed and cleared tissue, autoFISH was performed on mouse lung tissue. Briefly, mouse lungs were inflated with cold 4% paraformaldehyde (PFA), fixed overnight at 4°C, cryopreserved in 30% sucrose, and embedded in optimal cutting temperature (OCT) compound before sectioning (Curras-Alonso et al., 2023).

Analysis: Sharpness measurements

3D DAPI image stacks were acquired and projected along the z-axis using maximum intensity projection. Nuclei were segmented from the projected images with CellPose (Stringer et al., 2020). For each nucleus, we determined its bounding box and computed the local contrast of every z-plane in the 3D images using Helml's and Scherer local contrast algorithm (Helml and Scherer, 2001; Tsanov et al., 2016) and Supplementary Note 1. For each nucleus, the plane with the highest focal

contrast was selected for subsequent analysis. Composite images were then created by including the optimal focal-plane image for each cell mask, and background pixels were filled with the median z-index across all cells. Finally, we constructed box plots summarizing the extracted focus measurements for all nuclei in each condition.

Analysis: Sequential smFISH

After completing the experiment, the first step in the analysis is to segment the acquired images into .tiff files, each containing the entire Z-stack for a field of view and a hybridization round. During the experiment, alignment between fields of view may drift from one round to the next. To correct this, we use the SimpleITK package (Beare et al., 2018), which employs image similarity metrics across rounds to achieve alignment.

For analysis of cell culture experiments, spots are detected using Big-FISH (Imbert et al., 2022). Big-Fish sharpens spot-like signals in the images with a Laplacian-of-Gaussian (LoG) filter and then detects RNA spots as the local maxima. For lung tissue experiments, spot detection was performed with the recent deep learning approach U-FISH (Xu et al., 2024). U-FISH replaces Big-FISH's LoG filter with a deep-learning U-Net signal amplifier. It enhances the true RNA spots signal and suppresses the artefact signal. This U-Net feature is particularly advantageous for complex tissues such as the lung. Spot quality metrics, such as signal-to-noise ratio, background, and spot intensity, were computed using Big-FISH (Imbert et al., 2022). Nuclei segmentation on DAPI and cell segmentation, when cell staining was present, are performed with Cellpose (Pachitariu and Stringer, 2022). Finally, all fields of view are stitched using the stitching ImageJ plugin (Preibisch et al., 2009). The final step is to generate the spatially resolved gene-by-cell expression matrix from the previously localized RNA spots. However, because cell boundary staining is not always available in our experiments, the cell expression matrix is computed using our recently published point cloud segmentation method, ComSeg (Defard et al., 2024b). To perform cell type calling, we calculated the cosine distance between the cell expression vectors and the median centroids of each cell type, as defined in the reference scRNA-seq data from (Curras-Alonso et al., 2023). These gene markers allow for the classification of five distinct cell types: Alveolar Macrophage (AM) with *Chil3*, Alveolar Type 1 (AT1) with *Rtn2*, Alveolar Type 2 (AT2) with *Lamp3*, and Epithelial Cells with *Pecam1* and *Ptpnb*. Cells are assigned to their nearest cell type cluster based on cosine distance. Given the limitations of our gene markers, not every cell can be classified. Therefore, we only classify cells with a cosine distance to the nearest cell-type centroid below 0.8 and with more than 10 detected RNAs. The analysis code is available at <https://github.com/fish-quant/autofish-analysis>.

Analysis: Image registration and RNA colocalization for SABER

To align images of the same gene across different rounds by creating binary spot maps (with a value of 1 at each spot coordinate) and applying phase cross-correlation to compute the optimal global shift (ΔX , ΔY). This shift is then applied to the spot coordinates of rounds (3 and 7) and (5 and 8) to precisely align them with their respective reference rounds 1 and 2. The second step aligns manually round 2 to the primary reference (round 1) using a Napari plugin (Gros et al., 2025), and the resulting translation is applied to all rounds in the second group (2, 5, and 8). This process achieves pixel-level precision by first using cell membranes for coarse alignment and then the signal present in all images for fine-tuning,

overcoming the limitations of automatic registration methods due to the lack of conserved RNA spot patterns and low membrane contrast. Finally, after achieving high-quality registration, we match gene pairs across rounds using the Hungarian algorithm to minimize inter-round distances. A pair of spots is considered colocalized if their separation is less than two pixels (≈ 206 nm). Colocalization is reported as the percentage of detected colocalized spots relative to the total number of spots identified in that round.

Statistics and Reproducibility

As this study focuses on technology development, biological replicates were not performed. The method's robustness was validated through high-throughput spatial sampling and automated sequential imaging. The statistical unit (n) is defined as individual RNA molecules/spots, nested within cells and fields of view (FOVs). Specific n values for cells and RNAs are provided in the figure legends. Data are presented as [e.g., Mean $\pm$ SD] or boxplots (median, interquartile range, and 1.5 IQR whiskers) as indicated. No data were excluded from analysis, and all representative images reflect consistent results across all acquired FOVs.

Code availability

Code to control the fluidics system is available at <https://github.com/fish-quant/autofish>. Code used for the analysis is available at <https://github.com/fish-quant/autofish-analysis>.

Data availability

The data that support the findings of this study are available on reasonable request from the corresponding author.

Competing interests

The authors declare no competing interests.

Acknowledgements

This work has received financial support through the Institut Pasteur (Single Cell Seed Grant), Agence Nationale de la Recherche (ANR) grants LUSTRA (ANR-19-CE14-0015-03) and TRANSFACT (ANR-19-CE12-0007-02), as well as (ANR-24-INBS-0005 FBI BIOGEN). Further support from by a government grant managed by the Agence Nationale de la Recherche under the France 2030 program, with the reference numbers ANR-24-EXCI-0001, ANR-24-EXCI-0002, ANR-24-EXCI-0003, ANR-24-EXCI-0004, ANR-24- EXCI-0005. T.D, F.M., and C.W.

acknowledge funding by Institut Pasteur. A.M. has received a PhD fellowship from the European Union's Horizon 2020 research and innovation program under the Marie Skłodowska-Curie grant agreement No 847718. H.L. is the recipient of a PhD fellowship from the International Student program from Paris-Saclay University. We would also like to thank Elric Esposito and Juliene Fernandez from the Photonic BioImaging platform (PB) at the Pasteur Institute for their help setting up the imaging experiments. We thank the reviewers for their constructive and detailed feedback. Their suggestions have helped us clarify several points and improve the manuscript's overall quality.

Author Contributions

C.W. performed wet-lab experiments and contributed to the development of the fluidics system. T.D. and J.B. performed data analysis. M.L. provided microscopy expertise and support. C.S. contributed to the development of the fluidics system. H.L. and A.M. provided mouse samples. M.I.G. contributed to the writing of the manuscript. J.-A.L.-V. and C.F. provided technical insights and supervised the mouse-related work. T.W. contributed to data analysis. F.M. conceived the study, implemented the fluidics system, and led the writing and coordination of the manuscript. All authors reviewed and approved the final version of the manuscript.

References

- Bahry, E., Breimann, L., Epstein, L., Kolyvanov, K., Harrington, K.I.S., Lionnet, T., Preibisch, S., 2021. RS-FISH: Precise, interactive and scalable smFISH spot detection using Radial Symmetry. <https://doi.org/10.1101/2021.03.09.434205>
- Beare, R., Lowekamp, B., Yaniv, Z., 2018. Image Segmentation, Registration and Characterization in R with SimpleITK. *J Stat Softw* 86, 8. <https://doi.org/10.18637/jss.v086.i08>
- Bintu, B., Mateo, L.J., Su, J.-H., Sinnott-Armstrong, N.A., Parker, M., Kinrot, S., Yamaya, K., Boettiger, A.N., Zhuang, X., 2018. Super-resolution chromatin tracing reveals domains and cooperative interactions in single cells. *Science* 362. <https://doi.org/10.1126/science.aau1783>
- Chen, K.H., Boettiger, A.N., Moffitt, J.R., Wang, S., Zhuang, X., 2015. RNA imaging. Spatially resolved, highly multiplexed RNA profiling in single cells. *Science* 348, aaa6090. <https://doi.org/10.1126/science.aaa6090>
- Cheng, M., Jiang, Y., Xu, J., Mentis, A.-F.A., Wang, S., Zheng, H., Sahu, S.K., Liu, L., Xu, X., 2023. Spatially resolved transcriptomics: a comprehensive review of their technological advances, applications, and challenges. *J Genet Genomics* 50, 625–640. <https://doi.org/10.1016/j.jgg.2023.03.011>
- Choi, H.M.T., Schwarzkopf, M., Fornace, M.E., Acharya, A., Artavanis, G., Stegmaier, J., Cunha, A., Pierce, N.A., 2018. Third-generation in situ hybridization chain reaction: multiplexed, quantitative, sensitive, versatile, robust. *Development* 145, dev165753. <https://doi.org/10.1242/dev.165753>
- Chouaib, R., Safieddine, A., Pichon, X., Imbert, A., Kwon, O.S., Samacoits, A., Traboulsi, A.-M., Robert, M.-C., Tsanov, N., Coleno, E., Poser, I., Zimmer, C., Hyman, A., Le Hir, H., Zibara, K., Peter, M., Mueller, F., Walter, T., Bertrand, E., 2020. A Dual Protein-mRNA Localization Screen Reveals Compartmentalized Translation and Widespread Co-

- translational RNA Targeting. *Dev Cell* 54, 773–791.e5.
<https://doi.org/10.1016/j.devcel.2020.07.010>
- Codeluppi, S., Borm, L.E., Zeisel, A., La Manno, G., van Lunteren, J.A., Svensson, C.I., Linnarsson, S., 2018. Spatial organization of the somatosensory cortex revealed by osmFISH. *Nat. Methods* 15, 932–935. <https://doi.org/10.1038/s41592-018-0175-z>
- Cui, Y., Yang, G., Goodwin, D.R., O’Flanagan, C.H., Sinha, A., Zhang, C., Kitko, K.E., Shin, T.W., Park, D., Aparicio, S., Boyden, E.S., 2023. Expansion microscopy using a single anchor molecule for high-yield multiplexed imaging of proteins and RNAs. *PLoS One* 18, e0291506. <https://doi.org/10.1371/journal.pone.0291506>
- Curras-Alonso, S., Soulier, J., Defard, T., Weber, C., Heinrich, S., Laporte, H., Leboucher, S., Lameiras, S., Dutreix, M., Favaudon, V., Massip, F., Walter, T., Mueller, F., Londoño-Vallejo, J.-A., Fouillade, C., 2023. An interactive murine single-cell atlas of the lung responses to radiation injury. *Nat Commun* 14, 2445. <https://doi.org/10.1038/s41467-023-38134-z>
- Defard, T., Desrentes, A., Fouillade, C., Mueller, F., 2024a. A DIY guide for image-based spatial transcriptomic: TLS as a case example. *bioRxiv* 2024–07.
- Defard, T., Laporte, H., Ayan, M., Soulier, J., Curras-Alonso, S., Weber, C., Massip, F., Londoño-Vallejo, J.-A., Fouillade, C., Mueller, F., Walter, T., 2024b. A point cloud segmentation framework for image-based spatial transcriptomics. *Commun Biol* 7, 823. <https://doi.org/10.1038/s42003-024-06480-3>
- Femino, A.M., Fay, F.S., Fogarty, K., Singer, R.H., 1998. Visualization of single RNA transcripts in situ. *Science* 280, 585–590.
- Gandin, V., Kim, J., Yang, L.-Z., Lian, Y., Kawase, T., Hu, A., Rokicki, K., Fleishman, G., Tillberg, P., Castrejon, A.A., Stringer, C., Preibisch, S., Liu, Z.J., 2025. Deep-tissue transcriptomics and subcellular imaging at high spatial resolution. *Science* eadq2084. <https://doi.org/10.1126/science.adq2084>
- Gros, A., Vanaret, J., Dunsing-Eichenauer, V., Rostan, A., Roudot, P., Lenne, P.-F., Guignard, L., Tlili, S., 2025. A quantitative pipeline for whole-mount deep imaging and analysis of multi-layered organoids across scales. *eLife* 14. <https://doi.org/10.7554/eLife.107154.1>
- Helmli, F.S., Scherer, S., 2001. Adaptive shape from focus with an error estimation in light microscopy, in: ISPA 2001. Proceedings of the 2nd International Symposium on Image and Signal Processing and Analysis. In Conjunction with 23rd International Conference on Information Technology Interfaces (IEEE Cat. Presented at the ISPA 2001 the 2nd International Symposium on Image and Signal Processing and Analysis. In conjunction with 23rd International Conference on Information Technology Interfaces (IEEE Cat., pp. 188–193. <https://doi.org/10.1109/ISPA.2001.938626>
- Imbert, A., Ouyang, W., Safieddine, A., Coleno, E., Zimmer, C., Bertrand, E., Walter, T., Mueller, F., 2022. FISH-quant v2: a scalable and modular tool for smFISH image analysis. *RNA* 28, 786–795. <https://doi.org/10.1261/rna.079073.121>
- Kishi, J.Y., Lapan, S.W., Beliveau, B.J., West, E.R., Zhu, A., Sasaki, H.M., Saka, S.K., Wang, Y., Cepko, C.L., Yin, P., 2019. SABER amplifies FISH: enhanced multiplexed imaging of RNA and DNA in cells and tissues. *Nat Methods* 16, 533–544. <https://doi.org/10.1038/s41592-019-0404-0>
- Lubeck, E., Coskun, A.F., Zhiyentayev, T., Ahmad, M., Cai, L., 2014. Single-cell in situ RNA profiling by sequential hybridization. *Nat. Methods* 11, 360–361. <https://doi.org/10.1038/nmeth.2892>
- Mah, C.K., Ahmed, N., Lopez, N.A., Lam, D.C., Pong, A., Monell, A., Kern, C., Han, Y., Prasad, G., Cesnik, A.J., Lundberg, E., Zhu, Q., Carter, H., Yeo, G.W., 2024. Bento: a toolkit for subcellular analysis of spatial transcriptomics data. *Genome Biol* 25, 82. <https://doi.org/10.1186/s13059-024-03217-7>
- Mateo, L.J., Sinnott-Armstrong, N., Boettiger, A.N., 2021. Tracing DNA paths and RNA profiles

- in cultured cells and tissues with ORCA. *Nat Protoc* 16, 1647–1713.
<https://doi.org/10.1038/s41596-020-00478-x>
- Moffitt, J.R., Hao, J., Bambah-Mukku, D., Lu, T., Dulac, C., Zhuang, X., 2016a. High-performance multiplexed fluorescence in situ hybridization in culture and tissue with matrix imprinting and clearing. *Proc. Natl. Acad. Sci. U.S.A.*
<https://doi.org/10.1073/pnas.1617699113>
- Moffitt, J.R., Hao, J., Wang, G., Chen, K.H., Babcock, H.P., Zhuang, X., 2016b. High-throughput single-cell gene-expression profiling with multiplexed error-robust fluorescence in situ hybridization. *Proc. Natl. Acad. Sci. U.S.A.* 113, 11046–11051.
<https://doi.org/10.1073/pnas.1612826113>
- Moffitt, J.R., Zhuang, X., 2018. RNA Imaging with MERFISH - Probe Construction.
- Moffitt, J.R., Zhuang, X., 2016. RNA Imaging with Multiplexed Error-Robust Fluorescence In Situ Hybridization (MERFISH). *Meth. Enzymol.* 572, 1–49.
<https://doi.org/10.1016/bs.mie.2016.03.020>
- Pachitariu, M., Stringer, C., 2022. Cellpose 2.0: how to train your own model. *Nat Methods* 1–8.
<https://doi.org/10.1038/s41592-022-01663-4>
- Pinkard, H., Stuurman, N., Ivanov, I.E., Anthony, N.M., Ouyang, W., Li, B., Yang, B., Tsuchida, M.A., Chhun, B., Zhang, G., Mei, R., Anderson, M., Shepherd, D.P., Hunt-Isaak, I., Dunn, R.L., Jahr, W., Kato, S., Royer, L.A., Thiagarajah, J.R., Eliceiri, K.W., Lundberg, E., Mehta, S.B., Waller, L., 2021. Pycro-Manager: open-source software for customized and reproducible microscope control. *Nat Methods* 18, 226–228. <https://doi.org/10.1038/s41592-021-01087-6>
- Preibisch, S., Saalfeld, S., Tomancak, P., 2009. Globally optimal stitching of tiled 3D microscopic image acquisitions. *Bioinformatics* 25, 1463–1465.
<https://doi.org/10.1093/bioinformatics/btp184>
- Rademacher, A., Huseynov, A., Bortolomeazzi, M., Wille, S.J., Schumacher, S., Sant, P., Keitel, D., Okonechnikov, K., Ghasemi, D.R., Pajtler, K.W., Mallm, J.-P., Rippe, K., 2024. Comparison of spatial transcriptomics technologies using tumor cryosections.
<https://doi.org/10.1101/2024.04.03.586404>
- Raj, A., van den Bogaard, P., Rifkin, S.A., van Oudenaarden, A., Tyagi, S., 2008. Imaging individual mRNA molecules using multiple singly labeled probes. *Nat Meth* 5, 877–879.
<https://doi.org/10.1038/nmeth.1253>
- Stringer, C., Wang, T., Michaelos, M., Pachitariu, M., 2020. Cellpose: a generalist algorithm for cellular segmentation. *Nat Methods*. <https://doi.org/10.1038/s41592-020-01018-x>
- Teng, H., Yuan, Y., Bar-Joseph, Z., 2022. Clustering spatial transcriptomics data. *Bioinformatics* 38, 997–1004. <https://doi.org/10.1093/bioinformatics/btab704>
- Traag, V.A., Waltman, L., van Eck, N.J., 2019. From Louvain to Leiden: guaranteeing well-connected communities. *Sci Rep* 9, 5233. <https://doi.org/10.1038/s41598-019-41695-z>
- Tsanov, N., Samacoits, A., Chouaib, R., Traboulsi, A.-M., Gostan, T., Weber, C., Zimmer, C., Zibara, K., Walter, T., Peter, M., Bertrand, E., Mueller, F., 2016. smiFISH and FISH-quant - a flexible single RNA detection approach with super-resolution capability. *Nucleic Acids Res.* 44, e165. <https://doi.org/10.1093/nar/gkw784>
- Xia, C., Babcock, H.P., Moffitt, J.R., Zhuang, X., 2019. Multiplexed detection of RNA using MERFISH and branched DNA amplification. *Sci Rep* 9, 7721.
<https://doi.org/10.1038/s41598-019-43943-8>
- Xu, W., Cai, H., Zhang, Q., Mueller, F., Ouyang, W., Cao, G., 2024. U-FISH: a universal deep learning approach for accurate FISH spot detection across diverse datasets.
<https://doi.org/10.1101/2024.03.06.583706>

Figures




Figure 1. Summary of the autoFISH workflow, from wetlab to analysis. **A** Schematic of the experimental workflow. Used oligonucleotides: Overview of primary, readout, imager, and stripping oligos and their nomenclature. Sequential FISH: Primary oligos are hybridized once; across N rounds, readout oligos reveal defined RNA subsets, while stripping oligos remove signals from previous rounds. SABER: Signal amplification is achieved via DNA concatemers providing multiple binding sites for fluorescent imagers. Clearing: RNA targets are anchored to a polymer, and the sample is cleared to reduce autofluorescence and nonspecific background. **B** Overview of the fluidics system with the different components. **C** Samples can be placed in an imaging chamber. **D** Python control software allows the fluidics system to initiate communication with the microscope. **E** Provided software enables the analysis of different hybridization rounds and yields RNA abundance measurements for each segmented cell. The figure was created with [BioRender.com](https://www.biorender.com/).




Figure 2. Across 20 hybridization rounds alternating *XPO1* and *KIF1C* detection, autoFISH ensures efficient probe stripping and consistent mRNA detection. A Images of the same cell with either *XPO1* and *KIF1C* for the respective first and last hybridization rounds. Please note that the nuclear background observed in *KIF1C* is due to nonspecific binding of

certain primary oligonucleotides. The same Intensity scaling [min, max] was used for both images of each gene: *XPO1* [374, 2216], *KIF1C* [2334, 5708]. **B** Zooms in (yellow rectangles in A) showing *XPO1* (round 1) and *KIF1C* (round 2). **C** RNA counts per cell across rounds for *XPO1* and *KIF1C*. **D** Example of image registration with images from round 1 (purple) and round 19 (green). **E** Percentage of colocalized RNAs with reference to the first respective hybridization round. Distance threshold to be considered co-localized = 0.5 μm **F-G-H** Mean Signal over noise ratio (SNR), background intensity, and spots intensity across rounds. Number of field of views = 25; number of cells = 453; number of RNAs: *KIF1C* = 305K, *XPO1* = 55K; percentile interval = 50%.





Figure 3. Signal amplification with SABER across hybridization rounds. **A** SABER amplifies a specific sequence by serving as a binding region for imaging oligos. Hybridization of these oligos leads to signal amplification. The oligo also contains a toehold region (red) that permits probe stripping. 2% agarose gel electrophoresis shows readout oligos with either 1 or 3 repeats, or SABER amplification with two durations (30 min and 2 h). Gel was visualized using a Fisher Bioblock Scientific UV imaging system. Full-gel image was captured on thermal paper and subsequently digitized. **B** Hybridization conditions for the eight rounds (R1-R8): small, filled dots indicate a standard FISH; larger dots indicate SABER FISH; x indicates stripping of the previous round. **C** Images of the same cell across different hybridization rounds. Images for the same gene are shown with the same intensity scaling. **D** RNA levels per cell across rounds. SABER runs are indicated with an asterisk. Boxplots: central line = median; box = Interquartile range (Q1–Q3); whiskers = max/min data points within 1.5× IQR. Outliers excluded. **E** Quantification of intensities of RNA and background across rounds. SABER runs are indicated with an asterisk. Boxplots: central line = median; box = Interquartile range (Q1–Q3); whiskers = max/min data points within 1.5× IQR. Outliers excluded. **F** Gene colocalization across imaging rounds. Colocalization of *XPO1* and *NFKB1A* detections against reference detections from round 1 (*XPO1*) and round 2 (*NFKB1A*): colocalization between the same gene is shown with a circle, between different genes with a star. These colocalization values for each condition (circle or stars) are calculated by pooling n=25 fields of view within the experiment. Small points represent colocalization coefficients for individual fields of view to indicate spatial variability. Note that the aggregated value differs in theory from the arithmetic mean of individual fields of view, as the ratio of sums is not mathematically equivalent to the mean of ratios.



4. Clearing reduces background and allows RNA detection in tissue. A Principle of

clearing. RNAs can be anchored with epoxide and/or Poly-T oligos carrying an acrydite moiety. The panel was created with BioRender.com. **B** Influence of epoxide on nuclear signals. DAPI and smFISH images after clearing with Poly-T anchoring and epoxide. **C** Nucleus segmentation and RNA spots detection across rounds. **D** Illustration of workflow from RNA detection to cell segmentation and cell type classification. The yellow rectangle shows a region provided with a zoom-in. **E** Louvain clustering of the spatially resolved RNA profiles. Cell-type mapping based on our reference scRNAseq data by matching the cosine distance between cluster centroids.

Supplementary Material

autoFISH: a modular toolbox for sequential single-molecule RNA FISH experiments

Christian Weber^{1,2,*}, Thomas Defard^{1,2,3,4,5*}, Chloé Sturmach¹, Mickael Lelek⁶, Hugo Laporte^{6,7}, Ayan Mallick⁶, Maria Isabella Gariboldi^{1,2}, José-Arturo Londoño-Vallejo⁶, Thomas Walter^{3,4,5}, Charles Fouillade⁶, Jacques Bourg¹, Florian Mueller^{1,2,*,#}

¹ Institut Pasteur, Université Paris Cité, Photonic Bio-Imaging, Centre de Ressources et Recherches Technologiques (UTechS-PBI, C2RT), F-75015 Paris, France

² Institut Pasteur, Université Paris Cité, Imaging and Modeling Unit, F-75015 Paris, France

³ Centre for Computational Biology (CBIO), Mines Paris, PSL University, 75006, Paris, France

⁴ Institut Curie, PSL University, 75005, Paris, France

⁵ INSERM, U900, 75005, Paris, France

⁶ Institut Curie, Inserm U1021-CNRS UMR 3347, University Paris-Saclay, PSL Research University, Centre Universitaire, Orsay Cedex, France.

⁷ Institute of Cell Biology (Cancer Research), University Hospital Essen, Essen, Germany.


Supplementary Figure 1

Supplementary Figure 2

Supplementary Figure 3

Supplementary Note 1

Supplementary Figure 1




SABER - impact of repeated binding sites and number of fluorophores on single RNA signal intensity

- **1X-2F**: standard FISH with a single binding site; visualized with imager oligos with two fluorophores.
- **3X-1F / 3X-2F**: SABER with three binding sites (ordered from IDT); visualized with imager oligos with either one or two fluorophores.
- **30'-1F / 30'-2F**: SABER with enzymatically amplified sequences (30 min for 8-10 binding sites); visualized with imager oligos with either one or two fluorophores.

A Images showing a typical field of view with the same intensity rescaling for each row. The upper row shows a rescaling for the 3X SABER, the lower row for the 30' SABER. Scale bars are 20 μ m.

B Optimization of signal intensity via multivalent and dual-labeled oligonucleotides. Comparison of single RNA spot intensity distributions. Boxplots: central line = median; box = Interquartile range(Q1–Q3); whiskers = max/min data points within 1.5 $\times$ IQR. Outliers excluded.

Supplementary Figure 2




Impact of clearing and anchoring strategy on nuclear DAPI signal.

A Composite images for three different fields of view, without epoxyde (GMA), with epoxyde, and with epoxyde and poly-T anchor oligos. Each nucleus is displayed at the plane in which the average Helmi and Scherer focus metric is maximal. Shared color and spatial scale across all images. Scale bar is 40 μm .

B Maximal mean focus distribution across cells of the field of view. Boxplots: central line = median; box = Interquartile range(Q1–Q3); whiskers = max/min data points within 1.5 $\times$ IQR. Outliers excluded.

Supplementary Figure 3



Evaluation of redesigned readout (RO) sequences using intermediary oligonucleotides

A Schematic of the different oligonucleotides used to test new RO sequences. The use of an intermediary oligonucleotide enables the reuse of the same primary-RO oligo pool.

B Comparison of FISH signal intensities used to validate the new RO sequences. Scale bar is 10 µm.

- **First:** RO1, reference smFISH without intermediary oligo (positive control).
- **Second:** smFISH with intermediary oligo, *retained* barcode (RO15).
- **Third:** smFISH with intermediary oligo, *removed* barcode (RO16).