## **Instrument A Calibration Data**

\*\*\*\*\*\*\*\*\*\*\*\*\*\*\*\*\*\*

Traveler Number \_\_\_\_\_

Traveler Operation(s) \_\_\_\_\_ Traveler Page(s) \_\_\_\_\_

Document Title

Nominal and Actual Calibration Data Instrument A 10/5/2017 12:10

Equipment Used: equipment notes

NOTE: calibration notes

32766.98

Min and Max PLC Nominal Counts: 0.00 32767.00 Min and Max Nominal EU (units): 0.00 100.00

Measured Counts Measured EU (units)
----0.00 0.00
8191.00 90.50
16383.00 128.00
24574.90 156.76

181.01

The least squares fit 2 degree polynomial is: -1.5035037E-07x^2 + 0.010154652x + 5.4224937

Measured EU	Calcd EU	Error	% of EU Range
0.00	5.42	5.422	5.422%
90.50	78.51	-11.988	-11.988%
128.00	131.43	3.432	3.432%
156.76	164.17	7.412	7.412%
181.01	176.73	-4.278	-4.278%

Calibrated engineering units for the min and max PLC counts are as follows:

EU at min and max PLC Counts: 5.4225 176.7323

## **Instrument A Calibration Data**

Compensate for a non-zero count value at zero EU. Shift the curve fit up or down by the count value of the zero EU value (the x-intercept of EU axis). The adjusted count values vs EU values are:

Measured EU (units)	
0.00	
90.50	
128.00	
156.76	
181.01	
	0.00 90.50 128.00 156.76

The least squares fit 2 degree polynomial for the adjusted counts is:  $-1.5035037E-07x^2 + 0.010313973x + -1.5003897E-06$ 

Calibrated engineering units for the adjusted min and max PLC counts are as follows:

EU at min and max PLC Counts: -0.0000 176.5303

Mfg Sign/Date	
QA Sign/Date	
*****	**************

## Nominal and Actual Calibration Curves Instrument A 10/5/2017 12:10

