



SRI LANKA ATOMIC ENERGY REGULATORY COUNCIL

Ministry of Power & Renewable Energy

LICENCE

Issued under Sri Lanka Atomic Energy Act No.40 of 2014



For the Medical Radiography Using Ionizing Radiation

1. Licence No. : HT/07/A/MR-01/L/01/2022
2. Effective date : 01.01.2022
3. Date of expiry : 31.12.2023
4. Name of the applicant : Proprietor
5. Address of the institute where the machine is used : Arogya Channeling Center, No. 22, Tekkawatta Upper Road, Tangalle.
6. Name of the radiation protection officer : K.K. Siripala
7. Personnel authorized to operate the under mentioned equipment : K.K. Siripala
8. Personnel authorized to work in the controlled area of the facility : G.G. Krishantha Laknath
9. X-ray equipment authorized to be used :

Location of the Equipment	Type	Make	Model No.	Tube Serial No.	Max. tube Voltage/ Current (kVp/mAs)	Nature of Authorization
X-ray Room	Static	Toshiba	DR-1603	7L0050	125/320	Use & Possession

Pursuant to the section 22 of the Sri Lanka Atomic Energy Act No. 40 of 2014 and Atomic Energy Safety Regulations No. 1 of 1999 and in reliance on statements and representation made by the applicant, a licence is hereby issued to the above applicant to possess the above irradiating apparatus for the diagnostic purposes at the place(s) and to use these equipment by the persons, authorized as indicated above. This licence is subject to all applicable regulations and Codes of Practices of the Sri Lanka Atomic Energy Regulatory Council (Council) now or hereafter will be in effect and to the conditions specified in Annexure I.

IT IS THE RESPONSIBILITY OF THE LICENCED HOLDER TO BE FULLY AWARE OF ALL RESPONSIBILITIES AND OBLIGATIONS IMPOSED BY THE SRI LANKA ATOMIC ENERGY ACT NO. 40 OF 2014 AND ATOMIC ENERGY SAFETY REGULATIONS NO. 1 OF 1999 UNDER WHICH THIS LICENCE IS ISSUED.

THE LICENCE IS NOT VALID FOR IMPORTATION OF ANY TYPE OF X-RAY MACHINES AND THE LICENCE IS NOT TRANSFERRABLE.

Date: 06.12.2021

Director General

Sri Lanka Atomic Energy Regulatory Council

H. L. Anil Ranjith

Director General / CEO

Sri Lanka Atomic Energy Regulatory Council




SRI LANKA ATOMIC ENERGY ACT NO. 40 OF 2014



LICENCE NOTICE

Arogya Channeling Center, No. 22, Tekkawatta Upper Road, Tangalle has been authorized by the Sri Lanka Atomic Energy HT/07/A/MR-01/L/01/2022 to possess and use a Diagnostic X-ray machine for human diagnostic purposes under the conditions stipulated in the licence till 31.12.2023.


Director General

Sri Lanka Atomic Energy Regulatory Council

H. L. Anil Ranjith

Director General / CEO

Sri Lanka Atomic Energy Regulatory Council

Date: 06.12.2021



SRI LANKA ATOMIC ENERGY REGULATORY COUNCIL

CONDITIONS OF LICENCE No. HT/07/A/MR-01/L/01/2022



1. The licensee must comply with all relevant provisions of the Sri Lanka Atomic Energy Act No. 40 of 2014 and Atomic Energy Safety Regulations No. 1 of 1999.
2. The licensee must use only the X-ray equipment specified in the licence and only for the purpose specified in the licence.
3. X-ray equipment should only be operated by the personnel authorized in this licence for operating the X-ray machines.
4. Any alterations to the facility buildings should not be carried out without prior approval of the Council.
5. If the X-ray facility is shifted to another location (for different address), this licence will not be valid for use of the machine and new licence for new location shall be obtained.
6. If the X-ray machine is removed to another room within the same premises, a modification to the exist licence shall be obtained.
7. The X-ray equipment, which is licenced with the Council, shall not be sold or transferred without prior approval of the Council.
8. Any X-ray equipment which are not included in this licence shall not be used without prior approval of the Council.
9. Modification to the licence shall be obtained prior to use of new X-ray tubes/machines, use of new workers including Radiation Protection Officer (RPO).
10. The licensee shall ensure the safety and security of the X-ray equipment.
11. The licensee must ensure that the X-ray equipment are repaired and serviced only by the persons having proper training on repair & servicing of X-ray machines.
12. The licensee shall notify and apply for renewal of the licence along with the application processing fee of Rs.2000/= not less than three months prior to the date of the expiry of the licence issued.
13. All personnel who operate the machines/work in the controlled areas shall be provided with a personal dosimeter approved by the Council to ascertain their radiation exposures and records of exposures should be maintained for an inspection by regulatory inspectors.
14. Any over exposures of workers due to unexpected occurrences shall immediately be notified to the Council and a report indicating findings of the over exposures including statements from the exposed persons should be sent to the Council within a week.
15. The licensee shall arrange an annual health surveillance program for workers listed under section 6, 7 & 8 of this licence to meet the requirements of regulation No. 42 of the Ionizing Radiation Protection Regulations gazetted on 28th July 2000.
16. In any case, Radiation Protection Officer or worker/s authorized in this licence are unable to perform duties entrusted to them by Sri Lanka Atomic Energy Act No. 40 of 2014 and Ionizing Radiation Protection Regulations 1999; suitably qualified persons should be appointed and Council should be informed immediately along with the modification application (Mod-01) to obtain an amendment to the existing licence.
17. Information contained in the licence shall not be given to any third party without prior approval of the Council.

H. L. Anil Ranjith

Director General / CEO

Director General

Date: 06.12.2021

Sri Lanka Atomic Energy Regulatory Council

Sri Lanka Atomic Energy Regulatory Council