



Filename: 2021-01-11 - 15.10.33.D5000



Default image (Contrast 100%)

Sample Info


| Well | Conc. [ng/ul] | Sample Description | Alert | Observations |
|------|---------------|---------------------|-------|--------------------------------------------|
| A1 | 48.0 | Ladder | ⚠ | Caution! Expired ScreenTape device; Ladder |
| B1 | 86.5 | E2 pochistone test | ⚠ | Caution! Expired ScreenTape device |
| C1 | 61.6 | C28 pochistone test | ⚠ | Caution! Expired ScreenTape device |
| D1 | 56.2 | C32 pochistone test | ⚠ | Caution! Expired ScreenTape device |

A1: Ladder**Sample Table**

| Well | Conc. [ng/μl] | Sample Description | Alert | Observations |
|------|---------------|--------------------|-------|--------------------------------------------|
| A1 | 48.0 | Ladder | ⚠ | Caution! Expired ScreenTape device; Ladder |

Peak Table

| Size [bp] | Calibrated Conc. [ng/μl] | Assigned Conc. [ng/μl] | Peak Molarity [nmol/l] | % Integrated Area | Peak Comment | Observations |
|-----------|--------------------------|------------------------|------------------------|-------------------|--------------|--------------|
| 15 | 6.48 | - | 665 | - | | Lower Marker |
| 100 | 4.56 | - | 70.1 | 9.50 | | |
| 250 | 5.09 | - | 31.3 | 10.60 | | |
| 400 | 5.32 | - | 20.5 | 11.09 | | |
| 600 | 5.58 | - | 14.3 | 11.64 | | |
| 1000 | 5.61 | - | 8.63 | 11.69 | | |
| 1500 | 5.53 | - | 5.67 | 11.52 | | |
| 2500 | 5.19 | - | 3.20 | 10.82 | | |
| 3500 | 5.49 | - | 2.41 | 11.45 | | |
| 5000 | 5.61 | - | 1.73 | 11.70 | | |
| 10000 | 3.25 | 3.25 | 0.500 | - | | Upper Marker |


B1: E2 pochistone test**Sample Table**

| Well | Conc. [ng/μl] | Sample Description | Alert | Observations |
|------|---------------|--------------------|-------|------------------------------------|
| B1 | 86.5 | E2 pochistone test | ⚠ | Caution! Expired ScreenTape device |

Peak Table

| Size [bp] | Calibrated Conc. [ng/μl] | Assigned Conc. [ng/μl] | Peak Molarity [nmol/l] | % Integrated Area | Peak Comment | Observations |
|-----------|--------------------------|------------------------|------------------------|-------------------|--------------|--------------|
| 15 | 5.39 | - | 553 | - | | Lower Marker |
| 668 | 66.9 | - | 154 | 77.33 | | |
| 1054 | 18.4 | - | 26.9 | 21.33 | | |
| 10000 | 3.25 | 3.25 | 0.500 | - | | Upper Marker |
| 18581 | 1.16 | - | 0.0961 | 1.34 | | |

C1: C28 pochistone test




Sample Table

| Well | Conc. [ng/μl] | Sample Description | Alert | Observations |
|------|---------------|---------------------|-------|------------------------------------|
| C1 | 61.6 | C28 pochistone test | ⚠ | Caution! Expired ScreenTape device |

Peak Table

| Size [bp] | Calibrated Conc. [ng/μl] | Assigned Conc. [ng/μl] | Peak Molarity [nmol/l] | % Integrated Area | Peak Comment | Observations |
|-----------|--------------------------|------------------------|------------------------|-------------------|--------------|--------------|
| 15 | 5.53 | - | 567 | - | | Lower Marker |
| 678 | 43.6 | - | 99.0 | 70.82 | | |
| 2160 | 11.7 | - | 8.32 | 18.97 | | |
| 4179 | 6.29 | - | 2.31 | 10.21 | | |
| 10000 | 3.25 | 3.25 | 0.500 | - | | Upper Marker |

D1: C32 pochistone test**Sample Table**

| Well | Conc. [ng/μl] | Sample Description | Alert | Observations |
|------|---------------|---------------------|-------|------------------------------------|
| D1 | 56.2 | C32 pochistone test | ⚠ | Caution! Expired ScreenTape device |

Peak Table

| Size [bp] | Calibrated Conc. [ng/μl] | Assigned Conc. [ng/μl] | Peak Molarity [nmol/l] | % Integrated Area | Peak Comment | Observations |
|-----------|--------------------------|------------------------|------------------------|-------------------|--------------|--------------|
| 15 | 6.77 | - | 694 | - | | Lower Marker |
| 651 | 44.6 | - | 105 | 79.32 | | |
| 2201 | 8.30 | - | 5.80 | 14.77 | | |
| 3862 | 3.32 | - | 1.32 | 5.92 | | |
| 10000 | 3.25 | 3.25 | 0.500 | - | | Upper Marker |