

Winston-Lutz Analysis

Metadata:

Físico: Victor

Unidade: 6EX

Data: 22_03_2024

Winston-Lutz Analysis

Number of images: 4

Maximum 2D CAX->BB distance: 1.63mm

Median 2D CAX->BB distance: 0.48mm

Mean 2D CAX->BB distance: 0.73mm

Shift to iso: facing gantry, move BB: LEFT 0.39mm; OUT 0.14mm; UP 0.17mm

Gantry 3D isocenter diameter: 0.85mm (3/4 images considered)

Maximum Gantry RMS deviation (mm): 0.54mm

Maximum EPID RMS deviation (mm): 4.15mm






Gantry+Collimator 3D isocenter diameter: 0.85mm (3/4 images considered)

Collimator 2D isocenter diameter: 0.00mm (1/4 images considered)

Maximum Collimator RMS deviation (mm): 0.00

Couch 2D isocenter diameter: 1.24mm (2/4 images considered)

Maximum Couch RMS deviation (mm): 1.63




Metadata:

Físico: Victor

Unidade: 6EX

Data: 22_03_2024

Gantry images



Metadata:

Físico: Victor

Unidade: 6EX

Data: 22_03_2024

Couch images

