

Metadata:

Físico: Físico

Unidade: Versa

MTF10: 0.28825524678036224

Data: 05-07-2026

- CBCT/CT 503 QA Test -

- CTP 404 Results -

HU Linearity tolerance: 40

HU Linearity ROIs:

Air: -933.0, PMP: -407.5, LDPE: -349.0, Poly:

-313.5, Acrylic: -217.0, Delrin: -70.5, Teflon:

294.0

HU Passed?: False

Low contrast visibility: 2.03



Geometric Line Average (mm): 48.87

Geometry Passed?: False

Measured Slice Thickness (mm): 2.612

Slice Thickness Passed? False

HU Linearity (44)



Metadata:

Físico: Físico

Unidade: Versa

MTF10: 0.28825524678036224

Data: 05-07-2026



- CTP528 Results -

MTF 80% (lp/mm): 0.13

MTF 50% (lp/mm): 0.19

MTF 30% (lp/mm): 0.23

Spatial Resolution (34)



Metadata:

Físico: Físico

Unidade: Versa

MTF10: 0.28825524678036224

Data: 05-07-2026

- CTP486 Results -

Uniformity tolerance: 40

Uniformity ROIs: Top: -253.0, Right: -255.0, Bottom: -255.0, Left: -253.0, Center: -256.0

Uniformity index: 0.403


Integral non-uniformity: 0.0020

Uniformity Passed?: False


Max Noise Power frequency: 0.0666666

Average Noise Power: 0.204243902762

HU Uniformity (7)



Uniformity Profiles



Metadata:


Físico: Físico

Unidade: Versa

MTF10: 0.28825524678036224

Data: 05-07-2026

Side View



Metadata:

Físico: Físico

Unidade: Versa

MTF10: 0.28825524678036224

Data: 05-07-2026