

Winston-Lutz Analysis

Metadata:

Físico: Físico

Unidade: AL1

Data: 12-12-2024

Winston-Lutz Analysis

=====

Number of images: 10

Maximum 2D CAX->BB distance: 0.85mm

Median 2D CAX->BB distance: 0.59mm

Mean 2D CAX->BB distance: 0.55mm

Shift to iso: facing gantry, move BB: RIGHT 0.10mm; IN 0.01mm; DOWN 0.12mm

Gantry 3D isocenter diameter: 0.28mm (3/10 images considered)

Maximum Gantry RMS deviation (mm): 0.74mm

Maximum EPID RMS deviation (mm): 0.25mm








Gantry+Collimator 3D isocenter diameter: 0.50mm (6/10 images considered)

Collimator 2D isocenter diameter: 0.20mm (3/10 images considered)

Maximum Collimator RMS deviation (mm): 0.72

Couch 2D isocenter diameter: 1.47mm (4/10 images considered)

Maximum Couch RMS deviation (mm): 0.85




Metadata:

Físico: Físico

Unidade: AL1

Data: 12-12-2024

Gantry images




Metadata:

Físico: Físico

Unidade: AL1

Data: 12-12-2024

Collimator images



Metadata:

Físico: Físico

Unidade: AL1

Data: 12-12-2024

Couch images

