



Certificate No.: CER : QA-18-2017-1

Date :

CER : QA-18-2017-1

10-Mar-18 Wika Item No

65897367

OA No. SO00020107

Customer : Cryolor Asia Pacific Pvt. Ltd.

Customer Item No AC000495

Line no 2

[illegible]

Prepared By:

.....
Ajay

S. N. Patil



Remark: 1 The Instruments are calibrated & found with in specified accuracy limits.

2 All Gauges are checked for over range protection 130% of F.S.& Found OK

Approved By:

(Quality Assurance)



Inspection Report (Calibration) - Pressure Gauges

Certificate No.: CER : QA-18-2017-1

Date :

10-Mar-18 Wika Item No

65687051

OA No. SO00020107

Customer : Cryolor Asia Pacific Pvt. Ltd.

Customer Item No AC000089

Line no 1

Device Under Test-712.15.160+212.20.100			Master test Gauge (Direct Reading)											
Accuracy : ±2.5% Connection :Through Swagelok adaptor			Range : 0 to 1 bar & 0 to 40 bar					Serial No. : 1158138 & 1158177		Accuracy :±0.1%				
Pos.	Range mbar & bar	Serial No./ Tag No.	Rising Pressure (mbar & bar)					Falling Pressure (mbar & bar)						
		0 to 100 %	0.00	20.00	40.00	80.00	100.00		0.00	20.00	40.00	80.00	100.00	
		mbar	0.00	125.80	251.60	503.20	629.00		0.00	125.80	251.60	503.20	629.00	
		bar	0.00	5.00	10.00	20.00	25.00		0.00	5.00	10.00	20.00	25.00	
1	0 to 629 mbar 0 to 25 bar	9000BHMx	0.00	122.00	244.00	504.00	630.00		0.00	120.00	242.00	502.00	630.00	
			0.00	5.00	10.00	20.00	25.10		0.00	4.90	9.90	20.00	25.10	
2	"	9000BHMY	0.00	122.00	244.00	504.00	632.00		0.00	120.00	242.00	504.00	632.00	
			0.00	5.00	10.10	20.10	25.10		0.00	4.90	10.00	20.00	25.10	
3	"	9000BHMZ	0.00	122.00	246.00	506.00	632.00		0.00	120.00	244.00	504.00	632.00	
			0.00	5.00	10.00	20.10	25.10		0.00	4.90	10.00	20.00	25.10	

Prepared By:

Ajay

S. N. Patil



Remark: 1 The Instruments are calibrated & found with in specified accuracy limits.

2 All Gauges are checked for over range protection 130% of F.S.& Found OK

Approved By:

(Quality Assurance)