

Calibration

Measured values

pH buffer 1:	TEC_4.000	165.6 mV	24.6 °C
pH buffer 2:	TEC_7.000	-8.2 mV	24.3 °C

Calibration data

Slope:	98.0 % / -58.0 mV/pH (a)
Zero point:	pH 6.87 / -8.1 mV (a)
Date and time:	07/26/19 8:40:38

Electrode data

Name:	A 162 2M-DIN-ID	Batch ID:	C180706002
-------	-----------------	-----------	------------

Device information

Device: TitroLine 7000
Serial number: 10012653
Software version: 1.18.1121.247