

## Calibration

### Measured values

pH buffer 1:	TEC_4.000	165.8 mV	25.7 °C
pH buffer 2:	TEC_7.000	-9.0 mV	26.0 °C

### Calibration data

Slope:	98.3 % / -58.2 mV/pH (a)
Zero point:	pH 6.84 / -9.0 mV (a)
Date and time:	08/09/19 9:02:57

### Electrode data

Name:	A 162 2M-DIN-ID	Batch ID:	C180706002
-------	-----------------	-----------	------------

### Device information

Device: TitroLine 7000  
Serial number: 10012653  
Software version: 1.18.1121.247