

## Calibration

### Measured values

pH buffer 1:	TEC_4.000	164.6 mV	24.5 °C
pH buffer 2:	TEC_7.000	-9.4 mV	24.5 °C

### Calibration data

Slope:	98.2 % / -58.1 mV/pH (a)
Zero point:	pH 6.85 / -9.4 mV (a)
Date and time:	08/10/19 8:52:48

### Electrode data

Name:	A 162 2M-DIN-ID	Batch ID:	C180706002
-------	-----------------	-----------	------------

### Device information

Device: TitroLine 7000  
Serial number: 10012653  
Software version: 1.18.1121.247