

Calibration

Measured values

pH buffer 1:	TEC_4.000	165.5 mV	25.4 °C
pH buffer 2:	TEC_7.000	-8.8 mV	25.2 °C

Calibration data

Slope:	98.1 % / -58.1 mV/pH (a)
Zero point:	pH 6.85 / -8.8 mV (a)
Date and time:	08/01/19 8:08:41

Electrode data

Name:	A 162 2M-DIN-ID	Batch ID:	C180706002
-------	-----------------	-----------	------------

Device information

Device: TitroLine 7000
Serial number: 10012653
Software version: 1.18.1121.247