

## Calibration

### Measured values

pH buffer 1:	TEC_4.000	165.5 mV	24.1 °C
pH buffer 2:	TEC_7.000	-8.0 mV	24.3 °C

### Calibration data

Slope:	98.0 % / -58.0 mV/pH (a)
Zero point:	pH 6.87 / -7.9 mV (a)
Date and time:	08/03/19 8:37:22

### Electrode data

Name:	A 162 2M-DIN-ID	Batch ID:	C180706002
-------	-----------------	-----------	------------

### Device information

Device: TitroLine 7000  
Serial number: 10012653  
Software version: 1.18.1121.247