

## Calibration

### Measured values

pH buffer 1:	TEC_4.000	166.0 mV	26.4 °C
pH buffer 2:	TEC_7.000	-8.7 mV	25.1 °C

### Calibration data

Slope:	97.7 % / -57.8 mV/pH (a)
Zero point:	pH 6.86 / -8.1 mV (a)
Date and time:	07/26/19 8:44:30

### Electrode data

Name:	A 162 2M-DIN-ID	Batch ID:	C180706002
-------	-----------------	-----------	------------

### Device information

Device: TitroLine 7000  
Serial number: 10012653  
Software version: 1.18.1121.247