

Calibration

Measured values

pH buffer 1:	TEC_4.000	165.6 mV	23.7 °C
pH buffer 2:	TEC_7.000	-8.2 mV	23.0 °C

Calibration data

Slope:	98.2 % / -58.1 mV/pH (a)
Zero point:	pH 6.87 / -7.9 mV (a)
Date and time:	08/05/19 8:33:20

Electrode data

Name:	A 162 2M-DIN-ID	Batch ID:	C180706002
-------	-----------------	-----------	------------

Device information

Device: TitroLine 7000
Serial number: 10012653
Software version: 1.18.1121.247