

## Calibration

### Measured values

|              |           |          |         |
|--------------|-----------|----------|---------|
| pH buffer 1: | TEC_4.000 | 165.6 mV | 25.2 °C |
| pH buffer 2: | TEC_7.000 | -9.3 mV  | 25.5 °C |

### Calibration data

|                |                          |
|----------------|--------------------------|
| Slope:         | 98.5 % / -58.3 mV/pH (a) |
| Zero point:    | pH 6.84 / -9.3 mV (a)    |
| Date and time: | 08/08/19 8:59:42         |

### Electrode data

|       |                 |           |            |
|-------|-----------------|-----------|------------|
| Name: | A 162 2M-DIN-ID | Batch ID: | C180706002 |
|-------|-----------------|-----------|------------|

### Device information

Device: TitroLine 7000  
Serial number: 10012653  
Software version: 1.18.1121.247