






MTF — Br36f (ROI=92 px)

Ensemble image — (W/L = 400/-949)



— Acquisition —
Patient ID : RZ_Fan_W50_W25_Body
Study date : 20230203
Scanner : Siemens Healthineers
Model : NAEOTOM Alpha
Station : CT5T-1018148
Kernel : Br36f
rFOV : 50 mm
Pixel size : 0.0488 mm

— Metrics —
FWHM (PSF) : 15.869 mm
f_peak : 0.000 1/mm
eta_peak : 1.000
f50 : 0.271 1/mm
f10 : 0.510 1/mm
f02 : 0.643 1/mm
Background : -1007.3 HU
PSF peak : 116.0 HU (above bg)

— Computation —
 Δf (freq) : 0.01395 1/mm
 Δx (space) : 0.12207 mm
ind_max : (184, 184)
Wire (x,y) : (511, 543)

— Source —
Z:\eb028936\CTPro_Images\images\
Phan_Wire\W_25um\Siemens\Alpha\
MayoClinic_5T_20230203_AdultBody\
SR_12mGy\T3D-Br36f_0.6_0.6_50