IQC CORRECTIVE ACTION FORM

(Recording schedule – Whenever there is IQC outlier OR Westgard Rules violation)

Date	Test Parameter	OUTLIER Daily QC Out / Monthly % CV high Or Westgard Rule	EQPT	Root Cause – Please tick the reason	Corrective Action Taken	Preventive Action Taken	Taken By (Sign of Tech)	Verified By
8/5/26	BHCG	Daily QC Out	Maglumi X3	1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault	Random error. QC - Repeat. Correct result obtained.		P 8/5/26	
				1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault				

Acceptance Criteria: Daily QC value within range specified by manufacturer for that control OR Westgard rule violation.
Monthly % CV: < MU OR 5 % for Hematology and < 10 % for Biochemistry whichever is higher.

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L1 CON	FSH	68630	0	6.77 mIU/mL		2026-05-08 09:58:00
BIORAD L1 CON	LH	9035	0	1.66 mIU/mL		2026-05-08 09:17:49
BIORAD L1 CON	PRL	125160	0	6.39 ng/ml		2026-05-08 09:12:22
BIORAD L1 CON	E2	1319369	0	75.97 pg/mL		2026-05-08 09:51:20
BIORAD L1 CON	T3	571932	0	0.62 ng/mL		2026-05-08 09:54:01
BIORAD L1 CON	T4	136071	0	5.50 ug/dl		2026-05-08 09:00:47
BIORAD L1 CON	TEST	6624951	0	111.5 ng/dl		2026-05-08 09:39:57
BIORAD L1 CON	HCG/B-HCG	24460	0	11.80 mIU/mL		2026-05-08 09:41:31
BIORAD L1 CON	TSH	23860	0	0.43 uIU/mL		2026-05-08 09:36:50

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L1 CON	HCG/B-HCG	7078	0	8.95 mIU/mL		2026-05-08 12:01:37