IQC CORRECTIVE ACTION FORM

(Recording schedule – Whenever there is IQC outlier OR Westgard Rules violation)

Date	Test Parameter	OUTLIER Daily QC Out / Monthly % CV high Or Westgard Rule	EQPT	Root Cause – Please tick the reason	Corrective Action Taken	Preventive Action Taken	Taken By (Sign of Tech)	Verified By
9/7/26	T4	Daily QC Out	Maglumi X3	1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault	Random error. QC - Repeat. Correct result obtained.		P 9/7/26	
				1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault				

Acceptance Criteria: Daily QC value within range specified by manufacturer for that control OR Westgard rule violation.
Monthly % CV: < MU OR 5 % for Hematology and < 10 % for Biochemistry whichever is higher.

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L1 CON	FSH	74322	0	7.26 mIU/mL		2026-07-09 09:28:57
BIORAD L1 CON	LH	8330	0	1.52 mIU/mL		2026-07-09 09:14:11
BIORAD L1 CON	PRL	139147	0	6.86 ng/ml		2026-07-09 09:50:11
BIORAD L1 CON	E2	1270956	0	83.59 pg/mL		2026-07-09 09:08:08
BIORAD L1 CON	T3	548750	0	0.67 ng/mL		2026-07-09 09:14:36
BIORAD L1 CON	T4	110055	0	8.45 ug/dl		2026-07-09 09:31:41
BIORAD L1 CON	TEST	6631314	0	121.5 ng/dl		2026-07-09 09:18:02
BIORAD L1 CON	HCG/B-HCG	7318	0	8.41 mIU/mL		2026-07-09 09:59:48
BIORAD L1 CON	TSH	25084	0	0.48 uIU/mL		2026-07-09 09:00:36

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L1 CON	T4	80099	0	6.15 ug/dl		2026-07-09 11:49:41

IQC CORRECTIVE ACTION FORM

(Recording schedule – Whenever there is IQC outlier OR Westgard Rules violation)

Date	Test Parameter	OUTLIER Daily QC Out / Monthly % CV high Or Westgard Rule	EQPT	Root Cause – Please tick the reason	Corrective Action Taken	Preventive Action Taken	Taken By (Sign of Tech)	Verified By
24/6/26	BHCG	Daily QC Out	Maglumi X3	1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault	Random error. QC - Repeat. Correct result obtained.		P 24/6/26	
				1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault				

Acceptance Criteria: Daily QC value within range specified by manufacturer for that control OR Westgard rule violation.
Monthly % CV: < MU OR 5 % for Hematology and < 10 % for Biochemistry whichever is higher.

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L2 CON	FSH	84115	0	23.60 mIU/mL		2026-06-24 09:55:17
BIORAD L2 CON	LH	124698	0	23.13 mIU/mL		2026-06-24 09:14:28
BIORAD L2 CON	PRL	244716	0	20.07 ng/ml		2026-06-24 09:52:07
BIORAD L2 CON	E2	436788	0	298.3 pg/mL		2026-06-24 09:55:42
BIORAD L2 CON	T3	365822	0	1.80 ng/mL		2026-06-24 09:01:33
BIORAD L2 CON	T4	56516	0	10.06 ug/dl		2026-06-24 09:28:27
BIORAD L2 CON	TEST	153794	0	292.8 ng/dl		2026-06-24 09:07:20
BIORAD L2 CON	HCG/B-HCG	62372	0	31.44 mIU/mL		2026-06-24 09:37:30
BIORAD L2 CON	TSH	261400	0	5.35 uIU/mL		2026-06-24 09:38:07

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L2 CON	HCG/B-HCG	45252	0	22.81 mIU/mL		2026-06-24 11:55:30

IQC CORRECTIVE ACTION FORM

(Recording schedule – Whenever there is IQC outlier OR Westgard Rules violation)

Date	Test Parameter	OUTLIER Daily QC Out / Monthly % CV high Or Westgard Rule	EQPT	Root Cause – Please tick the reason	Corrective Action Taken	Preventive Action Taken	Taken By (Sign of Tech)	Verified By
12/6/26	T3	Daily QC Out	Maglumi X3	1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault	Random error. QC - Repeat. Correct result obtained.		P 12/6/26	
				1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault				

Acceptance Criteria: Daily QC value within range specified by manufacturer for that control OR Westgard rule violation.
Monthly % CV: < MU OR 5 % for Hematology and < 10 % for Biochemistry whichever is higher.

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L2 CON	FSH	91758	0	25.98 mIU/mL		2026-06-12 09:34:27
BIORAD L2 CON	LH	137842	0	26.01 mIU/mL		2026-06-12 09:59:57
BIORAD L2 CON	PRL	274904	0	21.85 ng/ml		2026-06-12 09:32:23
BIORAD L2 CON	E2	445039	0	330.1 pg/mL		2026-06-12 09:05:34
BIORAD L2 CON	T3	277829	0	2.53 ng/mL		2026-06-12 09:47:56
BIORAD L2 CON	T4	51907	0	11.47 ug/dl		2026-06-12 09:24:09
BIORAD L2 CON	TEST	144713	0	305.8 ng/dl		2026-06-12 09:58:25
BIORAD L2 CON	HCG/B-HCG	32737	0	22.77 mIU/mL		2026-06-12 09:53:57
BIORAD L2 CON	TSH	295453	0	5.94 uIU/mL		2026-06-12 09:50:35

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L2 CON	T3	200960	0	1.83 ng/mL		2026-06-12 12:05:56

IQC CORRECTIVE ACTION FORM

(Recording schedule – Whenever there is IQC outlier OR Westgard Rules violation)

Date	Test Parameter	OUTLIER Daily QC Out / Monthly % CV high Or Westgard Rule	EQPT	Root Cause – Please tick the reason	Corrective Action Taken	Preventive Action Taken	Taken By (Sign of Tech)	Verified By
10/6/26	TESTO	Daily QC Out	Maglumi X3	1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault	Random error. QC - Repeat. Correct result obtained.		P 10/6/26	
				1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault				

Acceptance Criteria: Daily QC value within range specified by manufacturer for that control OR Westgard rule violation.
Monthly % CV: < MU OR 5 % for Hematology and < 10 % for Biochemistry whichever is higher.

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L2 CON	FSH	94328	0	27.53 mIU/mL		2026-06-10 09:42:42
BIORAD L2 CON	LH	144308	0	27.22 mIU/mL		2026-06-10 09:12:10
BIORAD L2 CON	PRL	283561	0	22.47 ng/ml		2026-06-10 09:36:49
BIORAD L2 CON	E2	452195	0	342.6 pg/mL		2026-06-10 09:33:07
BIORAD L2 CON	T3	308283	0	1.90 ng/mL		2026-06-10 09:07:17
BIORAD L2 CON	T4	50471	0	11.94 ug/dl		2026-06-10 09:51:44
BIORAD L2 CON	TEST	147326	0	414.0 ng/dl		2026-06-10 09:11:28
BIORAD L2 CON	HCG/B-HCG	32491	0	23.5 mIU/mL		2026-06-10 09:47:12
BIORAD L2 CON	TSH	307632	0	6.18 uIU/mL		2026-06-10 09:40:19

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L2 CON	TEST	106758	0	300.0 ng/dl		2026-06-10 11:29:28

IQC CORRECTIVE ACTION FORM

(Recording schedule – Whenever there is IQC outlier OR Westgard Rules violation)

Date	Test Parameter	OUTLIER Daily QC Out / Monthly % CV high Or Westgard Rule	EQPT	Root Cause – Please tick the reason	Corrective Action Taken	Preventive Action Taken	Taken By (Sign of Tech)	Verified By
10/6/26	FSH	Daily QC Out	Maglumi X3	1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault	Random error. QC - Repeat. Correct result obtained.		P 10/6/26	
				1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault				

Acceptance Criteria: Daily QC value within range specified by manufacturer for that control OR Westgard rule violation.
Monthly % CV: < MU OR 5 % for Hematology and < 10 % for Biochemistry whichever is higher.

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L1 CON	FSH	41923	0	10.32 mIU/mL		2026-06-10 09:01:52
BIORAD L1 CON	LH	8869	0	1.61 mIU/mL		2026-06-10 09:06:41
BIORAD L1 CON	PRL	123274	0	6.08 ng/ml		2026-06-10 09:44:35
BIORAD L1 CON	E2	1343661	0	76.44 pg/mL		2026-06-10 09:47:49
BIORAD L1 CON	T3	562829	0	0.60 ng/mL		2026-06-10 09:30:56
BIORAD L1 CON	T4	137422	0	5.42 ug/dl		2026-06-10 09:03:12
BIORAD L1 CON	TEST	6702477	0	108 ng/dl		2026-06-10 09:47:35
BIORAD L1 CON	HCG/B-HCG	6670	0	7.52 mIU/mL		2026-06-10 09:52:23
BIORAD L1 CON	TSH	23593	0	0.42 uIU/mL		2026-06-10 09:03:26

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L1 CON	FSH	30630	0	7.54 mIU/mL		2026-06-10 11:19:52

IQC CORRECTIVE ACTION FORM

(Recording schedule – Whenever there is IQC outlier OR Westgard Rules violation)

Date	Test Parameter	OUTLIER Daily QC Out / Monthly % CV high Or Westgard Rule	EQPT	Root Cause – Please tick the reason	Corrective Action Taken	Preventive Action Taken	Taken By (Sign of Tech)	Verified By
12/5/26	PRL	Daily QC Out	Maglumi X3	1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault	Random error. QC - Repeat. Correct result obtained.		P 12/5/26	
				1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault				

Acceptance Criteria: Daily QC value within range specified by manufacturer for that control OR Westgard rule violation.
Monthly % CV: < MU OR 5 % for Hematology and < 10 % for Biochemistry whichever is higher.

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L2 CON	FSH	83732	0	23.97 mIU/mL		2026-05-12 09:22:42
BIORAD L2 CON	LH	122260	0	23.66 mIU/mL		2026-05-12 09:25:47
BIORAD L2 CON	PRL	483724	0	29.07 ng/ml		2026-05-12 09:53:27
BIORAD L2 CON	E2	439650	0	312.9 pg/mL		2026-05-12 09:48:48
BIORAD L2 CON	T3	334950	0	1.80 ng/mL		2026-05-12 09:46:21
BIORAD L2 CON	T4	55916	0	10.60 ug/dl		2026-05-12 09:34:20
BIORAD L2 CON	TEST	147778	0	292.5 ng/dl		2026-05-12 09:57:49
BIORAD L2 CON	HCG/B-HCG	31530	0	22.21 mIU/mL		2026-05-12 09:36:46
BIORAD L2 CON	TSH	276962	0	5.60 uIU/mL		2026-05-12 09:41:26

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L2 CON	PRL	350105	0	21.04 ng/ml		2026-05-12 12:11:27

IQC CORRECTIVE ACTION FORM

(Recording schedule – Whenever there is IQC outlier OR Westgard Rules violation)

Date	Test Parameter	OUTLIER Daily QC Out / Monthly % CV high Or Westgard Rule	EQPT	Root Cause – Please tick the reason	Corrective Action Taken	Preventive Action Taken	Taken By (Sign of Tech)	Verified By
8/5/26	BHCG	Daily QC Out	Maglumi X3	1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault	Random error. QC - Repeat. Correct result obtained.		P 8/5/26	
				1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault				

Acceptance Criteria: Daily QC value within range specified by manufacturer for that control OR Westgard rule violation.
Monthly % CV: < MU OR 5 % for Hematology and < 10 % for Biochemistry whichever is higher.

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L1 CON	FSH	68630	0	6.77 mIU/mL		2026-05-08 09:58:00
BIORAD L1 CON	LH	9035	0	1.66 mIU/mL		2026-05-08 09:17:49
BIORAD L1 CON	PRL	125160	0	6.39 ng/ml		2026-05-08 09:12:22
BIORAD L1 CON	E2	1319369	0	75.97 pg/mL		2026-05-08 09:51:20
BIORAD L1 CON	T3	571932	0	0.62 ng/mL		2026-05-08 09:54:01
BIORAD L1 CON	T4	136071	0	5.50 ug/dl		2026-05-08 09:00:47
BIORAD L1 CON	TEST	6624951	0	111.5 ng/dl		2026-05-08 09:39:57
BIORAD L1 CON	HCG/B-HCG	24460	0	11.80 mIU/mL		2026-05-08 09:41:31
BIORAD L1 CON	TSH	23860	0	0.43 uIU/mL		2026-05-08 09:36:50

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L1 CON	HCG/B-HCG	7078	0	8.95 mIU/mL		2026-05-08 12:01:37

IQC CORRECTIVE ACTION FORM

(Recording schedule – Whenever there is IQC outlier OR Westgard Rules violation)

Date	Test Parameter	OUTLIER Daily QC Out / Monthly % CV high Or Westgard Rule	EQPT	Root Cause – Please tick the reason	Corrective Action Taken	Preventive Action Taken	Taken By (Sign of Tech)	Verified By
23/4/26	TSH	Daily QC Out	Maglumi X3	1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault	Random error. QC - Repeat. Correct result obtained.		P 23/4/26	
				1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault				

Acceptance Criteria: Daily QC value within range specified by manufacturer for that control OR Westgard rule violation.
Monthly % CV: < MU OR 5 % for Hematology and < 10 % for Biochemistry whichever is higher.

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L1 CON	FSH	84597	0	8.19 mIU/mL		2026-04-23 09:00:48
BIORAD L1 CON	LH	9614	0	1.79 mIU/mL		2026-04-23 09:26:18
BIORAD L1 CON	PRL	156039	0	7.54 ng/ml		2026-04-23 09:29:08
BIORAD L1 CON	E2	1215994	0	94.67 pg/mL		2026-04-23 09:37:47
BIORAD L1 CON	T3	523615	0	0.72 ng/mL		2026-04-23 09:34:01
BIORAD L1 CON	T4	124048	0	6.73 ug/dl		2026-04-23 09:33:39
BIORAD L1 CON	TEST	6399632	0	137 ng/dl		2026-04-23 09:16:45
BIORAD L1 CON	HCG/B-HCG	8397	0	9.28 mIU/mL		2026-04-23 09:41:08
BIORAD L1 CON	TSH	34330	0	0.68 uIU/mL		2026-04-23 09:42:14

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L1 CON	TSH	25242	0	0.50 ulU/mL		2026-04-23 12:00:14

IQC CORRECTIVE ACTION FORM

(Recording schedule – Whenever there is IQC outlier OR Westgard Rules violation)

Date	Test Parameter	OUTLIER Daily QC Out / Monthly % CV high Or Westgard Rule	EQPT	Root Cause – Please tick the reason	Corrective Action Taken	Preventive Action Taken	Taken By (Sign of Tech)	Verified By
10/4/26	T3	Daily QC Out	Maglumi X3	1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault	Random error. QC - Repeat. Correct result obtained.		P 10/4/26	
				1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault				

Acceptance Criteria: Daily QC value within range specified by manufacturer for that control OR Westgard rule violation.
Monthly % CV: < MU OR 5 % for Hematology and < 10 % for Biochemistry whichever is higher.

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L1 CON	FSH	81211	0	7.93 mIU/mL		2026-04-10 09:29:05
BIORAD L1 CON	LH	11169	0	2.02 mIU/mL		2026-04-10 09:40:29
BIORAD L1 CON	PRL	153442	0	7.45 ng/ml		2026-04-10 09:40:52
BIORAD L1 CON	E2	1257624	0	91.23 pg/mL		2026-04-10 09:53:09
BIORAD L1 CON	T3	599904	0	0.96 ng/mL		2026-04-10 09:00:58
BIORAD L1 CON	T4	126849	0	6.54 ug/dl		2026-04-10 09:52:00
BIORAD L1 CON	TEST	6361765	0	133.2 ng/dl		2026-04-10 09:26:00
BIORAD L1 CON	HCG/B-HCG	8292	0	9.16 mIU/mL		2026-04-10 09:23:54
BIORAD L1 CON	TSH	25064	0	0.51 uIU/mL		2026-04-10 09:29:33

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L1 CON	T3	437430	0	0.70 ng/mL		2026-04-10 11:18:58

IQC CORRECTIVE ACTION FORM

(Recording schedule – Whenever there is IQC outlier OR Westgard Rules violation)

Date	Test Parameter	OUTLIER Daily QC Out / Monthly % CV high Or Westgard Rule	EQPT	Root Cause – Please tick the reason	Corrective Action Taken	Preventive Action Taken	Taken By (Sign of Tech)	Verified By
27/3/26	FSH	Daily QC Out	Maglumi X3	1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault	Random error. QC - Repeat. Correct result obtained.		P 27/3/26	
				1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault				

Acceptance Criteria: Daily QC value within range specified by manufacturer for that control OR Westgard rule violation.
Monthly % CV: < MU OR 5 % for Hematology and < 10 % for Biochemistry whichever is higher.

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L2 CON	FSH	9416	0	16.74 mIU/mL		2026-03-27 09:03:26
BIORAD L2 CON	LH	134041	0	26.19 mIU/mL		2026-03-27 09:55:52
BIORAD L2 CON	PRL	276684	0	21.9 ng/ml		2026-03-27 09:45:07
BIORAD L2 CON	E2	447772	0	339.9 pg/mL		2026-03-27 09:14:57
BIORAD L2 CON	T3	335298	0	1.94 ng/mL		2026-03-27 09:34:01
BIORAD L2 CON	T4	50984	0	11.56 ug/dl		2026-03-27 09:28:14
BIORAD L2 CON	TEST	147809	0	315.6 ng/dl		2026-03-27 09:27:19
BIORAD L2 CON	HCG/B-HCG	32534	0	24.11 mIU/mL		2026-03-27 09:26:40
BIORAD L2 CON	TSH	289494	0	5.84 uIU/mL		2026-03-27 09:26:08

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L2 CON	FSH	14242	0	25.32 mIU/mL		2026-03-27 11:21:26

IQC CORRECTIVE ACTION FORM

(Recording schedule – Whenever there is IQC outlier OR Westgard Rules violation)

Date	Test Parameter	OUTLIER Daily QC Out / Monthly % CV high Or Westgard Rule	EQPT	Root Cause – Please tick the reason	Corrective Action Taken	Preventive Action Taken	Taken By (Sign of Tech)	Verified By
11/3/26	E2	Daily QC Out	Maglumi X3	1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault	Random error. QC - Repeat. Correct result obtained.		P 11/3/26	
				1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault				

Acceptance Criteria: Daily QC value within range specified by manufacturer for that control OR Westgard rule violation.
Monthly % CV: < MU OR 5 % for Hematology and < 10 % for Biochemistry whichever is higher.

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L1 CON	FSH	72724	0	7.15 mIU/mL		2026-03-11 09:18:40
BIORAD L1 CON	LH	9749	0	1.80 mIU/mL		2026-03-11 09:27:14
BIORAD L1 CON	PRL	137047	0	6.82 ng/ml		2026-03-11 09:27:39
BIORAD L1 CON	E2	1004063	0	118.4 pg/mL		2026-03-11 09:00:23
BIORAD L1 CON	T3	537118	0	0.68 ng/mL		2026-03-11 09:32:24
BIORAD L1 CON	T4	133138	0	5.95 ug/dl		2026-03-11 09:40:02
BIORAD L1 CON	TEST	6434944	0	120.9 ng/dl		2026-03-11 09:10:42
BIORAD L1 CON	HCG/B-HCG	7394	0	8.44 mIU/mL		2026-03-11 09:32:53
BIORAD L1 CON	TSH	25397	0	0.48 uIU/mL		2026-03-11 09:15:48

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L1 CON	E2	730999	0	86.2 pg/mL		2026-03-11 11:18:23

IQC CORRECTIVE ACTION FORM

(Recording schedule – Whenever there is IQC outlier OR Westgard Rules violation)

Date	Test Parameter	OUTLIER Daily QC Out / Monthly % CV high Or Westgard Rule	EQPT	Root Cause – Please tick the reason	Corrective Action Taken	Preventive Action Taken	Taken By (Sign of Tech)	Verified By
5/3/26	TSH	Daily QC Out	Maglumi X3	1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault	Random error. QC - Repeat. Correct result obtained.		P 5/3/26	
				1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault				

Acceptance Criteria: Daily QC value within range specified by manufacturer for that control OR Westgard rule violation.
Monthly % CV: < MU OR 5 % for Hematology and < 10 % for Biochemistry whichever is higher.

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L2 CON	FSH	92467	0	26.72 mIU/mL		2026-03-05 09:23:44
BIORAD L2 CON	LH	143106	0	28.01 mIU/mL		2026-03-05 09:07:41
BIORAD L2 CON	PRL	289786	0	23.5 ng/ml		2026-03-05 09:06:44
BIORAD L2 CON	E2	453359	0	357.6 pg/mL		2026-03-05 09:33:32
BIORAD L2 CON	T3	324153	0	2.12 ng/mL		2026-03-05 09:44:05
BIORAD L2 CON	T4	47220	0	12.16 ug/dl		2026-03-05 09:00:12
BIORAD L2 CON	TEST	147460	0	337.6 ng/dl		2026-03-05 09:04:11
BIORAD L2 CON	HCG/B-HCG	34745	0	25.13 mIU/mL		2026-03-05 09:29:51
BIORAD L2 CON	TSH	386000	0	7.89 uIU/mL		2026-03-05 09:04:42

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L2 CON	TSH	281305	0	5.75 ulU/mL		2026-03-05 11:22:42

IQC CORRECTIVE ACTION FORM

(Recording schedule – Whenever there is IQC outlier OR Westgard Rules violation)

Date	Test Parameter	OUTLIER Daily QC Out / Monthly % CV high Or Westgard Rule	EQPT	Root Cause – Please tick the reason	Corrective Action Taken	Preventive Action Taken	Taken By (Sign of Tech)	Verified By
27/2/26	PRL	Daily QC Out	Maglumi X3	1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault	Random error. QC - Repeat. Correct result obtained.		P 27/2/26	
				1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault				

Acceptance Criteria: Daily QC value within range specified by manufacturer for that control OR Westgard rule violation.
Monthly % CV: < MU OR 5 % for Hematology and < 10 % for Biochemistry whichever is higher.

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L1 CON	FSH	73984	0	7.2 mIU/mL		2026-02-27 09:12:38
BIORAD L1 CON	LH	9801	0	1.84 mIU/mL		2026-02-27 09:14:47
BIORAD L1 CON	PRL	126000	0	9.69 ng/ml		2026-02-27 09:39:13
BIORAD L1 CON	E2	1265073	0	82.57 pg/mL		2026-02-27 09:37:24
BIORAD L1 CON	T3	571017	0	0.68 ng/mL		2026-02-27 09:32:45
BIORAD L1 CON	T4	129121	0	5.90 ug/dl		2026-02-27 09:54:38
BIORAD L1 CON	TEST	6385618	0	117.8 ng/dl		2026-02-27 09:30:55
BIORAD L1 CON	HCG/B-HCG	7514	0	8.45 mIU/mL		2026-02-27 09:44:11
BIORAD L1 CON	TSH	25981	0	0.49 uIU/mL		2026-02-27 09:27:16

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L1 CON	PRL	91932	0	7.07 ng/ml		2026-02-27 11:57:13

IQC CORRECTIVE ACTION FORM

(Recording schedule – Whenever there is IQC outlier OR Westgard Rules violation)

Date	Test Parameter	OUTLIER Daily QC Out / Monthly % CV high Or Westgard Rule	EQPT	Root Cause – Please tick the reason	Corrective Action Taken	Preventive Action Taken	Taken By (Sign of Tech)	Verified By
21/2/26	TSH	Daily QC Out	Maglumi X3	1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault	Random error. QC - Repeat. Correct result obtained.		P 21/2/26	
				1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault				

Acceptance Criteria: Daily QC value within range specified by manufacturer for that control OR Westgard rule violation.
Monthly % CV: < MU OR 5 % for Hematology and < 10 % for Biochemistry whichever is higher.

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L1 CON	FSH	74743	0	7.23 mIU/mL		2026-02-21 09:08:55
BIORAD L1 CON	LH	9484	0	1.83 mIU/mL		2026-02-21 09:18:12
BIORAD L1 CON	PRL	135183	0	6.93 ng/ml		2026-02-21 09:00:14
BIORAD L1 CON	E2	1278921	0	83.94 pg/mL		2026-02-21 09:47:28
BIORAD L1 CON	T3	560839	0	0.67 ng/mL		2026-02-21 09:12:29
BIORAD L1 CON	T4	128720	0	6.15 ug/dl		2026-02-21 09:53:42
BIORAD L1 CON	TEST	6313979	0	116.3 ng/dl		2026-02-21 09:18:12
BIORAD L1 CON	HCG/B-HCG	7598	0	8.52 mIU/mL		2026-02-21 09:46:53
BIORAD L1 CON	TSH	34543	0	0.68 uIU/mL		2026-02-21 09:28:40

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L1 CON	TSH	25399	0	0.50 ulU/mL		2026-02-21 11:46:40

IQC CORRECTIVE ACTION FORM

(Recording schedule – Whenever there is IQC outlier OR Westgard Rules violation)

Date	Test Parameter	OUTLIER Daily QC Out / Monthly % CV high Or Westgard Rule	EQPT	Root Cause – Please tick the reason	Corrective Action Taken	Preventive Action Taken	Taken By (Sign of Tech)	Verified By
15/1/26	E2	Daily QC Out	Maglumi X3	1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault	Random error. QC - Repeat. Correct result obtained.		P 15/1/26	
				1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault				

Acceptance Criteria: Daily QC value within range specified by manufacturer for that control OR Westgard rule violation.
Monthly % CV: < MU OR 5 % for Hematology and < 10 % for Biochemistry whichever is higher.

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L2 CON	FSH	95607	0	27.44 mIU/mL		2026-01-15 09:09:48
BIORAD L2 CON	LH	147019	0	25.52 mIU/mL		2026-01-15 09:03:33
BIORAD L2 CON	PRL	290752	0	22.63 ng/ml		2026-01-15 09:15:06
BIORAD L2 CON	E2	775133	0	439.1 pg/mL		2026-01-15 09:53:37
BIORAD L2 CON	T3	321710	0	1.96 ng/mL		2026-01-15 09:45:25
BIORAD L2 CON	T4	47898	0	11.72 ug/dl		2026-01-15 09:35:32
BIORAD L2 CON	TEST	147140	0	321.2 ng/dl		2026-01-15 09:26:51
BIORAD L2 CON	HCG/B-HCG	33897	0	24.77 mIU/mL		2026-01-15 09:15:01
BIORAD L2 CON	TSH	305791	0	6.20 uIU/mL		2026-01-15 09:31:35

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L2 CON	E2	563829	0	319.4 pg/mL		2026-01-15 12:11:37

IQC CORRECTIVE ACTION FORM

(Recording schedule – Whenever there is IQC outlier OR Westgard Rules violation)

Date	Test Parameter	OUTLIER Daily QC Out / Monthly % CV high Or Westgard Rule	EQPT	Root Cause – Please tick the reason	Corrective Action Taken	Preventive Action Taken	Taken By (Sign of Tech)	Verified By
4/1/26	PRL	Daily QC Out	Maglumi X3	1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault	Random error. QC - Repeat. Correct result obtained.		P 4/1/26	
				1) Reagent expiry, deterioration 2) QC expiry, deterioration 3) Equipment flag, functioning etc 4) Environmental conditions 5) Procedure violation 6) Technician Fault				

Acceptance Criteria: Daily QC value within range specified by manufacturer for that control OR Westgard rule violation.
Monthly % CV: < MU OR 5 % for Hematology and < 10 % for Biochemistry whichever is higher.

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L1 CON	FSH	70886	0	6.97 mIU/mL		2026-01-04 09:54:44
BIORAD L1 CON	LH	10585	0	1.90 mIU/mL		2026-01-04 09:13:26
BIORAD L1 CON	PRL	125941	0	9.69 ng/ml		2026-01-04 09:42:49
BIORAD L1 CON	E2	1295069	0	78.28 pg/mL		2026-01-04 09:48:59
BIORAD L1 CON	T3	559359	0	0.65 ng/mL		2026-01-04 09:05:21
BIORAD L1 CON	T4	133254	0	5.65 ug/dl		2026-01-04 09:43:05
BIORAD L1 CON	TEST	6295944	0	141.1 ng/dl		2026-01-04 09:08:29
BIORAD L1 CON	HCG/B-HCG	8340	0	9.40 mIU/mL		2026-01-04 09:05:39
BIORAD L1 CON	TSH	25093	0	0.46 uIU/mL		2026-01-04 09:59:30

Sample Result Report

Sample ID	Assay	RLU	RLU CV (%)	Conc.	Flag	Finish Time
BIORAD L1 CON	PRL	91889	0	7.07 ng/ml		2026-01-04 12:00:49