

Laboratory Instrument  
Register and Utilization,  
March–August 2026  
Loblolly Ridge Therapeutics

|                        |                                                                                                                                                                                                                                                                                                                                               |
|------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Effective date         | 2026-09-10                                                                                                                                                                                                                                                                                                                                    |
| Purpose                | Instrument register with maintenance data and weekly available, booked, used and down hours for 20 bookable instruments.                                                                                                                                                                                                                      |
| Coverage               | See each dataset dictionary and summary. Missing numeric values are excluded, not zero.                                                                                                                                                                                                                                                       |
| Summary interpretation | Descriptive statistics of selected records only. Totals of percentages and adjustment values are not operating KPIs.                                                                                                                                                                                                                          |
| Introduction           | Register of the 68 laboratory instruments and weekly utilization of the 20 bookable instruments from March 2 to August 28, 2026. Prepared by Lab Operations to support the bioanalytical capacity review.                                                                                                                                     |
| Contents               | Data 1 – instrument register: type, laboratory area, bookable flag, installation date, last preventive maintenance, maintenance interval, service vendor and status for all 68 instruments.                                                                                                                                                   |
| Contents               | Data 2 – instrument-week utilization: 520 records, one per bookable instrument per week, with capacity, available, booked, used and downtime hours and the number of runs.                                                                                                                                                                    |
| Hour definitions       | Available hours: scheduled hours the instrument could run in the week. For each of the two LC-MS/MS triple-quadrupole systems this includes the unattended overnight queue, about 100 hours a week; holiday weeks are lower.                                                                                                                  |
| Hour definitions       | Downtime hours: recorded breakdown, repair and maintenance time.                                                                                                                                                                                                                                                                              |
| Hour definitions       | Capacity hours: available hours minus downtime. Utilization is measured against capacity hours.                                                                                                                                                                                                                                               |
| Hour definitions       | Booked hours: hours reserved in the instrument calendar or the overnight queue.                                                                                                                                                                                                                                                               |
| Hour definitions       | Used hours: hours with a run logged on the instrument.                                                                                                                                                                                                                                                                                        |
| Utilization            | Utilization is not precomputed. Divide booked or used hours by capacity hours for the period you need. From June to August the two LC-MS/MS systems were booked for 91% and used for 84% of capacity, up from 74% booked and 67% used from March to May. The other 18 bookable instruments were used for 48% of capacity from June to August. |
| Downtime               | Recorded downtime across the 20 bookable instruments was 265 hours from March to August, including 58 hours on INS-002 in July for an ion-source repair spread over two weeks.                                                                                                                                                                |
| Maintenance            | Preventive maintenance intervals are 6 or 12 months by instrument type. Maintenance dates are copied from the Lab Operations spreadsheet and vendor portals, which remain the records of maintenance and calibration.                                                                                                                         |
| Not recorded here      | Booking-level user names and instrument purchase prices.                                                                                                                                                                                                                                                                                      |

| Dataset                        | Measure            | Unit   | Total     | Mean  | Records | Period start | Period end |
|--------------------------------|--------------------|--------|-----------|-------|---------|--------------|------------|
| Laboratory instrument register | pm_interval_months | months |           |       |         | 2014-01-01   | 2026-08-31 |
| Instrument-week utilization    | used_hours         | hours  | 756.00    | 11.12 | 68      | 2026-03-02   | 2026-08-28 |
| Instrument-week utilization    | available_hours    | hours  | 13,954.20 | 26.84 | 520     | 2026-03-02   | 2026-08-28 |
| Instrument-week utilization    | downtime_hours     | hours  | 27,412.50 | 52.72 | 520     | 2026-03-02   | 2026-08-28 |
| Instrument-week utilization    | capacity_hours     | hours  | 264.80    | 0.51  | 520     | 2026-03-02   | 2026-08-28 |
| Instrument-week utilization    | booked_hours       | hours  | 27,147.70 | 52.21 | 520     | 2026-03-02   | 2026-08-28 |
| Instrument-week utilization    | runs               | runs   | 15,230.00 | 29.29 | 520     | 2026-03-02   | 2026-08-28 |
|                                |                    |        | 4,626.00  | 8.90  | 520     |              |            |

| Dataset                        | Field              | Definition                      | Unit   | Type     | Reference |
|--------------------------------|--------------------|---------------------------------|--------|----------|-----------|
| Laboratory instrument register | instrument_id      | Instrument id                   |        | id       |           |
| Laboratory instrument register | pm_interval_months | Preventive-maintenance interval | months | integer  |           |
| Laboratory instrument register | instrument_type    | Instrument type                 |        | category |           |
| Laboratory instrument register | lab_area           | Lab area                        |        | category |           |
| Laboratory instrument register | bookable           | Bookable                        |        | category |           |
| Laboratory instrument register | install_date       | Install date                    |        | date     |           |
| Laboratory instrument register | last_pm            | Last pm                         |        | category |           |
| Laboratory instrument register | service_vendor     | Service vendor                  |        | category |           |
| Laboratory instrument register | status             | Status                          |        | category |           |
| Instrument-week utilization    | record_id          | Record id                       |        | id       |           |
| Instrument-week utilization    | used_hours         | Hours used                      | hours  | number   |           |
| Instrument-week utilization    | instrument_id      | Instrument id                   |        | category |           |
| Instrument-week utilization    | instrument_type    | Instrument type                 |        | category |           |
| Instrument-week utilization    | week_start         | Week start                      |        | date     |           |
| Instrument-week utilization    | available_hours    | Scheduled available hours       | hours  | number   |           |
| Instrument-week utilization    | downtime_hours     | Maintenance or breakdown hours  | hours  | number   |           |
| Instrument-week utilization    | capacity_hours     | Available minus downtime        | hours  | number   |           |
| Instrument-week utilization    | booked_hours       | Hours booked in the calendar    | hours  | number   |           |
| Instrument-week utilization    | runs               | Runs or sessions                | runs   | integer  |           |

| instrument_id | pm_interval_months | instrument_type              | lab_area                   | bookable | install_date | last_pm    | service_vendor | status     |
|---------------|--------------------|------------------------------|----------------------------|----------|--------------|------------|----------------|------------|
| INS-001       |                    | 6 LC-MS/MS triple quadrupole | bioanalytical & analytical | yes      | 2019-08-12   | 2026-07-23 | V-023          | in service |
| INS-002       |                    | 6 LC-MS/MS triple quadrupole | bioanalytical & analytical | yes      | 2022-03-07   | 2026-07-30 | V-023          | in service |
| INS-003       |                    | 6 high-resolution MS         | bioanalytical & analytical | yes      | 2025-04-17   | 2026-04-20 | V-023          | in service |
| INS-004       |                    | 6 UPLC                       | bioanalytical & analytical | yes      | 2021-12-30   | 2026-06-29 | V-024          | in service |
| INS-005       |                    | 6 UPLC                       | bioanalytical & analytical | yes      | 2017-05-10   | 2026-07-21 | V-024          | in service |
| INS-006       |                    | 6 UPLC                       | bioanalytical & analytical | yes      | 2022-04-06   | 2026-07-30 | V-024          | in service |
| INS-007       |                    | 12 plate reader              | assay development          | yes      | 2016-04-01   | 2025-12-05 | V-024          | in service |
| INS-008       |                    | 12 plate reader              | assay development          | yes      | 2022-06-11   | 2025-12-09 | V-024          | in service |
| INS-009       |                    | 12 plate reader              | assay development          | yes      | 2015-05-26   | 2026-05-26 | V-024          | in service |
| INS-010       |                    | 12 plate reader              | assay development          | yes      | 2016-12-14   | 2026-01-19 | V-024          | in service |
| INS-011       |                    | 6 liquid handler             | assay development          | yes      | 2016-03-08   | 2026-07-05 | V-025          | in service |
| INS-012       |                    | 6 liquid handler             | assay development          | yes      | 2020-09-26   | 2026-08-11 | V-025          | in service |
| INS-013       |                    | 6 liquid handler             | assay development          | yes      | 2021-04-15   | 2026-07-24 | V-025          | in service |
| INS-014       |                    | 6 flow cytometer             | cell biology               | yes      | 2018-08-12   | 2026-04-29 | V-024          | in service |
| INS-015       |                    | 12 automated imager          | cell biology               | yes      | 2016-08-21   | 2025-10-13 | V-024          | in service |
| INS-016       |                    | 12 automated imager          | cell biology               | yes      | 2015-10-21   | 2026-03-03 | V-024          | in service |
| INS-017       |                    | 12 qPCR                      | cell biology               | yes      | 2016-01-10   | 2026-05-22 | V-024          | in service |
| INS-018       |                    | 12 qPCR                      | cell biology               | yes      | 2015-01-26   | 2026-04-08 | V-024          | in service |
| INS-019       |                    | 12 HPLC                      | chemistry support          | yes      | 2023-10-28   | 2026-02-14 | V-024          | in service |
| INS-020       |                    | 12 HPLC                      | chemistry support          | yes      | 2025-03-01   | 2026-04-12 | V-024          | in service |
| INS-021       |                    | 12 biosafety cabinet         | cell biology               | no       | 2023-07-23   | 2026-03-06 | V-024          | in service |
| INS-022       |                    | 12 biosafety cabinet         | cell biology               | no       | 2018-10-08   | 2026-05-06 | V-024          | in service |
| INS-023       |                    | 12 biosafety cabinet         | cell biology               | no       | 2017-03-24   | 2026-07-30 | V-024          | in service |
| INS-024       |                    | 12 biosafety cabinet         | cell biology               | no       | 2018-12-19   | 2026-04-20 | V-024          | in service |
| INS-025       |                    | 12 biosafety cabinet         | cell biology               | no       | 2016-08-13   | 2026-07-20 | V-024          | in service |
| INS-026       |                    | 12 biosafety cabinet         | cell biology               | no       | 2025-07-05   | 2026-04-08 | V-024          | in service |
| INS-027       |                    | 12 biosafety cabinet         | cell biology               | no       | 2017-01-05   | 2025-11-21 | V-024          | in service |
| INS-028       |                    | 12 biosafety cabinet         | cell biology               | no       | 2017-12-11   | 2026-08-06 | V-024          | in service |
| INS-029       |                    | 12 biosafety cabinet         | cell biology               | no       | 2024-10-05   | 2026-03-03 | V-024          | in service |
| INS-030       |                    | 12 biosafety cabinet         | cell biology               | no       | 2025-07-24   | 2026-04-28 | V-024          | in service |

| record_id | used_hours | instrument_id | instrument_type            | week_start | available_hours | downtime_hours | capacity_hours | booked_hours | runs |    |
|-----------|------------|---------------|----------------------------|------------|-----------------|----------------|----------------|--------------|------|----|
| IW-0001   |            | 72.9 INS-001  | LC-MS/MS triple quadrupole | 2026-03-02 |                 | 100            | 0              | 100          | 75.9 | 8  |
| IW-0002   |            | 67.5 INS-001  | LC-MS/MS triple quadrupole | 2026-03-09 |                 | 100            | 0              | 100          | 77.7 | 8  |
| IW-0003   |            | 71.8 INS-001  | LC-MS/MS triple quadrupole | 2026-03-16 |                 | 100            | 0              | 100          | 78.6 | 8  |
| IW-0004   |            | 67.6 INS-001  | LC-MS/MS triple quadrupole | 2026-03-23 |                 | 100            | 0              | 100          | 75.5 | 8  |
| IW-0005   |            | 63.4 INS-001  | LC-MS/MS triple quadrupole | 2026-03-30 |                 | 90             | 1.7            | 88.3         | 69.3 | 7  |
| IW-0006   |            | 65 INS-001    | LC-MS/MS triple quadrupole | 2026-04-06 |                 | 100            | 0.7            | 99.3         | 71.7 | 7  |
| IW-0007   |            | 59.4 INS-001  | LC-MS/MS triple quadrupole | 2026-04-13 |                 | 100            | 0              | 100          | 67.8 | 7  |
| IW-0008   |            | 73.1 INS-001  | LC-MS/MS triple quadrupole | 2026-04-20 |                 | 100            | 0.3            | 99.7         | 76.6 | 8  |
| IW-0009   |            | 72.9 INS-001  | LC-MS/MS triple quadrupole | 2026-04-27 |                 | 100            | 0              | 100          | 75.5 | 8  |
| IW-0010   |            | 64.3 INS-001  | LC-MS/MS triple quadrupole | 2026-05-04 |                 | 100            | 0              | 100          | 71   | 7  |
| IW-0011   |            | 64.7 INS-001  | LC-MS/MS triple quadrupole | 2026-05-11 |                 | 100            | 0              | 100          | 67.3 | 7  |
| IW-0012   |            | 65.6 INS-001  | LC-MS/MS triple quadrupole | 2026-05-18 |                 | 100            | 6              | 94           | 74.4 | 7  |
| IW-0013   |            | 48.5 INS-001  | LC-MS/MS triple quadrupole | 2026-05-25 |                 | 80             | 0              | 80           | 54.1 | 5  |
| IW-0014   |            | 80 INS-001    | LC-MS/MS triple quadrupole | 2026-06-01 |                 | 100            | 0              | 100          | 88.9 | 9  |
| IW-0015   |            | 89 INS-001    | LC-MS/MS triple quadrupole | 2026-06-08 |                 | 100            | 0              | 100          | 96.4 | 10 |
| IW-0016   |            | 84.7 INS-001  | LC-MS/MS triple quadrupole | 2026-06-15 |                 | 100            | 0              | 100          | 92.9 | 9  |
| IW-0017   |            | 86.1 INS-001  | LC-MS/MS triple quadrupole | 2026-06-22 |                 | 100            | 0              | 100          | 94.7 | 10 |
| IW-0018   |            | 58.3 INS-001  | LC-MS/MS triple quadrupole | 2026-06-29 |                 | 80             | 0              | 80           | 67   | 6  |
| IW-0019   |            | 84.8 INS-001  | LC-MS/MS triple quadrupole | 2026-07-06 |                 | 100            | 0              | 100          | 88.8 | 9  |
| IW-0020   |            | 87 INS-001    | LC-MS/MS triple quadrupole | 2026-07-13 |                 | 100            | 0.2            | 99.8         | 95   | 10 |
| IW-0021   |            | 82.8 INS-001  | LC-MS/MS triple quadrupole | 2026-07-20 |                 | 100            | 1.1            | 98.9         | 85.4 | 9  |
| IW-0022   |            | 83.1 INS-001  | LC-MS/MS triple quadrupole | 2026-07-27 |                 | 100            | 6.6            | 93.4         | 86   | 9  |
| IW-0023   |            | 91.1 INS-001  | LC-MS/MS triple quadrupole | 2026-08-03 |                 | 100            | 0              | 100          | 97.8 | 10 |
| IW-0024   |            | 89.6 INS-001  | LC-MS/MS triple quadrupole | 2026-08-10 |                 | 100            | 0              | 100          | 93.5 | 10 |
| IW-0025   |            | 73.3 INS-001  | LC-MS/MS triple quadrupole | 2026-08-17 |                 | 100            | 0              | 100          | 83.1 | 8  |
| IW-0026   |            | 83 INS-001    | LC-MS/MS triple quadrupole | 2026-08-24 |                 | 100            | 0.8            | 99.2         | 92.4 | 9  |
| IW-0027   |            | 59.3 INS-002  | LC-MS/MS triple quadrupole | 2026-03-02 |                 | 100            | 0              | 100          | 68.5 | 7  |
| IW-0028   |            | 65.8 INS-002  | LC-MS/MS triple quadrupole | 2026-03-09 |                 | 100            | 0              | 100          | 73.1 | 7  |
| IW-0029   |            | 69.8 INS-002  | LC-MS/MS triple quadrupole | 2026-03-16 |                 | 100            | 0              | 100          | 74.8 | 8  |
| IW-0030   |            | 65.9 INS-002  | LC-MS/MS triple quadrupole | 2026-03-23 |                 | 100            | 0.2            | 99.8         | 73.8 | 7  |