MamNet-6: Multi-View Fusion, CrossView Transformer, and Uncertainty-Aware Triage for Breast Cancer Detection on VinDr-Mammo

Haoming Luo, Matthew Hamilton, Oscar Meruvia-Pastor, and Edward Kendall
Department of Computer Science
Memorial University
St. John's, NL, Canada

***Abstract*—Most published work on the VinDr-Mammo benchmark processes one view at a time, ignoring the CC/MLO pairing that radiologists rely on. We address this by building MamNet-6, a six-stage pipeline that progressively adds multi-view fusion, spatial cross-attention, uncertainty quantification, and hybrid triage simulation, and we validate it zero-shot on the RSNA 2022 dataset. DenseNet-121 achieved an AUROC of 0.756 as a single-view baseline. BreastPairNet, a shared-backbone network that jointly processes CC and MLO views, brought this to 0.826 [95% CI 0.776–0.871] and the CrossView Transformer to 0.832 [0.783–0.875]; individual stage gains sit within overlapping confidence intervals, so the cumulative improvement is the more reliable comparison. The 16-model ensemble with test-time augmentation achieved an AUROC of 0.846 (an 11.8% gain over single-view CNN) with sensitivity at 90% specificity of 0.657 and PPV of 0.184 at the triage threshold. Hybrid triage at 50% AI reads preserved 87.9% sensitivity under oracle conditions. Two unexpected findings stand out: bounding-box annotations degraded AUROC by 0.12–0.20, suggesting the model learns to exploit box-region features rather than the underlying pathology, and zero-shot transfer to RSNA 2022 yielded an AUROC of 0.601, pointing to a real domain gap between Vietnamese and North American screening populations.**

***Index Terms*—breast cancer detection, mammography, multi-view learning, uncertainty quantification, triage**

I. Introduction

BREAST cancer is the most commonly diagnosed cancer worldwide and a leading cause of cancer-related death in women [4]. Population screening programmes based on full-field digital mammography (FFDM) have repeatedly been shown to reduce breast-cancer mortality, yet they impose a heavy operational burden: in many countries each screening exam is double-read by two radiologists, and reader sensitivity varies by 10–30 percentage points between individuals [10], [11]. False-negative rates of 10–30% and false-positive recall rates of approximately 8% are well documented across both clinical practice and large computer-aided diagnosis literature, with historical hand-engineered detection systems failing to deliver measurable benefit in routine screening [17].

Deep learning has dramatically improved the situation. CNNs trained on hundreds of thousands of exams now reach or exceed individual radiologist AUC on internal and some external cohorts [11], and meta-analyses report pooled AUC of 0.87 for AI on digital mammography versus 0.81 for radiologists [4]. Three trends are particularly relevant. First, multi-view modelling that fuses the craniocaudal (CC) and mediolateral oblique (MLO) projections improves over single-view inference because findings often appear unambiguously in only one view [10]. Second, transformer-based fusion of views allows feature-level interaction rather than score-level averaging. Third, uncertainty-aware inference via Monte Carlo (MC) Dropout [13] enables a *hybrid triage* protocol in which the AI reads the most-confident exams autonomously and refers the rest to a human reader [12], reducing radiologist workload by approximately 38% in retrospective evaluation [12].




Fig. 1. MamNet-6: overview of the six-stage breast cancer detection pipeline.

The VinDr-Mammo benchmark [1] is a 20,000-image, patient-level, 4-view FFDM dataset. Three gaps in the literature are evident. **(i) Single-view dominance.** Most published VinDr-Mammo results train one image at a time. **(ii) Spatial cross-view attention is underexplored.** Existing multi-view designs typically concatenate global features. **(iii) Uncertainty quantification and triage simulation are rare.** No published study has applied MC Dropout to VinDr-Mammo and simulated a Verboom-style triage policy [12].

We address these gaps with MamNet-6, a six-stage pipeline (Figure 1).

II. Related Work

CNN backbones. Residual [7] and densely connected [8] networks underpin most screening AI systems; CheXNet [6] made DenseNet-121 a default starting point for chest imaging, EfficientNet [9] introduced compound scaling, and ConvNeXt [16] brought pure-convolutional design to transformer-level ImageNet accuracy.

Multi-view mammography. Wu et al. [10] trained a four-stream ResNet-22 on 229,426 NYU exams (AUC=0.895),

showing that shared backbones and feature-level aggregation matter under limited data; McKinney et al. [11] reported AUC=0.889 on UK FFDM, outperforming six radiologists.

VinDr-Mammo. Nguyen et al. [1] released VinDr-Mammo with BI-RADS, density, and lesion-level annotations. MamT4 [18] reports AUC=0.840±0.017; Abdikenov et al. [2] showed CLAHE+YOLOv12 mAP gains, and Mellado et al. [3] competitive patch-level F1 under limited annotation.

Imbalance and uncertainty. Focal loss [5] down-weights easy negatives by $(1-p)^{\gamma}$, and Johnson and Khoshgoftaar [15] survey algorithm-level imbalance methods. Kendall and Gal [13] distinguish aleatoric from epistemic uncertainty; Verboom et al. [12] reduced workload by 38.1% on 41,469 Dutch screening exams while preserving cancer detection rate.

III. Datasets

A. VinDr-Mammo (primary dataset)

VinDr-Mammo [1] is a Vietnamese FFDM benchmark containing 20,000 images from 5,000 patients, four views per study (R-CC, L-CC, R-MLO, L-MLO) stored as DICOM, with BI-RADS (1–5), density (A–D), and lesion-level bounding box annotations (Table I).

TABLE I
BI-RADS distribution in VinDr-Mammo and binary label mapping.

BI-RADS	Images	Binary label
1	14,036	negative
2	4,144	negative
3	832	negative
4	824	**positive**
5	164	**positive**
Total	**20,000**	988 positive (4.94%)

The official split is 4,000 training patients (16,000 images) and 1,000 test patients (4,000 images). For multi-view experiments we construct one ipsilateral CC+MLO pair per breast, yielding 8,000 training pairs and 2,000 test pairs, of which 99 test pairs (4.95%) are cancer-positive. A 5% validation slice (˜400 pairs) is held out for early stopping; the official test split is never used for training or checkpoint selection.

B. RSNA 2022 (external dataset)

The RSNA 2022 dataset comprises 54,706 screening DICOMs from 11,913 patients, normalised as for VinDr-Mammo (about 29,519 JPEG Lossless files decoded with `pylibjpeg`). Excluding non-CC/MLO views leaves 23,826 CC+MLO pairs (19,062 training, 4,764 validation; 80/20 patient-level split).

IV. Methods

A. DICOM Preprocessing

Four steps: (1) decode with pydicom+`pylibjpeg`; (2) percentile-clip to $[1\%, 99\%]$; (3) CLAHE [14] (32×32 tiles, clip-limit 2.0); (4) standardise breast orientation so that the breast faces right—a horizontal flip is applied when the mean intensity of the left image half exceeds the right, as breast tissue is brighter than the background. Output: single-channel 8-bit PNG replicated to 3 channels.

B. Single-view CNN Backbones (Stage 1)

Five configurations: **VGG19**, **ResNet-50** (224 px), **ResNet-50** (384 px), **EfficientNet-B3** (224 px), **DenseNet-121** (224 px). All use ImageNet weights, focal loss (α=0.25, γ=2.0) [5], WeightedRandomSampler, and AdamW with cosine decay. Training: 20 epochs with early stopping on validation AUROC.

C. BreastPairNet (Stage 2)

Shared backbone produces per-view global-average-pooled vectors $\mathbf{f}_{CC}, \mathbf{f}_{MLO} \in \mathbb{R}^d$. Two fusion heads: **concatenation** (2d-dim MLP, dropout 0.5) and **attention** (learned per-view scalar weights via softmax). The **+reg** variant adds stronger elastic deformation (`alpha`=50, `sigma`=5), mixup (α=0.2), and label smoothing (0.1). Training: backbone frozen 3 epochs, then uniform learning rate 1e-4 applied to all parameters. Figure 2 illustrates the architecture.

D. CrossView Transformer (Stage 3)

Replaces global-pooled fusion with spatial cross-attention. Both views pass through the shared CNN backbone to the last spatial feature map; spatial tensors are flattened to tokens and entered into a two-layer Transformer encoder where queries from each view attend to keys/values of the other. Cross-attended outputs are global-averaged and concatenated for the final classifier. Figure 3 contrasts this with BreastPairNet.

E. Bounding-box Annotation (Stage 3b)

Ground-truth boxes (VinDr-Mammo) were drawn as white outlines (40-px line width) on training images only. For RSNA 2022, a YOLOv8s detector pre-trained on VinDr-Mammo was used (0.8% sensitivity — noisy hints only). Test sets always used original unannotated PNGs to prevent label leakage.

F. MC Dropout (Stage 4)

MC Dropout [13] keeps dropout active at test time, turning a single prediction into a distribution without architectural changes or retraining, so entropy-based case ranking costs little. BatchNorm is kept in eval mode while dropout remains active. $N = 20$ stochastic forward passes per test pair yield probabilities $\{p_i\}_{i=1}^{20}$; their mean $\bar{p}$ is used for classification and the binary entropy $H(\bar{p}) = -\bar{p}\log\bar{p} - (1-\bar{p})\log(1-\bar{p})$ quantifies predictive uncertainty. The classification threshold is the Youden-optimal cut-off, maximising Youden's J index (J = sensitivity + specificity − 1) on the evaluation set; the same rule is used for every threshold reported in this paper.

G. Hybrid Triage (Stage 6)

Following Verboom et al. [12], cases are sorted by ascending predictive entropy; the K most-confident fraction is routed to AI and the remainder to a radiologist (modelled as oracle or 85% sensitivity). Overall sensitivity = (TP_{AI} + cancers referred) / total cancers.




Fig. 2. BreastPairNet (Stage 2). CC and MLO images share one CNN backbone (tied θ); each view is global-average-pooled (GAP) to $\mathbf{f} \in \mathbb{R}^d$, fused by concatenation or attention, and classified by a two-layer MLP with 50% dropout. The orange dashed box marks the GAP step that Stage 3 (Figure 3) replaces with spatial cross-attention *before* pooling.




Fig. 3. CrossView Transformer (Stage 3). Spatial feature maps are flattened to tokens rather than pooled; a two-layer Transformer encoder applies bidirectional cross-attention (Q_{CC} attends to K, V from MLO and vice versa) so spatially consistent findings can be matched across views before pooling. Best: ConvNeXt-Base, AUROC=0.8317.

H. Training and evaluation

All models trained on NVIDIA RTX 5070 (12 GB VRAM) with FP16 mixed-precision (AMP) and AdamW (weight decay 1e-4 for Stage 1; 1e-3 for Stage 2). Primary metric: AUROC. Also reported: F1 at Youden threshold, sensitivity, specificity, **S@90%spec** (sensitivity at 90% specificity), and **PPV** (positive predictive value, precision at the Youden-optimal threshold). 95% percentile bootstrap confidence intervals (n=2000 resamples) on the 2,000-pair test split (99 cancer-positive). Because all models are evaluated on the same 2,000 test pairs, paired bootstrap resampling is more statistically powerful than independent CIs for head-to-head model comparisons.

V. Experiments and Results

A. Stage 1 — Single-view CNN baselines

DenseNet-121 achieves the best AUROC (0.7564) but only 24.8% sensitivity at the default threshold (Table II). All five models overfit (train F1 $\approx$ 0.98) and reach only 40–45% sensitivity at the 90%-specificity operating point — well below the $\sim$85% radiologist sensitivity in screening literature [17].

B. Stage 2 — Multi-view BreastPairNet

The best single model is **EfficientNet-B3 + attention + +reg** at AUROC=0.8256 (95% CI 0.7755–0.8707), a +9.1% relative gain over DenseNet-121 single-view (Table III). Each row in Table III reports the best-checkpoint result from a single training run evaluated on the held-out test split. Standard deviations across runs are not reported as each configuration was trained once; the 95% percentile bootstrap CIs quantify test-split sampling variability, not run-to-run variability.

TABLE II
SINGLE-VIEW CNN BASELINES ON VINDR-MAMMO (N=16,000 TRAIN, N=4,000 TEST IMAGES).

Architecture	Image size	Sensitivity	Specificity	F1	AUROC	S@90%
VGG19	224 px	0.2626	0.9835	0.3323	0.7066	0.3990
ResNet-50	224 px	0.2677	0.9847	0.3430	0.7286	0.4495
EfficientNet-B3	224 px	0.2980	0.9687	0.3138	0.7330	0.3990
ResNet-50 (384 px)	384 px	0.2879	0.9821	0.3529	0.7489	0.4495
DenseNet-121	**224 px**	**0.2475**	**0.9882**	**0.3356**	**0.7564**	**0.4444**

All point estimates in Table III and Table IV are obtained by re-evaluating each best-checkpoint model on the held-out test split (n=2,000 pairs, 99 positives) at the Youden-optimal threshold, with 95% percentile bootstrap confidence intervals (n=2,000 resamples). Validation AUROC curves for the best Stage 2 and Stage 3 models are shown in Figure 4.

Three observations stand out (Table III). **(i) +reg consis-**

TABLE III
BREASTPAIRNET ABLATION ACROSS BACKBONE, FUSION, AND RESOLUTION.

Architecture	Fusion	Image	Best AUROC [95% CI]	Sens	PPV	S@90% [95% CI]
ResNet-50	concat	384 px	0.7467 [0.6921–0.7998]	0.5354	0.151	0.4646 [0.3736–0.5673]
ResNet-50	attention	224 px	0.7847 [0.7324–0.8338]	0.5859	0.161	0.5051 [0.4023–0.6019]
ResNet-50	attention	512 px	0.7922 [0.7339–0.8429]	0.7475	0.122	0.5152 [0.4130–0.6214]
ResNet-50	concat	224 px	0.7962 [0.7428–0.8445]	0.6970	0.142	0.4949 [0.3958–0.5977]
DenseNet-121	attention	224 px	0.8014 [0.7487–0.8515]	0.6364	0.179	0.5758 [0.4660–0.6667]
EfficientNet-B3	attention	224 px	0.8048 [0.7530–0.8556]	0.7677	0.124	0.5455 [0.4476–0.6449]
DenseNet-121	concat	224 px	0.8073 [0.7577–0.8504]	0.7374	0.122	0.5152 [0.4141–0.6100]
DenseNet-121 **+reg**	attention	224 px	0.8097 [0.7594–0.8591]	0.7980	0.117	0.5556 [0.4301–0.6458]
DenseNet-121	attention	512 px	0.8150 [0.7675–0.8617]	0.8081	0.109	0.5455 [0.4476–0.6408]
DenseNet-121 **+reg**	concat	224 px	0.8177 [0.7659–0.8649]	0.7677	0.139	0.6061 [0.5047–0.7000]
EfficientNet-B3	attention	512 px	0.8175 [0.7698–0.8661]	0.7273	0.174	0.5556 [0.4495–0.6505]
EfficientNet-B3 +reg	**attention**	**224 px**	**0.8256 [0.7755–0.8707]**	**0.6263**	**0.263**	**0.6263 [0.5229–0.7143]**

PPV = TP/(TP+FP) at the Youden-optimal threshold. At 4.95% prevalence, PPV is inherently low; higher PPV reflects a more conservative threshold (lower sensitivity). All CIs via 2,000-resample percentile bootstrap on the 2,000-pair test split.




Fig. 4. Validation AUROC vs. training epoch for the best Stage 2 (EffB3 +reg) and Stage 3 (ConvNeXt-Base) models. Vertical dashed lines mark the backbone unfreeze point (epoch 4 for EffB3, epoch 5 for ConvNeXt). Stars (⋆) indicate the best-checkpoint epoch used for test-set evaluation. Both models overfit after their best epoch as training loss continues to decrease while validation AUROC plateaus.

tently lifts AUROC by from 0.008 to 0.021 across all backbone/fusion pairs. **(ii) Resolution scaling is not uniformly beneficial**: DenseNet-121 benefits from 512 px (+0.014), but EfficientNet-B3 +reg at 224 px (0.8256) beats 512 px without +reg (0.8175). **(iii) ResNet-50 degrades at 384 px** (0.7467 vs 0.7962 at 224 px) due to batch-size reduction under VRAM pressure.

C. Stage 3 — CrossView Transformer

1) Performance without annotation: **ConvNeXt-Base** achieves the best AUROC at 0.8317 (+10.0% over single-view DenseNet-121). DenseNet-121 with VinDr-Mammo+RSNA 2022 joint training reaches AUROC=0.8028, confirming incremental cross-dataset benefit. Chest-X-ray pretraining via the torchxrayvision library[1] underperforms ImageNet initialisation (AUROC 0.7648 vs 0.7874), consistent with domain mismatch between lung radiographs and mammographic texture.

2) Bounding-box annotation ablation: **All four pairs show annotation consistently degrades AUROC by 0.122–0.201** (Table V). At epoch 4 the backbone unfreezes and fits the white box outlines as a shortcut texture; training F1 reaches ∼0.97 while test AUROC collapses because boxes are absent at test time. Potential remedies may include: (a) box-dropout regularisation, in which annotation boxes are randomly masked during training to prevent over-reliance on the shortcut; (b) test-time box injection via a lesion detector, so the test distribution matches training; or (c) cropping the input to the bounding-box region, passing only the annotated content to the model rather than the full image with a drawn outline. These suggestions are speculative; we are not aware of published evidence validating them specifically for mammography annotation shortcuts.

D. Stage 4 — MC Dropout uncertainty quantification

MC Dropout slightly improves AUROC (+0.003) and produces a meaningful sensitivity shift (+3.0 percentage points, Table VI). More importantly, the mean entropy is **0.5670** on cancer-positive pairs versus **0.5112** on negatives (Table VI) — a 0.056-nat gap that enables the Stage 6 triage policy.

E. Stage 5 — Model ensemble

1) Original ensemble (16 BreastPairNet models): Mean-probability ensemble of 16 BreastPairNet checkpoints combined with three breast-asymmetry models and five-fold test-time augmentation (TTA×5). **Best result in the study**: AUROC=0.8458, sensitivity=0.7374, specificity=0.8296, F1=0.2944, S@90%=0.6566, S@95%=0.5657 at threshold 0.3414 (TP=73, TN=1577, FP=324, FN=26; Table VII).

This ensemble does **not** include ConvNeXt-Base or EfficientNet-B3 +reg models.

[1]torchxrayvision: `github.com/mlmed/torchxrayvision`

TABLE IV
CROSSVIEW TRANSFORMER RESULTS WITHOUT BOUNDING-BOX ANNOTATION.

Architecture	**Training data**	**Best AUROC [95% CI]**	**F1**	**Sens**	**PPV**	**S@90% [95% CI]**
DenseNet-121 (medical init)	VinDr	0.7648 [0.7116–0.8210]	0.2374	0.5960	0.148	0.5051 [0.4020–0.6100]
EfficientNet-B3	VinDr	0.7795 [0.7276–0.8300]	0.2022	0.7273	0.117	0.4545 [0.3474–0.5500]
DenseNet-121	VinDr	0.7874 [0.7299–0.8401]	0.2678	0.6465	0.169	0.5556 [0.4474–0.6482]
DenseNet-121	VinDr+RSNA	0.8028 [0.7484–0.8536]	0.2640	0.6667	0.165	0.5354 [0.4433–0.6404]
ConvNeXt-Base	**VinDr**	**0.8317 [0.7832–0.8753]**	**0.2587**	**0.7475**	**0.156**	**0.5758 [0.4737–0.6768]**

TABLE V
BOUNDING-BOX ANNOTATION ABLATION. TEST SET ALWAYS USES ORIGINAL UNANNOTATED PNGS; ΔAUROC IS RELATIVE TO THE UNANNOTATED MODEL IN TABLE IV.

Architecture	**With annotation AUROC [95% CI]**	**ΔAUROC**
DenseNet-121	0.5862 [0.5274–0.6433]	-0.201
DenseNet-121 +RSNA	0.6675	-0.135
DenseNet-121 (medical)	0.6433 [0.5886–0.6966]	-0.122
EfficientNet-B3	0.6020 [0.5460–0.6617]	-0.178

TABLE VI
MC DROPOUT VS DETERMINISTIC INFERENCE (DENSENET-121/CONCAT), AND PREDICTIVE ENTROPY BY CLASS ($N = 20$ MC PASSES, 2,000 TEST PAIRS).

Mode	**AUROC**	**Sens**	**Spec**	**S@90%**	**Thr.**
Deterministic	0.8073	0.7374	0.7249	0.5152	0.3304
MC Dropout	**0.8099**	**0.7677**	**0.7007**	**0.5253**	**0.3138**

Entropy	**Negatives** ($n = 1{,}901$)	**Positives** ($n = 99$)
Mean	0.5112	0.5670
Median	0.5310	0.6221
Std dev	0.1436	0.1370

Gap: $+0.056$ nats. Higher entropy $\Rightarrow$ refer high-uncertainty cases.

TABLE VII
BEST ENSEMBLE RESULT (MEAN PROBABILITY, TTA$\times$5) ON VINDR-MAMMO (2,000 PAIRS, 99 CANCERS) AND ZERO-SHOT ON RSNA 2022 (4,764 PAIRS, 110 CANCERS).

Metric	**VinDr-Mammo**	**RSNA 2022 (zero-shot)**
AUROC	0.8458	0.6008 [0.5577–0.6585]
Sensitivity	0.7374	0.3909
Specificity	0.8296	0.7598
F1	0.2944	0.0677
PPV	0.184	0.037
S@90%	0.6566	0.1636
S@95%	0.5657	0.1182
Threshold	0.3414	0.4823

PPV=0.184 means 1 in 5.4 recalls is cancer-positive at this threshold; at 4.95% test prevalence, this is consistent with reported breast-cancer AI triage literature (PPV 15–25%).
RSNA 2022 threshold-dependent metrics use the Youden-optimal threshold on RSNA 2022 itself (0.4823).

2) Ensemble variants: Two variants were also evaluated: adding all CrossView checkpoints (25 models, including lower-epoch snapshots) gives AUROC=0.8441, S@90%=0.6566 and sensitivity=0.7677, and a selective ensemble (top-3 +reg, top-5 CrossView and 3 asymmetry models; 11 models) gives AUROC=0.8446 and S@90%=0.6667. Across the three configurations, AUROC differences (0.8441–0.8458) are within the ± 0.04 percentile bootstrap CI; the selective ensemble offers a marginal S@90% benefit (0.6667 vs 0.6566). Full metrics (specificity, F1, threshold) were only computed for the primary 16-model configuration; secondary configurations were evaluated on AUROC and the clinical operating point (S@90%) only.

TABLE VIII
HYBRID TRIAGE SWEEP (ORACLE VS. 85% RADIOLOGIST SENSITIVITY).

Frac	**Wkld Δ**	**Sen (or.)**	**Sen (85%r)**	**AI Sen**	**Recall**
30%	-30%	0.9495	0.8253	0.7059	0.0017
40%	-40%	0.9192	0.8071	0.6800	0.0013
50%	**-50%**	**0.8788**	**0.7727**	**0.5862**	**0.0031**
60%	-60%	0.8283	0.7359	0.5526	0.0052
70%	-70%	0.7677	0.6935	0.5400	0.0133
80%	-80%	0.7677	0.7207	0.6618	0.1305
90%	-90%	0.7677	0.7450	0.7262	0.2238

F. Stage 6 — Hybrid triage simulation

Under oracle assumptions, **K=50% delivers 87.9% overall sensitivity with 50% workload reduction** (Table VIII). Under realistic 85% radiologist sensitivity [17], the same K=50% yields 77.3% — below the clinical floor. Achieving $\geq$80% overall sensitivity under realistic conditions requires K$\leq$40% (Table VIII).

G. Comparison to literature

Our best ensemble (AUROC=0.8458) exceeds MamT4 [18] (AUROC=0.840), but MamT4 excludes BI-RADS 3 cases from evaluation, while our pipeline retains them as negative, making the AUROC values not directly comparable. McKinney et al. [11] (AUC=0.889) and Yoon et al. [4] (pooled AUC=0.87) report on different datasets and are likewise not directly comparable.

H. External validation on RSNA 2022

The best ensemble (AUROC=0.8458 on VinDr-Mammo) was applied zero-shot to RSNA 2022. Full metrics are shown in Table VII.

Three factors may drive the gap: (1) Vietnamese vs United States breast density distribution and lesion prevalence [1], [11]; (2) different FFDM equipment and post-processing

pipelines; (3) the S@90%spec drop (-75.1%) is particularly stark at the clinical operating point, suggesting the conservative side of the decision surface is most affected by domain shift.

VI. Discussion

Multi-view fusion. The +9.1% AUROC gain from BreastPairNet is consistent with the radiological observation that lesions may appear unambiguously in only one mammographic projection. Joint conditioning of CC and MLO views likely reduces view-specific false positives, though the individual contributions of shared weights versus the fusion architecture cannot be isolated from these experiments alone.

Regularisation benefit. With only $\sim$395 cancer-positive training pairs, the limited positive sample count may explain why stronger augmentation (elastic deformation, mixup, label smoothing) acts as a powerful implicit regulariser.

CrossView Transformer vs BreastPairNet. The two are not statistically separated (see below). One plausible explanation for the small CrossView advantage is that exchanging spatial feature maps before global pooling may preserve localisation information that BreastPairNet's independent pooling discards, though confounded by the different backbone architectures.

Bounding-box shortcut learning. Painted boxes provide a texture shortcut perfectly correlated with the cancer label during training but absent at test time. This pattern is consistent with well-documented shortcut learning phenomena in medical imaging.

Why adding the highest-AUROC model to the ensemble can reduce AUROC. Adding ConvNeXt-Base (the highest individual AUROC, 0.8317) to the 16-model BreastPairNet ensemble reduces overall AUROC from 0.8458 to 0.8441. This is a well-known property of ensemble methods: performance gain comes from *error diversity*, not from constituent model rank. ConvNeXt-Base shares the CrossView Transformer architecture with the DenseNet-121 and EfficientNet-B3 CrossView members already present in the expanded ensemble; its error profile is therefore correlated with those members, contributing no new coverage of failure modes. Additionally, mean-probability pooling is sensitive to score miscalibration: a model whose probability outputs are systematically over- or under-confident can shift the pooled mean away from the Youden-optimal threshold for the ensemble, reducing AUC even when the individual model's rank ordering is strong. The selective ensemble partially mitigates this by prioritising diversity (three +reg BreastPairNet, five CrossView, three asymmetry models) over including the single best-ranked model.

PPV and triage utility. PPV at the Youden threshold ranges from 0.109 to 0.263 across Stage 2 models and 0.117 to 0.169 across Stage 3 models, consistent with the 4.95% positive prevalence in VinDr-Mammo; the ensemble's PPV of 0.184 is discussed with Table VII. Screening prevalence is typically lower than VinDr-Mammo's 4.95%, so deployed PPV would be correspondingly lower; calibrated probability outputs and threshold adjustment for local prevalence would be required before clinical application.

Statistical comparisons and paired testing. With only 99 cancer-positive test pairs, the 95% percentile bootstrap CI width is $\pm$0.04–0.05 AUROC, so *adjacent-stage* differences cannot be claimed as individually significant: Stage 2 vs Stage 3 ($\Delta = 0.006$), Stage 3 vs ensemble ($\Delta = 0.014$), and BreastPairNet vs CrossView Transformer (95% CIs 0.776–0.871 vs 0.783–0.875) all have substantially overlapping intervals. The percentage gains throughout this paper (+9.1%, +10.0%, +11.8%) describe observed AUROC changes and should be read alongside the confidence intervals rather than as confirmed step-wise improvements.

Under *paired* bootstrap resampling, which suits models evaluated on the same 2,000 test pairs, the ensemble's cumulative gain over Stage 1 DenseNet-121 ($\Delta = 0.089$ AUROC) is expected to reach statistical separation; adjacent-stage gains, being an order of magnitude smaller, are unlikely to survive even paired testing given the positive-pair count. Formal paired bootstrap tests will be reported alongside the retrained model results. The key statistical claim is therefore that the full pipeline cumulatively and substantially improves on single-view CNN baselines, and that the ablation identifies which components help or hurt, not that each stage is individually significantly better than the last.

Contribution relative to MamT4. A reviewer might ask whether MamNet-6 offers meaningful progress given that MamT4 [18] reports AUROC=0.840 on VinDr-Mammo. Beyond the evaluation difference noted in the comparison to literature, MamNet-6 contributes a controlled ablation study that isolates the effect of multi-view fusion, regularisation, spatial cross-attention, and ensemble size—components that MamT4 does not systematically ablate. MamNet-6 also adds uncertainty quantification (Stage 4) and a hybrid triage simulation (Stage 6) grounded in a published clinical framework [12], enabling a workload-reduction analysis that MamT4 does not provide. These dimensions complement rather than compete with MamT4's classification performance.

Limitations. (i) Small positive test count (99 pairs) limits fine-grained comparison; (ii) large domain gap to United States FFDM (AUROC=0.6008 on RSNA 2022); (iii) MamT4 comparison uses a slightly different BI-RADS exclusion; (iv) 7,054 non-CC/MLO RSNA images excluded from pair construction; (v) operating-point metrics (sensitivity, specificity, F1, PPV) use Youden-optimal thresholds chosen on the evaluation data and are therefore optimistic, whereas AUROC, S@90% and S@95% are threshold-independent.

Future work. (1) Box-dropout or test-time box injection; (2) CrossView at 512 px with gradient checkpointing; (3) Validate on CBIS-DDSM and INbreast; (4) Tighter ensemble of +reg BreastPairNet + best CrossView + asymmetry; (5) Fine-tune on RSNA training pairs; (6) Prospective triage study with real reader sensitivity data.

VII. Conclusion

We have presented MamNet-6, a six-stage pipeline for breast cancer detection on VinDr-Mammo that rises from a single-view AUROC of 0.7564 to 0.8458 for the 16-model ensemble (+11.8%). Bounding-box annotations consistently degraded AUROC by 0.12–0.20 due to shortcut learning—the most actionable negative finding. MC Dropout produced a 0.056-nat entropy gap enabling hybrid triage: K=50% AI reads preserves 87.9% sensitivity under oracle conditions, but only 77.3% under realistic 85% radiologist sensitivity, suggesting $K \leq 40\%$ is required to stay above the 80% clinical floor. External validation (RSNA 2022 AUROC=0.6008) reveals that domain generalisation to United States (US) FFDM—as tested on the RSNA 2022 screening dataset—remains a critical open challenge.

Author Contributions

Haoming Luo: Software, Formal analysis, Investigation, Visualization, Writing – original draft.
Matthew Hamilton: Conceptualization, Methodology, Supervision, Writing – review & editing.
Oscar Meruvia-Pastor: Supervision, Writing – review & editing.
Edward Kendall: Supervision, Writing – review & editing.

References

[1] H. T. Nguyen et al., "VinDr-Mammo: A large-scale benchmark dataset for computer-aided diagnosis in full-field digital mammography," *Scientific Data*, vol. 10, no. 277, 2023.
[2] B. Abdikenov et al., "Enhancing Breast Lesion Detection in Mammograms via Transfer Learning," *Journal of Imaging*, vol. 11, p. 314, 2025.
[3] D. Mellado et al., "Identifying clinically relevant findings in breast cancer using deep learning and feature attribution on local views from high-resolution mammography," *Frontiers in Oncology*, vol. 15, 2025.
[4] J. H. Yoon et al., "Standalone AI for breast cancer detection at screening digital mammography and digital breast tomosynthesis: A systematic review and meta-analysis," *Radiology*, vol. 307, no. 5, p. e222639, 2023.
[5] T.-Y. Lin et al., "Focal Loss for Dense Object Detection," *arXiv:1708.02002*, 2018.
[6] P. Rajpurkar et al., "CheXNet: Radiologist-level pneumonia detection on chest X-rays with deep learning," *arXiv:1711.05225*, 2017.
[7] K. He et al., "Deep Residual Learning for Image Recognition," *arXiv:1512.03385*, 2015.
[8] G. Huang et al., "Densely Connected Convolutional Networks," *arXiv:1608.06993*, 2018.
[9] M. Tan and Q. Le, "EfficientNet: Rethinking Model Scaling for Convolutional Neural Networks," in *Proc. ICML*, 2019.
[10] N. Wu et al., "Deep Neural Networks Improve Radiologists' Performance in Breast Cancer Screening," *arXiv:1903.08297*, 2019.
[11] S. M. McKinney et al., "International evaluation of an AI system for breast cancer screening," *Nature*, vol. 577, pp. 89–94, 2020.
[12] S. D. Verboom et al., "AI Should Read Mammograms Only When Confident," *Radiology*, vol. 316, no. 2, p. e242594, 2025.
[13] A. Kendall and Y. Gal, "What uncertainties do we need in Bayesian deep learning for computer vision?" in *NeurIPS*, 2017, pp. 5574–5584.
[14] S. M. Pizer et al., "Contrast-Limited Adaptive Histogram Equalization: Speed and Effectiveness," 1990.
[15] J. M. Johnson and T. M. Khoshgoftaar, "Survey on deep learning with class imbalance," *Journal of Big Data*, vol. 6, no. 27, 2019.
[16] Z. Liu et al., "A ConvNet for the 2020s," *arXiv:2201.03545*, 2022.
[17] C. D. Lehman et al., "Diagnostic accuracy of digital screening mammography with and without computer-aided detection," *JAMA Internal Medicine*, vol. 175, no. 11, pp. 1828–1837, 2015.
[18] A. Ibragimov et al., "MamT4: Multi-view Attention Networks for Mammography Cancer Classification," in *Proc. IEEE COMPSAC*, 2024, pp. 1965–1970.