Designing Micro-fabricated fiducials for cellular tracking with MRI

Aaron T. Becker

Abstract—A Magnetic Resonance Imaging (MRI) scanner typically has a minimum resolution of 0.2x0.2mm² with a high-field 7T scanner and 0.5x0.5mm² with a clinical 3T scanner[1].

Cells are generally much smaller than this resolution: 1 red blood cells are 6-8 μ m, human muscle cells are 1-40 mm in length and 10-50 μ m in width [2] squamous Epithelial Cells are 30-60 μ m, irregular thin flat flagstone shape with distinct edges [3], an E. coli bacterium is rod-shaped and approximately 0.7-1.4 μ m in diameter by 2-4 μ m in length [4].

Small particles of magnetic material are used as contrast agents in MRI. These particles can be smaller than the resolution of the scanner because the particles produce dark spots in the image that are larger than the particles. Individual particles as small as 1.68 μ m can be detected using 7T MRI scanners, and as small as 15 μ m using clinical 3T scanners. Particles produced by micro-fabrication processes can be more uniform and be easier to detect than those produced by chemical synthesis.

I. OVERVIEW

MRI scanners offer a safe technique for imaging tissue deep within a living body. However, it is often difficult to differentiate a tissue of interest from its surroundings using only magnetic resonance. Additionally, the tissue of interest may be too small to detect. To solve these problems, MRI markers (also called tags or fiducials) are used, along with contrast agents. This literature review compares MRI marking methods.

MRI markers include nano-scale MRI-contrast agents [5], micro-scale particles, micro-fabricated MRI labels, and centimeter-scale fiducials.

MRI contrast agents and micro-particles usually require large concentrations of the material because

individual particles are difficult to resolve with MRI.

Many groups are designing MRI labels for single-cell tracking. These usually require either *ultra-high field MRI*, defined as scanners with field strengths \geq 7T, specialized coil systems, or large particles.

Zabow et al. designed $\approx 2\mu$ micro-fabricated markers for MRI with a key property that by varying lengths of the design, a scanning sequence could be designed to detect only one variety of the markers at a time. They designed a stacked disc structure and a cylindrical shell[6]. Each design owed its high signal to noise ratio on the fact that water would flow through the shape.

Later the same team designed special microfabricated cuboids composed of iron or nickel covered in gold. These are more effective than chemically produced spheres because they contain a.) purer magnetic material and b.) more magnetic material than similarly sized spheres [7].

II. ENGINEERED PARTICLES THAT CAN BE INDIVIDUALLY DETECTED

A. Micro-engineered particles for multi-spectral MRI

Zabow et al. [6] designed labels that frequency shift water by discrete amounts. The phase shift can be designed to make multiple distinguishable markers. The markers must be aligned with the MRI B_0 field, but were designed anisotropically so they passively align. They presented a design using two nickel disks of radius R separated by distance S

Their Figure 3 showed 1.24mm discs designed for three different frequencies. Each was distinguishable, Resolution was 500 x 750 μ meter. Fig-

 $^{^{1}\}mathrm{For}$ a catalog of useful biological numbers, see bioNumbers.org

ure 5 resolution was 100 x 100 |mumeter with much smaller particles 5μ meter disks separated by 2μ meters. Each voxel contained 10 to 20 particles.

They claim that their design with an open interior have a 10^4 greater signal-to-noise ratio than enclosed designs.

Competing methods: labeling cells with superparamagnetic iron oxide (SPIO) nanoparticles or dentrimers[8], or micrometer-sized particles of iron oxide (MPIOs). These methods magnetically dephase the signal from water surrounding these particles.

Specially engineered micro coils can be used for high resolution MRI imaging, but this requires the coils to be placed close to the target [9]. This paper referred to tests with 5 nano-liter samples for NMR spectroscopy. The coils used surrounded the capillary containing the liquid under interrogation. The coils were 1mm long, with an OD of 0.5mm. Since these are not for imaging, they could instead be used to indicate whether a compound was present or not.

B. Magnetodendrimers

Magnetodendrimers are magnetic tags used to efficiently label mammalian cells. These tags are diffused into the cell culture, and brought into the cells through a membrane adsorption process with subsequent localization in endosomes. They are visible in MR with doses as low as 1 μ g iron/ml culture medium. Labelled cells were detected with 9-14 pg/cell for up to 6 weeks after transplant [8].

C. Stainless steel microparticles in capillaries

This work by Olamaeni et al. [1] is inspired by the rapeutic magnetic micro carriers – $\approx 50 \mu \text{m}$ diameter drug capsules that can be steered and tracked by an MRI scanner.

Magnetic bodies create susceptibility artifacs in the MRI image. The paper[1] includes equations describing how the particle distorts the MRI image. Testing used 0.4mm diameter chrome steel spheres with Gradient-Echo imaging. Pairs of particles are distinguishable when their distance is ≈ 40 diameters apart. The imaging used 0.5 mm pixels on a 1.5 T Magnetom Siemens scanner. The

authors demonstrate a 15 μ m sphere is detectable in a clinical scanner [10]. Such a particle can navigate in large capillaries, which are 20μ m in diameter, but cannot reach small capillaries, which are 4-5 μ m in diameter.

Later they demonstrated tracking $50\mu m$ diameter droplets of ferrofluid using a 1.5T scanner [11]. The droplets moved along a $50\mu m$ capillary and were tracked with Gradient-Echo imaging with 0.6mm voxels. The droplets contained 10% MNP (Fe₃O₄).

III. CONTRAST AGENTS AND MRI FIDUCIALS

A. nano-particle MRI contrast agents

These typically use iron-oxide nanoparticles as MRI contrast agents. Often the particles are collected by cells in the body, e.g. macrophages, and show are dark pixels in an MRI image [?].

B. Commercial MRI fiducial Markers

There are many products available today for MRI fiducials, used for image registration to link current scanned data with previous scans and targets. MRI fiducials used in routine care range from large 20 mm markers designed to be stuck on the skin, or \approx 20mm gold markers inserted via needle into the skin. These have a variety of forms, but are centimeter-scale large objects.

MRI fiducials inserted by a 17-Gauge needle, made of 99.95% gold construction with a snake-like design., 1.15mm x 1 cm, to 0.85mm x 3cm²

Skin markers: ³, 1.75 cm RadianceTMfilled flexible packets, 1.5cm RadianceTMfilled tubes, to 3cm RadianceTM-filled tubes.

Multi-modality Center-hole markers: ⁴, a flat disc with 2mm center hole, 15 mm outer diameter, 3.5mm thick, bright in MRI image.

²http://www.mriequip.com/store/pc/ viewPrd.asp?idproduct=929

³http://www.beekley.com/MRI/MRSPOTS.asp

⁴http://www.universalmedicalinc.com/ Center-Hole-Mammography-Skin-Marker-p/ mm3002.htm

REFERENCES

- N. Olamaei, F. Cheriet, and S. Martel, "Accurate positioning of magnetic microparticles beyond the spatial resolution of clinical mri scanners using susceptibility artifacts," in Engineering in Medicine and Biology Society, EMBC, 2011 Annual International Conference of the IEEE. IEEE, 2011, pp. 2800–2803.
- [2] H. Lodish, A. Berk, S. L. Zipursky, P. Matsudaira, D. Baltimore, J. Darnell *et al.*, "Muscle: A specialized contractile machine," 2000. [Online]. Available: http://www.ncbi.nlm.nih.gov/books/NBK21670/
- [3] N. A. Brunzel, "Fundamentals of urine and body fluid analysis," 2004.
- [4] A. Zaritsky and C. L. Woldringh, "Chromosome replication rate and cell shape in escherichia coli: lack of coupling." *Journal of bacteriology*, vol. 135, no. 2, pp. 581–587, 1978.
- [5] M. R. Carroll, R. C. Woodward, M. J. House, W. Y. Teoh, R. Amal, T. L. Hanley, and T. G. St Pierre, "Experimental validation of proton transverse relaxivity models for superparamagnetic nanoparticle mri contrast agents," *Nanotechnology*, vol. 21, no. 3, p. 035103, 2010.
- [6] G. Zabow, S. Dodd, J. Moreland, and A. Koretsky, "Microengineered local field control for high-sensitivity multispectral mri," *Nature*, vol. 453, no. 7198, pp. 1058–1063, 2008.
- [7] G. Zabow, A. Koretsky, and J. Moreland, "Design and fabrication of a micromachined multispectral magnetic resonance imaging agent," *Journal of Micromechanics and Microengineering*, vol. 19, no. 2, p. 025020, 2009.
- [8] J. W. Bulte, T. Douglas, B. Witwer, S.-C. Zhang, E. Strable, B. K. Lewis, H. Zywicke, B. Miller, P. van Gelderen, B. M. Moskowitz et al., "Magnetodendrimers allow endosomal magnetic labeling and in vivo tracking of stem cells," *Nature biotechnology*, vol. 19, no. 12, pp. 1141–1147, 2001.
- [9] D. L. Olson, T. L. Peck, A. G. Webb, R. L. Magin, and J. V. Sweedler, "High-resolution microcoil 1h-nmr for mass-limited, nanoliter-volume samples," *Science*, vol. 270, no. 5244, pp. 1967–1970, 1995.
- [10] N. Olamaei, F. Cheriet, G. Beaudoin, and S. Martel, "Mri visualization of a single 15 μm navigable imaging agent and future microrobot," in Engineering in Medicine and Biology Society (EMBC), 2010 Annual International Conference of the IEEE. IEEE, 2010, pp. 4355–4358.
- [11] N. Olamaei, F. Cheriet, and S. Martel, "Magnetic resonance imaging of microvessels using iron-oxide nanoparticles," *Journal of Applied Physics*, vol. 113, no. 12, p. 124701, 2013.