



OR # _____

Amount _____

Date _____

SSDL CALIBRATION REQUEST FORM FOR ACTIVITY METER

Instructions:

1. Fill out *all* the fields below and check ☒ the box ☐ that corresponds to your facility.
2. **Please write legibly.** All information written in this form will be the basis of the certificate/report issued.
3. **A maximum of two (2) units per client may be submitted to each service request.**

I. Customer Information			
Company Name:			
Address:			
Contact Person:			
Mobile Number:		Designation:	
Landline Number:		Email Address:	
II. Field of Application			
☐ Nuclear Medicine		☐ Radioisotope production/distribution (e.g., Production/ distribution of I-131, Tc-99m, etc.)	
☐ OTHERS (Please specify): _____			
III. Instrument Information			
Manufacturer		Model	Serial Number
Reference sources available in facility:		☐ Cesium-137 ☐ Cobalt-57 ☐ Barium-133 ☐ OTHERS (Please specify): _____	
Highest activity being measured in Becquerel (Bq):			
IV. Terms and Conditions			
1. The customer shall be responsible for all logistical arrangements required to facilitate the PNRI Calibration Team's service activities. These responsibilities include, but are not limited to, arranging transportation between PNRI and the customer's facility, providing accommodation when necessary, and ensuring the availability of any required incidentals (e.g., personal protective equipment and related materials) during the agreed-upon service schedule. 2. The customer shall be responsible for preparing the Tc-99m or any other radioisotope required for the Linearity Test. The activity level of the prepared source shall be within the range normally used during routine operations. In the event that the required radioisotope is not available at the time of calibration, the Linearity Test shall not be performed, and this test will not be reflected in the Certificate of Instrument Calibration to be issued. 3. The customer shall ensure that the designated work area is free from contamination and elevated radiation levels at the time of calibration. Should background radiation levels exceed acceptable limits, PNRI shall implement the procedures outlined in the Workflow for Handling High Background Levels During Calibration. 4. Once the PNRI Calibration Team has arrived at the customer's facility and has commenced calibration activities, the service shall be deemed delivered. Following the completion of the calibration work, and depending on the evaluation results, either a Certificate of Calibration or a Certificate of Test shall be issued after ten (10) working days from the date of service. Any rescheduling not for reasons not of fault by PNRI will be considered as a new service request and customer shall pay corresponding fees. 5. In the event of severe natural phenomena (e.g., storms, typhoons, earthquakes, volcanic activity), public health emergencies, security incidents (e.g., bomb threats, civil unrest), or any other unforeseen circumstances that pose risks to PNRI personnel or impede transportation to or from the service location—including but not limited to flight disruptions—PNRI reserves the right to defer, modify, or temporarily suspend service deliverables. Service activities shall be resumed at the earliest feasible and safe opportunity, subject to coordination and mutual agreement with the client. 6. If no complaints regarding the calibration are raised within one (1) week after the release of the certificate, the result of the calibration shall be considered final and acceptable. The customer may review the calibration results and raise any concerns or issues they may have. 7. The PNRI is implementing a cash payment policy. The services being requested will be provided only upon presentation of the official receipt. Upon payment, contact RPSS for scheduling of service through tps@pnri.dost.gov.ph or (02) 8929-6011 to 19 local 246/262.			
I have read and agreed with all the terms and conditions stated above; and all information provided are correct and updated.			
Signature over Printed Name of Applicant		Date	RPSS Receiving Personnel