

Vaccine Temperature Monitoring Requirements:

- The National Vaccine Storage and Handling Guidelines (2015), as well as many provincial guidelines, require manual documentation of vaccine temperatures, where the: current, minimum and maximum refrigerator temperatures, and the ambient room temperature must be recorded at the beginning and end of each day (min/max temperatures must be reset after each recording) – log templates are available, refer to the guidelines, or contact us for assistance
- The digital display should be placed on the **outside** of the refrigerator (Ontario requires a display with a minimum resolution of 0.1° C), with the probe in the centre of the refrigerator (preferably in a vaccine box)
- Bottle probes (bio-safe glycol fluid bottle probes) are recommended for a more accurate reading
- Reset button required to reset the minimum and maximum temperature after each recording
- Thermometers should be set to alarm if temperatures fall to 3°C or rise to 7°C to signify approaching out of range
- Thermometers should be checked for calibration at least annually, and it is recommended they be calibrated within $\pm 0.5^{\circ}\text{C}$ ($\pm 1^{\circ}\text{F}$) of accuracy
- Thermometers should have a low battery indicator, and batteries should be replaced every 6 months



General Vaccine Storage Requirements:

- ✓ Maintain required vaccine storage temperatures
- ✓ Hold sufficient inventory
- ✓ Have a calibrated temperature-monitoring device inside each storage compartment
- ✓ Refrigerator is dedicated to the storage of vaccines only
- ✓ Be placed in a secure location
- ✓ Recommended thermometer calibration should be accurate within $\pm 0.5^{\circ}\text{C}$ ($\pm 1^{\circ}\text{F}$)
- ✓ Record refrigerator temperature twice daily
- ✓ Contact public health officials if vaccines are exposed to temperatures outside the recommended range