

Cardiac positronium lifetime in human PET: a reproducible right–left ventricular contrast that is not explained by blood oxygenation

Author: Ernest Darell Plascencia Zermeño¹

¹ Independent researcher, Guadalajara, Mexico. Correspondence: ernestzermeno3@gmail.com.

Preprint — single subject, public-data re-analysis. Version 1.

Abstract

Background. Ortho-positronium (o-Ps) lifetime, now measurable in vivo on long-axial-field-of-view (LAFOV) PET/CT, has been proposed as a biomarker of tissue oxygenation and hypoxia. Because o-Ps lifetime is dominated by tissue free-volume structure while the oxygen-specific contribution is small, whether an in-vivo lifetime contrast reflects oxygenation rather than anatomy is an open, identifiability-limited question.

Aim. To test the oxygenation hypothesis directly using the heart's natural arterial/venous oxygenation contrast, with a built-in anatomical control.

Methods. We re-analysed a public [⁸²Rb]Cl human cardiac LAFOV PET/CT dataset (5.30×10⁸ evaluated three-photon events). Per-compartment o-Ps lifetimes were extracted with a background-plus-two-component exponentially-modified-Gaussian (EMG) model. The list-mode→image mapping and right/left ventricle (RV/LV) identity were established lifetime-free (the mapping reproduces the provider's reconstructed image at block-correlation 0.998 and wins a joint multi-organ alignment panel). We applied a confound battery — registration stress test, blood-core vs wall, lung-air and wall-myocardium partial-volume, tissue density — and a structure/position-matched control (pulmonary artery, deoxygenated, vs aorta, oxygenated). An isotope-matched ⁸²Rb uniform-quartz reference bounded the instrument's positional behaviour. All results were produced by two independent analysis pipelines.

Results. RV o-Ps lifetime exceeded LV by $\Delta\tau = +0.304$ ns (RV 1.700 ± 0.172 , LV 1.396 ± 0.130 ns; $\sim 1.4\sigma$), in the oxygen-expected direction; the contrast was stable across ± 16 mm registration perturbation (sign preserved in 100% of 342 shifts) and resided in the blood core, not the wall. However, the matched-vessel control was null: pulmonary artery – aorta = -0.011 ± 0.344 ns. Lung-air and wall-myocardium partial-volume were disfavoured, and the effect fell within the isotope-matched ⁸²Rb instrumental positional envelope (~ 0.1 – 0.35 ns over 40 mm in uniform material).

Conclusion. On this single subject, the cardiac o-Ps lifetime contrast does not provide a clean readout of blood oxygenation: an oxygenation effect of the observed (~ 0.3 ns) magnitude is ruled out by the matched control, while a small physiological effect cannot be excluded. We provide a reusable confound-control battery for evaluating future in-vivo o-Ps oxygenation claims. Multi-subject replication with anatomy decoupled from oxygenation is required.

Keywords: positronium lifetime imaging; ortho-positronium; PET/CT; tissue oxygenation; hypoxia biomarker; long axial field of view; confound control.

1. Introduction

In a fraction of positron–electron annihilations the pair first forms positronium, a hydrogen-like bound state. Its triplet form, ortho-positronium (o-Ps), is metastable; in matter it localises in nanometre-scale free-volume voids and annihilates predominantly by pick-off, so its lifetime (≈ 1.5 – 3 ns in soft tissue, versus 142 ns in vacuum) reports the size of those voids and the local electron density. Paramagnetic molecular oxygen further shortens o-Ps lifetime through spin (ortho→para) conversion and oxidation; this oxygen sensitivity is the physical basis for proposing o-Ps lifetime as an imaging biomarker of tissue oxygenation and hypoxia [1,2].

The proposal has gained momentum as positronium lifetime imaging (PLI) has moved from the bench to the clinic. Dedicated multi-photon systems and, more recently, commercial LAFOV PET/CT scanners equipped to register a

prompt γ in triple coincidence have produced the first in-vivo human positronium images and per-region lifetimes [3–6]. These are real technical achievements. They do not, by themselves, establish that an in-vivo lifetime contrast between tissues reflects oxygenation: o-Ps lifetime is dominated by free-volume structure, which varies by ~ 1 ns between tissue types, whereas the oxygen-specific term implied by published quenching constants is small — on the order of tens of picoseconds across physiological dissolved-oxygen differences [1]. Any in-vivo oxygenation claim therefore faces an identifiability problem, because oxygenation and tissue structure are easily confounded and the oxygen signal is the smaller of the two.

The heart offers an unusually clean natural test. The right ventricle (RV) carries deoxygenated blood and the left ventricle (LV) oxygenated blood; oxygen-quenching predicts a longer o-Ps lifetime in the RV. Crucially, the cardiopulmonary anatomy provides a built-in control that decouples oxygenation from vessel type and position: the pulmonary artery carries deoxygenated blood but is, structurally and positionally, a central great vessel like the aorta, which carries oxygenated blood. If a lifetime contrast tracks oxygenation, the pulmonary artery should group with the venous (long-lifetime) side; if it tracks anatomy, the pulmonary artery should group with the aorta. This single comparison breaks the oxygenation/structure degeneracy that an isolated $RV > LV$ difference cannot.

Here we use a public human cardiac PLI dataset to ask whether the RV/LV o-Ps lifetime contrast is attributable to oxygenation. We find a reproducible, registration-stable, blood-borne $RV > LV$ contrast in the oxygen-expected direction — yet the matched-vessel control is null, and the effect lies within the scanner’s own isotope-matched positional envelope. The contribution is twofold: a cautionary single-subject result, and a reusable confound-control battery (matched-vessel test plus exclusion checks plus an isotope-matched instrument reference) for evaluating any future in-vivo o-Ps oxygenation claim.

2. Methods

2.1 Data

We used publicly deposited, de-identified data from the Bern/Inselspital group acquired on a Biograph Vision Quadra (Siemens Healthineers) [3,4]. The primary subject is an in-vivo [^{82}Rb]CI cardiac acquisition (Zenodo 10.5281/zenodo.11243763): 5.30×10^8 evaluated three-photon events in binary list mode (per event: float64 x , y , z in mm and the prompt $\rightarrow$ annihilation time difference t in ns), together with the attenuation-correction CT and the reconstructed activity (histo-)image. For instrument characterisation we used the ^{82}Rb uniform quartz-glass reference measurement from the same group (Zenodo 10.5281/zenodo.12636019; 3.91×10^8 events). No new data were acquired.

2.2 List-mode coordinate mapping and chamber identity

The list-mode axes are not the image axes; we determined the mapping empirically rather than assuming it. Among all 48 axis-permutation/sign combinations, the mapping that reproduces the provider’s own reconstructed activity image — scored by coarse (16-voxel-block, flip-sensitive) correlation — is unambiguous (correlation 0.998; all 47 alternatives ≤ 0.60). Because the CT, activity image and segmentations share an LPS orientation, this fixes RV/LV identity with no flip; the conclusion was cross-checked by a joint multi-organ alignment panel (heart, liver, spleen, both kidneys), which the same mapping wins. No lifetime information enters this determination. (For completeness: the mapping that reproduces the activity image differs from a column-reorder helper distributed with the dataset, which appears to target a CT-aligned frame; the activity-image reproduction is the operative ground truth here.)

2.3 Anatomical masks

Organ and chamber masks were generated from the subject’s CT with TotalSegmentator [7] (the `total` task for organs including liver, spleen and kidneys, and the `heartchambers_highres` task for RV, LV, right and left atria, aorta, pulmonary artery and myocardium).

2.4 Lifetime model

For each region we accumulated a Δt histogram (0.1 ns bins) and fitted a background plus a two-component model: a flat accidental background B , a short component, and the o-Ps component, each a decay exponential

convolved with a Gaussian instrument response (an EMG), over $t \in [-3, 12]$ ns with Poisson weights. τ_{oPs} is the long component. A model-independent tail single-exponential fit (after the prompt peak) served as a cross-check; it agreed on sign and approximate magnitude. The same model was applied identically to every region, so model systematics partially cancel in the contrast.

2.5 Registration stress test

We modelled registration uncertainty as a global integer-voxel translation of the event cloud relative to the masks (equivalently, shifting both masks together), and refitted every shift. Two sweeps were run: a dense ± 3 -voxel grid (≤ 8 mm, 342 shifts) and a per-axis sweep to ± 10 voxels (± 16 mm).

2.6 Confound battery

We tested, individually: (i) chamber-wall partial volume, by eroding masks to their blood core versus the near-wall rim; (ii) lung-air partial volume, by correlating τ with a lung-air fraction field (lung-lobe masks convolved with the PET point-spread function); (iii) wall-myocardium leakage, by correlating τ with a myocardium point-spread field; (iv) tissue density, via per-compartment CT Hounsfield units; and (v) instrument/reconstruction positional behaviour, using the isotope-matched ^{82}Rb uniform-quartz reference, in which any spatial lifetime variation is by construction non-biological.

2.7 Reproducibility

The analysis was implemented and executed by two independent pipelines that converged on every reported number; during development the second pipeline twice identified and corrected an over-statement (see Declarations and the public code's audit log). Software: Python 3.10, NumPy, SciPy, NiBabel; scripts and intermediate products are released (Section: Code availability).

3. Results

3.1 A reproducible, registration-stable RV>LV contrast

Per-chamber lifetimes were $\tau_{oPs}(\text{RV}) = 1.700 \pm 0.172$ ns ($n = 2.36 \times 10^6$ events) and $\tau_{oPs}(\text{LV}) = 1.396 \pm 0.130$ ns ($n = 1.95 \times 10^6$), giving $\text{RV} - \text{LV} = +0.304$ ns ($\approx 1.4\sigma$) — the direction oxygen-quenching predicts (Figure 1). The contrast was robust to registration: across the dense ± 3 -voxel grid the sign was preserved in 100% of 342 shifts with a dispersion of only 0.09 ns, and it remained positive along every single axis out to ± 16 mm (Figures 2–4). Registration is therefore not the limiting factor.

3.2 The contrast resides in blood, not the wall

Eroding each chamber to its blood core increased the contrast (core $\Delta\tau = +0.43$ ns), whereas the near-wall rim erased it (RV rim $1.54 \approx \text{LV rim } 1.51$ ns, $\Delta\tau = +0.03$ ns). Partial volume from the chamber wall was diluting, not creating, the contrast.

3.3 Chamber identity is confirmed

The mapping underlying these assignments reproduces the provider's activity image at correlation 0.998 (versus ≤ 0.60 for all alternative orientations) and wins the joint multi-organ alignment panel (Section 2.2). RV is anatomically the right ventricle; there is no flip. This was established without using any lifetime information.

3.4 Oxygenation ladder and the matched-vessel control (key result)

Extending to the cardiopulmonary compartments (Table 1, Figure 5) shows that the venous>arterial tendency is carried by the right-heart chambers, not by oxygenation per se. The decisive comparison is the structure- and position-matched control: the pulmonary artery (deoxygenated) and the aorta (oxygenated), two central great vessels, have indistinguishable lifetimes — pulmonary artery – aorta = -0.011 ± 0.344 ns. The ± 0.34 ns uncertainty exceeds the dissolved-oxygen effect expected on physiological grounds (tens of ps), so this control

rules out an oxygenation effect of the RV/LV magnitude (~ 0.3 ns) but does not exclude a small physiological effect.

3.5 Confound exclusions

Lung-air partial volume is disfavoured (the eroded blood cores are essentially air-free and τ does not correlate with the lung-air fraction, $r = -0.04$). Wall-myocardium leakage is refuted directly: the LV core has $\sim 3\times$ more nearby myocardium than the RV core (myocardium point-spread 0.044 vs 0.014) yet a lower lifetime ($r(\tau, \text{myocardium-PSF}) = +0.06$). CT Hounsfield units do not separate blood from muscle (86% overlap) and are uninformative.

3.6 Isotope-matched instrumental positional envelope

In the uniform ^{82}Rb quartz reference — where any spatial lifetime variation is non-biological — the instrument response is consistent with the human fits (prompt offset $t_0 = 0.067$ ns $\approx$ human 0.066; Gaussian width $\sigma = 0.088$ ns) and the absolute lifetime is reasonable (1.71 ns versus 1.589 ns by the original Bayesian/voxel-wise method; the difference reflects model and volume-of-interest choices). Critically, the reconstructed lifetime in this single uniform material varies by $\sim 0.1\text{--}0.35$ ns over ~ 40 mm (a systematic envelope, not a calibrated gradient — its direction is unstable across core sizes). This isotope-matched envelope is comparable to the entire RV–LV effect; the previously cited ^{124}I phantom bound (≤ 0.08 ns) understates it, plausibly because ^{82}Rb 's high positron energy degrades spatial localisation. The cardiac contrast therefore falls within the scanner's own positional behaviour.

3.7 Residual right-heart elevation: mechanism open

After excluding oxygenation (of the observed magnitude), lung-air, wall-myocardium, registration and gross instrumental bias, what remains is a $\sim 1.5\sigma$ elevation specific to the right-heart chambers whose mechanism we leave open. An interim hypothesis (internal trabecular muscle plus ^{82}Rb myocardial avidity) is plausible but unproven; regional scatter and bolus kinetics are not excluded.

4. Discussion

The natural reading of an isolated RV>LV lifetime difference in the oxygen-expected direction is that o-Pos lifetime is sensing blood oxygenation. Three lines of evidence argue against that reading here. First, the magnitude (~ 0.3 ns) is roughly an order of magnitude larger than the dissolved-oxygen effect implied by published quenching constants for the arterial–venous difference [1] — so, if real, it is too large to be dissolved-oxygen quenching alone. Second, and decisively, the structure- and position-matched control (pulmonary artery versus aorta) is null: when oxygenation is varied while structure and position are held fixed, no lifetime difference appears. Third, the apparent venous>arterial trend is carried by the right-heart chambers, which differ from the great vessels in trabeculation, wall geometry and surrounding tissue, and the effect lies within the scanner's own isotope-matched positional envelope.

Why is the contrast so easily confounded? In cardiac anatomy, “deoxygenated” travels with “right heart / specific position / specific structure”, so a single RV>LV comparison cannot separate oxygenation from anatomy. The pulmonary vessels are the one place where oxygenation and vessel type disagree, which is exactly why the pulmonary-artery-versus-aorta control is informative. We also note that deoxyhaemoglobin, the dominant paramagnetic species in venous blood, would shorten the RV lifetime — the opposite of what is observed — so it is not driving the contrast either. Finally, ^{82}Rb is a high-positron-energy, myocardium-avid tracer; both properties (degraded localisation and muscle uptake) work against a clean blood-pool oxygenation read.

The practical implication for in-vivo PLI is methodological: a lifetime contrast between two anatomical regions is not evidence of an oxygenation difference unless oxygenation has been varied with structure and position controlled, and unless the contrast exceeds the scanner's isotope-matched positional envelope at adequate effective statistics. The confound-control battery used here — a matched-vessel control, blood-core isolation, partial-volume checks, and an isotope-matched uniform-material reference — is portable to any future o-Pos oxygenation claim and is released with this work.

5. Limitations

This is a single subject, and the central contrasts are weak ($\sim 1.5\sigma$): the robust findings are the no-detectable-difference of the matched control and the disavowing of named confounds, not a positive claim of zero oxygenation effect. The matched-control uncertainty (± 0.34 ns) exceeds the expected physiological dissolved-oxygen effect (tens of ps), so a small oxygenation contribution cannot be excluded. The instrumental positional envelope is a systematic bound, not a calibrated gradient (its direction is unstable across analysis volumes). The o-Ps component sits on a large flat accidental background, so its effective statistics are modest, and the absolute lifetime is model-dependent (~ 0.1 ns; our quartz value 1.71 vs the reference 1.589 ns). Masks were derived from the subject's CT; thin systemic vessels were too noisy to use. The mechanism of the residual right-heart elevation is unresolved. The strongest version of this test requires replication across subjects with anatomy decoupled from oxygenation, and, ultimately, an independent oxygenation ground truth.

6. Conclusion

In a public human cardiac LAFOV PET/CT dataset, a reproducible, registration-stable, blood-borne and chamber-identity-confirmed RV>LV o-Ps lifetime contrast (~ 0.3 ns) in the oxygen-expected direction is not attributable to blood oxygenation at the achievable precision: the structure/position-matched pulmonary-artery-versus-aorta control is null, and the effect lies within the scanner's own isotope-matched positional envelope. In-vivo o-Ps lifetime did not yield a clean oxygenation readout here. We release a reusable confound-control methodology to evaluate such claims, and call for multi-subject, oxygenation-decoupled replication before cardiac o-Ps lifetime contrasts are interpreted as oxygenation maps.

Declarations

Ethics approval and consent. This work is a secondary analysis of publicly available, fully de-identified data; it involved no new human-subjects data collection and required no additional ethical approval. The original in-vivo data were acquired and shared by the Bern/Inselspital group under their institutional approvals [3,4] and are distributed under CC-BY-4.0; the analysed CT does not include the face. The author had no access to identifiable participant information.

Data availability. All inputs are public on Zenodo: the [^{82}Rb]CI human dataset (10.5281/zenodo.11243763) and the ^{82}Rb quartz reference (10.5281/zenodo.12636019); a ^{124}I phantom record (10.5281/zenodo.13443797) is referenced for context. No raw data are redistributed here.

Code availability. Analysis code, intermediate result tables, figures, a key-numbers file, environment specification and step-by-step reproduction instructions are available at <https://github.com/Jefemaestro33/rv-lv-paper> (MIT for code; CC-BY-4.0 for text and figures). An archival DOI can be added after public release.

Competing interests. The author declares no competing interests.

Funding. No specific funding was received for this work.

Author contributions. E.D.P.Z. conceived the study, performed the analysis, and wrote the manuscript. Analyses were implemented and independently cross-checked by two AI-assisted pipelines under the author's direction; the AI systems are tools and are not authors.

Acknowledgements. We thank L. Mercolli, W. Steinberger and colleagues (Inselspital Bern) for openly depositing the positronium-lifetime data that made this re-analysis possible.

Figure legends

Figure 1. Per-chamber positronium-lifetime spectra (Δt histograms, log scale) for the right and left ventricles at baseline registration, with the background-plus-two-component EMG fits and the fitted flat background. The o-Ps signal is a modest excess over a large accidental background; $\tau_{\text{oPs}}(\text{RV}) = 1.700$ ns, $\tau_{\text{oPs}}(\text{LV}) = 1.396$ ns.




Figure 2. Distribution of the RV–LV lifetime difference across the dense ± 3 -voxel (≤ 8 mm) registration grid ($n = 342$ shifts). The contrast is positive in 100% of shifts (dispersion 0.09 ns); dashed line, baseline; the difference never crosses zero.




Figure 3. RV–LV lifetime difference versus single-axis registration shift (± 16 mm) along each axis. The difference stays positive across the full tested range.




Figure 4. RV-LV lifetime difference versus the magnitude of the registration shift; positive throughout.




Figure 5. Oxygenation ladder. (Left) per-compartment o-Ps lifetime, coloured by oxygenation class, with the pulmonary crossover controls highlighted; (right) lifetime versus nominal oxygen saturation. The flat trend and the null pulmonary-artery-aorta pair indicate the contrast does not track oxygenation.





Figure 6. Lung-air partial-volume check: o-Ps lifetime versus the local lung-air fraction (and the matched-stratum comparison). The blood cores are essentially air-free and lifetime does not track the lung-air fraction.



Tables

Table 1. Per-compartment o-Ps lifetime (eroded blood cores).

Compartment	Class	τ_{oPs} (ns)	n ($\times 10^6$)
Right ventricle	deoxygenated	1.731 ± 0.197	2.05
Right atrium	deoxygenated	1.640 ± 0.332	1.25
Superior vena cava	deoxygenated	1.444 ± 0.505	0.30
Inferior vena cava	deoxygenated	1.425 ± 0.412	0.33
Pulmonary artery	deoxygenated	1.259 ± 0.312	0.54
Left atrium	oxygenated	1.302 ± 0.270	0.90
Left ventricle	oxygenated	1.358 ± 0.134	1.71
Aorta	oxygenated	1.270 ± 0.145	1.74
Myocardium (control)	muscle	1.662 ± 0.206	1.10

Table 2. Confound-control battery.

Confound	Test	Result
Registration error	translational sweep ± 16 mm	sign-stable (100%/342 shifts, σ 0.09 ns) — not the driver
Chamber-wall partial volume	core vs near-wall rim	core $\Delta +0.43$; rim $\Delta +0.03$ — signal is in blood, not wall
Blood oxygenation	matched control (pulmonary artery vs aorta)	$\Delta = -0.01 \pm 0.34$ ns — no effect of the RV/LV magnitude
Lung-air partial volume	lung-PSF field vs τ	air ≈ 0 in cores; $r = -0.04$ — disfavoured
Wall-myocardium leakage	myocardium-PSF field vs τ	LV has 3 $\times$ more yet lower τ ; $r = +0.06$ — refuted
Instrument/reconstruction position	isotope-matched ^{82}Rb uniform quartz	positional envelope $\sim 0.1\text{--}0.35$ ns / 40 mm — effect lies within it
Tissue density	per-compartment CT HU	blood/muscle 86% overlap — uninformative

References

1. Shibuya K, Saito H, Nishikido F, Takahashi M, Yamaya T. Oxygen sensing ability of positronium atom for tumor hypoxia imaging. *Commun Phys* 3, 173 (2020). doi:10.1038/s42005-020-00440-z
2. Moskal P, Kisielewska D, Curceanu C, Czerwiński E, Dulski K, Gajos A, Gorgol M, Hiesmayr B, Jasińska B, Kacprzak K, Kapłon Ł, Korcyl G, Kowalski P, Krzemień W, Kozik T, Kubicz E, Mohammed M, Niedźwiecki Sz, Pałka M, Pawlik-Niedźwiecka M, Raczyński L, Raj J, Sharma S, Shivani, Shopa RY, Silarski M, Skurzok M, Stępień E, Wiślicki W, Zgardzińska B. Feasibility study of the positronium imaging with the J-PET tomograph. *Phys Med Biol* 64, 055017 (2019). doi:10.1088/1361-6560/aafe20
3. Mercolli L, Steinberger WM, Sari H, Afshar-Oromieh A, Caobelli F, Conti M, Felgosa Cardoso ÂR, Mingels C, Moskal P, Pyka T, Rathod N, Schepers R, Stępień EŁ, Viscione M, Rominger A, Shi K, Seifert R. In vivo positronium lifetime measurements with intravenous tracer administration and a long axial field-of-view PET/CT. *medRxiv* (2024). doi:10.1101/2024.10.19.24315509
4. Steinberger WM, Mercolli L, Breuer J, Sari H, Parzych S, Niedźwiecki S, Lapkiewicz G, Moskal P, Stępień E, Rominger A, Shi K, Conti M. Positronium lifetime validation measurements using a long-axial field-of-view positron emission tomography scanner. *EJNMMI Phys* 11, 76 (2024). doi:10.1186/s40658-024-00678-4
5. Mercolli L, Steinberger WM, Rathod N, Conti M, Moskal P, Rominger A, Seifert R, Shi K, Stępień EŁ, Sari H. Phantom imaging demonstration of positronium lifetime with a long axial field-of-view PET/CT and ^{124}I . *EJNMMI Phys* 12 (2025). doi:10.1186/s40658-025-00790-z
6. Moskal P, Baran J, Bass S, Choiński J, Chug N, Curceanu C, Czerwiński E, Dadgar M, Das M, Dulski K, Eliyan KV, Fronczewska K, Gajos A, Kacprzak K, Kajetanowicz M, Kaplanoglu T, Kapłon Ł, Klimaszewski K, Kobylecka M, Korcyl G, Kozik T, Krzemień W, Kubat K, Kumar D, Kunikowska J, Mączewska J, Migdał W, Moskal G, Mryka W, Niedźwiecki S, Parzych S, del Rio EP, Raczyński L, Sharma S, Shivani, Shopa RY, Silarski M, Skurzok M, Tayefi F, Ardebili KT, Tanty P, Wiślicki W, Królicki L, Stępień EŁ. Positronium image of the human brain in vivo. *Sci Adv* 10, eadp2840 (2024). doi:10.1126/sciadv.adp2840
7. Wasserthal J, Breit HC, Meyer MT, Pradella M, Hinck D, Sauter AW, Heye T, Boll DT, Cyriac J, Yang S, Bach M, Segeroth M. TotalSegmentator: robust segmentation of 104 anatomic structures in CT images. *Radiol Artif Intell* 5, e230024 (2023). doi:10.1148/ryai.230024