

Q1 Interlock Status Display

11:30:20

Mon Nov 16 1998

He

U1e

Tread

N2

PSU

Temp

Coil

Command Line

psu:standby

Overall State NORMAL
N2 system state N2COLD
He System state HeCOLD
PSU State Enabled

Main Coil Quench	<input type="radio"/>	Watchdog	<input type="radio"/>
Cur lead overvoltage	<input type="radio"/>	OVC vacuum high	<input type="radio"/>
Dump switch open	<input type="radio"/>	He neck flow low	<input type="radio"/>
Trin coil Quench	<input type="radio"/>	He reservoir low	<input type="radio"/>
He pressure high	<input type="radio"/>	N2 pressure high	<input type="radio"/>
CEBAF Shutdown	<input type="radio"/>		

F6: ALARM POINT

F7: ALARM ACK