



Downstream Processing using JMP

Example 1



Example 2




Downstream Processing using JMP

Example 3

Attribute Gauge				
Agreement Report				
Rater	% Agreement	95% Lower CI	95% Upper CI	
Call	100.000	24.3293	100.000	
Expected Call	100.000	24.3293	100.000	
Number Inspected	Number Matched	% Agreement	95% Lower CI	95% Upper CI
2184	2184	100.00	99.824	100.00

Example 4



Example 5

Exclusion passing criteria												
Bacteria (VIC) ≤ 5% detection rate at Ct < 35.5												
Fungi (FAM) ≤ 5% detection rate at Ct < 36												
Sample Name & Concentrati	Bacteria (VIC)						Fungi (FAM)					
	7500 Fast			QS5			7500 Fast			QS5		
	Mean Ct	Reps	Detection	Mean Ct	Reps	Detection	Mean Ct	Reps	Detection	Mean Ct	Reps	Detec
xxx	37.51	0/20	0.00%	37.22	1/20	5.00%	39.85	0/20	0.00%	39.94	0/20	0.00%
xxx	38.03	1/20	5.00%	38.49	1/20	5.00%	39.94	0/20	0.00%	40.00	0/20	0.00%
xxx	38.05	0/20	0.00%	38.17	1/20	5.00%	39.77	0/20	0.00%	40.00	0/20	0.00%
xxx	37.88	0/20	0.00%	37.66	0/20	0.00%	39.93	0/20	0.00%	39.89	0/20	0.00%
xxx	38.01	1/20	5.00%	38.42	0/20	0.00%	39.47	1/20	5.00%	39.95	0/20	0.00%
xxx	37.03	1/20	5.00%	38.33	0/20	0.00%	38.76	0/20	0.00%	38.72	0/20	0.00%
xxx	40.00	0/20	0.00%	40.00	0/20	0.00%	37.87	0/20	0.00%	37.88	0/20	0.00%
xxx	40.00	0/20	0.00%	40.00	0/20	0.00%	39.53	0/20	0.00%	39.48	0/20	0.00%

Downstream Processing using JMP

Example 6

