

Sample Report



Sample Metadata

Sample Name	Scan 04212
Hazard Level	Unknown
Last Modified	2024-03-07 11:46:14-05:00
Created	2024-03-07 11:46:14-05:00


Instrument

Serial Number	192600200502138
Firmware Package	8.1.3.55
Last Calibration	2024-03-07 09:58:26-05:00
Last System Suitability Test	Fail, 2024-03-07 09:58:38-05:00


Acquisition Parameters

Operating Procedure Name	Default Procedure
Operating Procedure Version	1
Integration Time	6.41 s
Auto Integration	On
Averages	3
Laser Power	5
Smart Tip Type	Stand Off Attachment Lens
Smart Tip Serial Number	6075060200202022

Spectral Data



Mixture Identification Results



Hazard Level	Match	Sample Type	Source Library	Spectral Weight %
Unknown	4-ANPP PSERS-Ag 1e-4		NATO SERS, Version: 1	71%