

Sample Report



Sample Metadata

Sample Name	Scan 00101
Hazard Level	Clear
Sample Type	Chemical
Last Modified	2024-03-07 10:28:21-05:00
Created	2024-03-07 10:28:21-05:00


Instrument

Serial Number	192601100601640
Firmware Package	8.0.3.56
Last Calibration	2024-03-06 10:11:28-05:00
Last System Suitability Test	Pass, 2024-03-06 10:11:38-05:00


Acquisition Parameters

Operating Procedure Name	Default Procedure
Operating Procedure Version	1
Integration Time	0.87 s
Auto Integration	On
Averages	3
Laser Power	5
Smart Tip Type	Intelligent UA, Surface (Position 1)
Smart Tip Serial Number	6075060600301151

Spectral Data



Identification Results



Mixture Identification Results

