# EyeGlynt — Session Summary

Session S-20261001-01 · Normal Mode

## Patient & session

| | | | |
|---|---|---|---|
| Patient | Juan Dela Cruz (P-001) | Date | October 01, 2026 |
| Condition | ALS | Start – End | 03:44 PM – 03:45 PM |
| Care setting | Home | Duration | 0m 22s |
| Caregiver | Jhovany Honorio (Speech-Language Therapist) | Dwell time | 1.5 s |

## Performance metrics

| Metric | Result | Benchmark | Status |
|---|---|---|---|
| Gaze Accuracy | 91.7% | Min 85% | **Meets** |
| False Positive Rate | 8.3% | Max 10% | **Meets** |
| Detection Latency | 360 ms | Max 500 ms | **Meets** |
| Recal. Success | 100.0% | Min 80% | **Meets** |
| Total selections | 12 | — | |
| Most-used interface | Quick-Fire | — | |

Normal Mode: accuracy assumes selections were intended unless a caregiver flagged them.

## Activity log (12 selections)

| Time | Interface | Selection | Note |
|---|---|---|---|
| 15:44:56 | Quick-Fire | I can't breathe well | |
| 15:44:59 | Quick-Fire | Where is my family? | |
| 15:45:01 | Quick-Fire | I can't breathe well | |
| 15:45:02 | Keyboard | Please | |
| 15:45:05 | Quick-Fire | I need the bathroom | |
| 15:45:06 | Keyboard | No | |
| 15:45:09 | Keyboard | Wait | Flagged unintended |
| 15:45:11 | Quick-Fire | Call my daughter | |
| 15:45:13 | Quick-Fire | Turn off the light | |
| 15:45:15 | Quick-Fire | Thank you | |
| 15:45:16 | Keyboard | Where is mama | |
| 15:45:18 | Keyboard | Watch TV please | |

## System alerts (10)

| Time | Alert | Detail | Latency |
|---|---|---|---|
| 15:44:56 | Lighting shift detected | Ambient light changed to 368.1 lux — recalibration triggered | 302 ms |
| 15:44:56 | Recalibration started | Recalibration started (auto lighting) | |
| 15:44:56 | Recalibration successful | Tracking restored | |
| 15:45:03 | Gaze lock lost | Eyes not detected — user looked away or closed eyes | 352 ms |
| 15:45:05 | Gaze lock restored | User back in view | |
| 15:45:12 | Recalibration started | Recalibration started (manual caregiver) | |

| Time | Alert | Detail | Latency |
|---|---|---|---|
| 15:45:13 | Recalibration successful | Tracking restored | |
| 15:45:14 | Lighting shift detected | Ambient light changed to 185.5 lux — recalibration triggered | 427 ms |
| 15:45:14 | Recalibration started | Recalibration started (auto lighting) | |
| 15:45:15 | Recalibration successful | Tracking restored | |

## Caregiver notes

**Before session:** Patient slightly fatigued after lunch.
**After session:** Responded well; used Quick-Fire mostly.