Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

Pressure Test Certificate

Test Date: **2020-09-12**   Description: **SUNA-V2**

Sensor Information:

Model Number: **SUNA-V2**

Serial Number: **1471**

Pressure Test Protocol:

Low Pressure Test: **40**   PSI   Held For: **15**   Minutes

High Pressure Test: **150**   PSI   Held For: **15**   Minutes

Passed Test: **True**

Tested By: **dn**

High pressure is generally equal to the maximum depth rating of the instrument

Pressure

Time

Typical Test Profile