



**SEA-BIRD**  
SCIENTIFIC

## SBE56 ECO Temperature Logger

### Instrument Configuration

Instrument Serial Number: 56-13393  
Instrument Firmware Version: 1.02  
Communications Format: RS232

### Installed Devices/Sensors

<i>Data Format</i>	<i>Measurement</i>	<i>Sensor Type</i>	<i>Serial Number</i>	<i>Rating</i>
Count	Temperature	Internal	N/A	N/A

Maximum Depth: **1500m**

**CAUTION** - The maximum deployment depth will be limited by the measurement range of the pressure sensor, if installed, an attached sensor, if installed, or the housing.



Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

SENSOR SERIAL NUMBER: 13393  
CALIBRATION DATE: 20-Aug-23

SBE 56 TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -1.332008e-003  
a1 = 3.554179e-004  
a2 = -6.936808e-006  
a3 = 2.096394e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
-1.5000	20176480.7	-1.5000	0.0000
1.0000	18037750.6	1.0000	-0.0000
4.5000	15463717.4	4.5000	0.0000
8.0000	13300824.2	8.0000	0.0000
11.5000	11477198.2	11.5000	-0.0000
15.0000	9934493.7	15.0000	0.0000
18.5000	8625238.3	18.5000	-0.0000
22.0000	7510608.0	22.0000	-0.0000
25.5000	6558736.3	25.5000	0.0000
29.0000	5743489.1	29.0000	-0.0000
32.5000	5043199.1	32.5000	-0.0000

n = Instrument Output (counts)

Temperature ITS-90 (°C) =  $1/\{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

