Calibration Date:	06/01/16			**Job No.:**	R12596
Model Number:	QSP2300				
Serial Number:	70361				
Operator:	TPC				
Standard Lamp:	V-035(3/4/15)				
Operating Voltage Range:	6	to	15	VDC (+)	

Note: The QSP2300 output is a voltage that is proportional to the log of the incident irradiance.
To calculate irradiance, use this formula:
Irradiance = Calibration factor * (10^Light Signal Voltage - 10^Dark Voltage)

Dry Calibration Factor:	**4.88E+12**	**quanta/cm²·sec per volt**	**8.11E-06**	**µEinsteins/cm²·sec per volt**
Wet Calibration Factor:	**8.62E+12**	**quanta/cm²·sec per volt**	**1.43E-05**	**µEinsteins/cm²·sec per volt**

Sensor Test Data and Results[2)]

Sensor Supply Current (Dark):	3.4	mA		
Supply Voltage:	6	Volts		
Lamp Integrated PAR Irradiance:	1.03E+16	quanta/cm²·sec	0.01713	µEinsteins/cm²sec
Immersion Coefficient:	0.566			

Nominal Filter OD	Expected Transmission	Calibrated Trans.	Sensor Voltage	Expected Voltage	Voltage % Error	Measured Trans.	Transmission Error (%)	Test Irrad. (quanta/ cm²·sec)
No Filter	100%	100.00%	3.325	3.325	0%	100.00%	0.0	1.03E+16
0.3	50%	36.10%	2.885	2.883	0%	36.24%	-0.4	3.74E+15
0.5	32%	27.60%	2.770	2.766	0%	27.80%	-0.7	2.87E+15
1	10%	9.27%	2.296	2.292	0%	9.31%	-0.4	9.61E+14
2	1%	1.11%	1.373	1.370	0%	1.07%	3.8	1.10E+14
3	0.10%	0.05%	0.216	0.053	76%	0.03%	76.9	3.15E+12
RG780	0.00%	0.00%	0.003	0.003	-5%	0.00%	-100.0	3.05E+10

Dark Before:	0.003	Volts
Light - No Filter Hldr.:	3.325	Volts
Dark After - NFH:	0.003	Volts
Average Dark	0.0028	Volts

Notes:
1. Annual calibration is recommended.
2) This section is for internal use and for more advanced analysis.