



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer Pacific Marine Environmental Lab
Job Number 62186
Date 2/1/2011
Technician DG

Serial Number 37SM25541-1863

Low Pressure (PSI) 50 PSI
Time (Minutes) 15 Minutes

High Pressure (PSI) 10000 PSI
Time (Minutes) 30 Minutes

Pass ☒

Fail ☐

Comments

Replaced the main piston "O"-Rings.

