PO Box 518 620 Applegate St. Philomath, OR 97370



(541) 929-5650 Fax (541) 929-5277 www.wetlabs.com

Chlorophyll WETStar Characterization

Date: January 7, 2014 S/N: WS3S-748P

Chlorophyll concentration expressed in µg/l can be derived using the equation:

 $CHL(\mu g/I) = Scale Factor \times (Output - Clean Water Offset)$

Analog output

Clean Water Offset (CWO) 0.066 VScale Factor (SF) $5.3 \mu g/I/V$

Maximum Output 5.46 V Resolution 0.28 mV Ambient Characterization Temperature 22 \pm 1°C

Current Draw 30 mA @ 12V (typical)

12-hour Stability 0.22 mV/hr Temperature Stability, 25–2 °C 0.33 mV/°C

Range	
15 μg/l	0
26 µg/l	Х
150 µg/l	0

Definitions:

CWO: Clean Water Offset value obtained using pure filtered de-ionized water.

SF: Scale Factor is used to convert the fluorescence response of the instrument into chlorophyll-a concentration. Scale Factor is determined at WET Labs during a cross calibration using a liquid fluorescent standard and a reference fluorometer whose chlorophyll fluorescence response has been characterized in a laboratory using a mono-species lab culture of *Thalassiosira weissflogii* phytoplankton.

Maximum Output: Maximum signal output of the fluorometer.

Resolution: Standard deviation of 1 minute of clean water data, sampled once per second.

Ambient Characterization Temperature: Room temperature at time of characterization.

Current Draw: The amount of current the instrument uses for operation.

12-hour Stability: Deviation of output averaged over 12 hours.

Temperature Stability: Measured output variation per degree.

WS3S-748P.xlsx Revision I 10/3/07

PO Box 518 620 Applegate St. Philomath, OR 97370



(541) 929-5650 Fax (541) 929-5277 www.wetlabs.com

WETStar Calibration and Repairs

Date January 7, 2014 Customer NOAA

S/N# WS3S-748P Repair Order 21594

Standard Service

- Performed noise test: 1 sample/sec for 60 sec
- Performed stability test: 1 sample/min for 12 hrs
- Performed temperature test: 25-2 °C
- Performed saturation test
- Shake-tested unit
- Pressure-tested unit
- Updated unit's calibration sheet

Diagnosis

Red Filter was delaminating.

Repairs

Replaced Red Filter and O-Rings.

Comments

WETStar was re-calibrated with 23.5ppb Uranine.

WS3S-748P.xlsx Revision I 4/17/08