



Sea-Bird Electronics, Inc.

13431 NE 20th St. Bellevue, Washington 98005 USA
www.seabird.com

Phone: (425) 643-9866

Fax: (425) 643-9954

Email: seabird@seabird.com

Pressure Test Certificate

Test Date: **2017-04-04**

Description: **SBE-16P SeaCat**

Sensor Information:

Replaced the main piston "O"-Rings.

Model Number: **SBE-16P**

Serial Number: **4424**

Pressure Test Protocol:

Low Pressure Test: **40** PSI Held For: **15** Minutes

High Pressure Test: **870** PSI Held For: **15** Minutes

Passed Test: **True**

Tested By: **nd**

High pressure is generally equal to the maximum depth rating of the instrument

