



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer Pacific Marine Environmental Lab

Job Number 88417

Date 3/22/2016

Technician BLT

Serial Number 3229804-0418

Low Pressure (PSI) 50 PSI

Time (Minutes) 60 Minutes

High Pressure (PSI) N/A PSI

Time (Minutes) N/A Minutes

Pass ☒

Fail ☐

Comments

Replaced the main piston "O"-Rings. Replaced main housing. * Due to the dimensions of the CAROUSEL we are unable to perform a high pressure test at this time.

