



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer Pacific Marine Environmental Lab

Job Number 86981

Date 12/3/2015

Technician CA

Serial Number 3928082-0817

Low Pressure (PSI) 45 PSI

Time (Minutes) 15 Minutes

High Pressure (PSI) 500 PSI

Time (Minutes) 30 Minutes

Pass ☒

Fail ☐

Comments

Replaced the main piston "O"-Rings.

