



**Sea-Bird Electronics, Inc.**

13431 NE 20<sup>th</sup> St. Bellevue, Washington 98005 USA  
www.seabird.com

Phone: (425) 643-9866

Fax: (425) 643-9954

Email: seabird@seabird.com

## Pressure Test Certificate

Test Date: **2017-01-20**

Description: **SBE-16 SeaCat**

### Sensor Information:

Replaced conductivity cell.

Replaced the main piston "O"-Rings.

Model Number: **SBE-16**

Serial Number: **0658**

### Pressure Test Protocol:

Low Pressure Test: **40**      PSI      Held For: **15**      Minutes

High Pressure Test: **870**      PSI      Held For: **30**      Minutes

Passed Test: **True**

Tested By: **JK**

**High pressure is generally equal to the maximum depth rating of the instrument**

