

a xylem brand

Electronic board: 3623 Product: RCM 9
Electronic board serial: 1145 Serial No: 905

Reference reading: 834 Calibration Date: May 21, 2015

For details; see the individual Calibration Sheets.

The calibration coefficients listed below are valid for sensors with the following serial numbers:

Sensor	Type	Serial No.		Range	
Doppler Current Sensor	3920	503			
Temperature Sensor	3621	1162			
			Arctic:		
			High:		
			Low:	-2.83 to 21.76	
			Wide:		
Conductivity Sensor	N/A				
Pressure Sensor	3815C	450		0 - 7000kPa	
Turbidity Sensor	3612	793		0-20	
Oxygen Sensor	3835	1777			

Calibration Coefficients:

Ch. No.	Parameter	A	В	С	D	Unit
1	Reference	0	1.00E+00	0		-
2	Current Speed	0	2.93E-01	0		cm/s
3	Current Direction	0	3.52E-01	0		Deg. M
4	Temperature Range					
	Arctic					Deg. C
	High					Deg. C
	Low	-2.826E+00	2.417E-02	-2.238E-06	2.056E-09	Deg. C
	Wide					Deg. C
	Other					
5	Conductivity					mS/cm
6	Pressure	-1.695E+02	7.085E+00	2.715E-05		kPa
7	Turbidity					NTU
8	Oxygen	-3.223E+00	1.149E+00			uM
	O2 SR10 Scaling Coef	0	0.488281			

Date: May 21, 2015 Sign: Jason B Baxter