



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer Pacific Marine Environmental Lab

Job Number 64188

Date 6/1/2011

Technician DG

Serial Number 3925210-0516

Low Pressure (PSI) 50 PSI

Time (Minutes) 30 Minutes

High Pressure (PSI) *N/A PSI

Time (Minutes) *N/A Minutes

Pass ☒

Fail ☐

Comments

Replaced the main piston "O"-Rings.

*The unit is not designed for high pressure applications.

