



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

Pressure Test Certificate

Test Date: **6/4/2024**

Description: **SBE-16P SeaCat**

Sensor Information:

Replaced the main piston "O"-Rings.
Replaced 6-pin connector.

Model Number: **16P**

Serial Number: **6902**

Pressure Test Protocol:

Low Pressure Test: **40** PSI Held For: **15** Minutes

High Pressure Test: **870** PSI Held For: **15** Minutes

Passed Test: **True**

Tested By: **JW**

