Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 0045
CALIBRATION DATE: 22-Dec-20

Prawler CTD CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

```
g = -1.011763e+000
h =  1.470170e-001
i = -2.173253e-004
j =  3.786653e-005

CPcor = -9.5700e-008
CTcor =  3.2500e-006
WBOTC =  9.5198e-008
```

| BATH TEMP (° C) | BATH SAL (PSU) | BATH COND (S/m) | INSTRUMENT OUTPUT (Hz) | INSTRUMENT COND (S/m) | RESIDUAL (S/m) |
|---|---|---|---|---|---|
| 22.0000 | 0.0000 | 0.00000 | 2626.11 | 0.00000 | 0.00000 |
| 1.0000 | 34.7383 | 2.96993 | 5206.03 | 2.96994 | 0.00001 |
| 4.5000 | 34.7198 | 3.27652 | 5402.09 | 3.27651 | -0.00001 |
| 15.0000 | 34.6805 | 4.25671 | 5985.28 | 4.25671 | -0.00000 |
| 18.5000 | 34.6731 | 4.60142 | 6177.03 | 4.60142 | 0.00000 |
| 23.9940 | 34.6655 | 5.15805 | 6474.40 | 5.15805 | -0.00000 |
| 29.0000 | 34.6626 | 5.67997 | 6741.01 | 5.67997 | 0.00000 |
| 32.5000 | 34.6617 | 6.05207 | 6924.64 | 6.05207 | -0.00000 |

f = Instrument Output(Hz) * sqrt(1.0 + WBOTC * t) / 1000.0

t = temperature (°C); p = pressure (decibars); $\delta$ = CTcor; $\varepsilon$ = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \varepsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity